

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122422\
 Data File : BN023445.D
 Acq On : 25 Dec 2022 12:22
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4359

Quant Time: Dec 26 01:23:21 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Dec 25 05:17:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.989	152	7247	0.400	ng/ul	0.00
4) Naphthalene-d8	10.804	136	24017	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.637	164	13603	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.387	188	29326	0.400	ng/ul	0.00
17) Chrysene-d12	21.576	240	18898	0.400	ng/ul	0.00
23) Perylene-d12	24.022	264	15197	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	2671	0.339	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.393	152	14994	0.404	ng/ul	0.00
18) Fluoranthene-d10	19.416	212	30985	0.453	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.378	88	2762	0.343	ng/ul#	93
5) Naphthalene	10.854	128	28056	0.398	ng/ul	100
7) 2-Methylnaphthalene	12.465	142	17986	0.399	ng/ul	100
8) 1-Methylnaphthalene	12.685	142	18460	0.397	ng/ul	100
10) Acenaphthylene	14.359	152	25741	0.427	ng/ul	99
11) Acenaphthene	14.702	153	20081	0.395	ng/ul	98
12) Fluorene	15.687	166	22365	0.392	ng/ul	99
14) Pentachlorophenol	17.036	266	3455	0.425	ng/ul	100
15) Phenanthrene	17.425	178	36473	0.382	ng/ul	100
16) Anthracene	17.518	178	32723	0.413	ng/ul	100
19) Fluoranthene	19.444	202	40331	0.440	ng/ul	99
20) Pyrene	19.807	202	40500	0.438	ng/ul	100
21) Benzo(a)anthracene	21.558	228	31605	0.430	ng/ul	99
22) Chrysene	21.614	228	30253	0.404	ng/ul	100
24) Benzo(b)fluoranthene	23.274	252	27183	0.346	ng/ul	99
25) Benzo(k)fluoranthene	23.324	252	25957	0.352	ng/ul	99
26) Benzo(a)pyrene	23.911	252	22834	0.367	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.578	276	27281	0.376	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.598	278	21725	0.386	ng/ul	99
29) Benzo(g,h,i)perylene	27.366	276	22683	0.383	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122422\
Data File : BN023445.D
Acq On : 25 Dec 2022 12:22
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 73 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4359

Quant Time: Dec 26 01:23:21 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Dec 25 05:17:59 2022
Response via : Initial Calibration

