


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122721\
 Data File : BN018205.D
 Acq On : 27 Dec 2021 16:16
 Operator : CG/JU
 Sample : DFTPP219
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP219

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122721.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 28 00:34:45 2021



AutoFind: Scans 2459, 2460, 2461; Background Corrected with Scan 2452

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	33.9	419929	PASS
68	69	0.00	2	1.2	4976	PASS
69	198	0.00	100	33.8	417920	PASS
70	69	0.00	2	0.6	2531	PASS
127	198	10	80	37.8	467397	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1237333	PASS
199	198	5	9	6.9	85075	PASS
275	198	10	60	29.6	366443	PASS
365	198	1	100	3.9	47872	PASS
441	198	0.01	100	14.2	175659	PASS
442	198	50	100	84.9	1050176	PASS
443	442	15	24	20.4	213739	PASS

m/z	Abundance
35.90	1193.0
37.10	486.0
38.10	983.0
39.10	3284.0
40.00	2686.0
41.00	3996.0
42.00	3656.0
43.00	8132.0
44.00	20904.0
45.00	2923.0
46.00	1712.0
46.80	320.0
49.00	1397.0
50.00	719.0
51.00	1788.0
52.10	977.0
53.00	1461.0
53.90	719.0
55.00	3045.0
56.00	976.0
57.00	3215.0
58.00	578.0
59.00	427.0
60.00	985.0
61.10	355.0
63.10	721.0
65.00	446.0
66.00	503.0
66.90	696.0
68.10	1299.0
69.10	1941.0
70.00	867.0
70.90	1228.0
71.90	349.0
73.10	2909.0
73.90	1143.0
75.10	415.0
76.10	315.0
77.10	742.0
78.10	886.0
79.00	1087.0
80.00	495.0
80.90	810.0
81.90	622.0
83.10	904.0
83.90	1835.0
85.00	885.0
86.00	804.0
87.00	550.0
87.90	564.0
88.90	452.0
91.00	1121.0
93.00	405.0
93.90	399.0
95.00	1472.0
96.00	649.0
97.10	653.0
98.00	1352.0
99.10	452.0
100.20	339.0
103.00	432.0
104.90	643.0
107.10	316.0
107.90	564.0
109.10	348.0
110.00	777.0
112.00	373.0
115.10	620.0
116.00	497.0
117.00	641.0
121.00	331.0
124.00	310.0
126.10	745.0
127.10	315.0
128.90	356.0
129.20	310.0
131.20	516.0
133.00	842.0
134.10	454.0
135.10	620.0
142.10	499.0
142.90	326.0
147.10	621.0
151.10	352.0
159.10	335.0
161.20	308.0
162.80	355.0
167.00	967.0
169.10	356.0
179.00	418.0
182.10	487.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

190.00	349.0
191.00	661.0
193.00	895.0
193.90	337.0
200.10	379.0
204.10	398.0
205.20	318.0
207.00	3494.0
208.00	870.0
209.00	705.0
223.00	337.0
280.90	889.0
283.00	358.0
340.90	304.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

m/z	Abundance
36.00	1380.0
37.00	2098.0
38.00	5044.0
39.00	34272.0
40.00	4535.0
41.00	3357.0
42.00	2539.0
43.00	8101.0
44.00	16664.0
45.00	2890.0
45.90	881.0
47.00	420.0
47.90	309.0
48.10	313.0
49.10	4351.0
50.00	116240.0
51.00	442816.0
52.00	22176.0
53.00	1662.0
54.00	724.0
55.00	4777.0
56.00	12543.0
57.00	30760.0
58.10	1843.0
59.10	477.0
59.90	1721.0
61.00	5975.0
62.00	6236.0
63.00	16173.0
64.00	3341.0
65.00	10097.0
65.90	836.0
67.00	1314.0
68.00	9165.0
69.00	426112.0
70.00	2738.0
71.10	839.0
72.10	728.0
73.00	6843.0
74.00	46872.0
75.00	72808.0
76.00	26080.0
77.00	509760.0
78.10	33240.0
79.00	33328.0
80.00	25048.0
81.00	35576.0
82.00	10026.0
83.00	8731.0
84.00	2027.0
85.00	8761.0
86.00	9855.0
87.00	4824.0
88.00	2127.0
89.00	1139.0
89.80	495.0
91.00	8151.0
92.00	8924.0
93.00	54424.0
94.00	4080.0
94.90	2308.0
96.00	2899.0
96.90	2281.0
98.00	44480.0
99.00	33720.0
100.00	3789.0
101.00	18224.0
102.10	1415.0
103.00	6489.0
104.00	10997.0
105.00	12229.0
106.00	4243.0
107.00	145984.0
108.00	25552.0
109.00	4943.0
110.00	252416.0
111.00	38528.0
112.10	4063.0
113.00	1834.0
114.20	684.0
115.10	1522.0
116.00	7904.0
117.00	126536.0
118.00	8816.0
119.00	1282.0
120.00	2296.0
120.90	691.0
121.20	757.0
121.90	9287.0
123.00	13577.0
123.90	6286.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

125.00	7149.0
127.00	445504.0
128.00	37144.0
129.00	178368.0
130.00	16520.0
131.00	3395.0
132.10	1373.0
133.10	1646.0
134.00	6029.0
135.00	15716.0
136.00	6456.0
137.00	8561.0
138.00	1654.0
139.00	1677.0
140.00	2421.0
141.00	23064.0
142.00	8230.0
143.00	5433.0
144.10	1350.0
145.00	2040.0
146.00	4401.0
147.00	11792.0
148.00	34344.0
149.00	7030.0
150.10	1519.0
151.10	2980.0
151.90	2123.0
153.10	7977.0
153.90	7323.0
155.00	13791.0
156.10	21736.0
157.10	3871.0
158.00	5561.0
159.00	4330.0
159.90	7520.0
161.00	10694.0
162.10	2621.0
163.20	1700.0
164.10	1519.0
165.00	10092.0
166.00	8285.0
167.00	56096.0
168.00	37264.0
169.00	5420.0
170.00	1758.0
171.00	1866.0
172.00	4623.0
173.10	5999.0
174.00	11843.0
175.00	21248.0
175.90	6051.0
177.00	9823.0
178.00	3297.0
179.00	42232.0
180.00	28920.0
181.00	13714.0
182.10	2823.0
183.00	1289.0
184.00	3517.0
185.00	22144.0
186.00	162624.0
187.00	45080.0
188.10	5326.0
189.00	9558.0
190.00	2176.0
191.00	6306.0
192.00	13723.0
193.00	15591.0
193.90	3637.0
195.10	1791.0
196.00	32000.0
198.00	1145856.0
199.00	82464.0
200.00	5528.0
201.70	4051.0
203.00	6869.0
204.00	41160.0
205.00	70064.0
206.00	309056.0
207.10	43448.0
208.10	9272.0
209.00	4209.0
210.10	4622.0
210.90	10912.0
213.00	1265.0
215.10	3429.0
216.10	6448.0
217.00	79168.0
217.90	10136.0
219.00	1303.0
221.00	53776.0
223.00	17800.0
224.00	176448.0
225.00	43480.0
226.10	4839.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

227.00	65480.0
228.00	9964.0
229.00	14121.0
229.90	2133.0
231.00	6856.0
231.90	1040.0
233.00	1694.0
234.00	4957.0
235.00	5083.0
236.00	3424.0
237.00	6343.0
238.00	1252.0
239.00	2821.0
239.90	2384.0
240.90	3742.0
242.00	9334.0
243.10	9556.0
244.00	136960.0
245.00	18184.0
246.00	26208.0
247.00	5404.0
248.10	1252.0
249.00	4492.0
250.00	1104.0
251.00	1675.0
252.20	1006.0
253.00	3032.0
255.00	670976.0
256.00	95856.0
257.00	7406.0
258.00	38088.0
259.00	6070.0
260.00	1366.0
260.90	1829.0
262.90	354.0
264.10	931.0
265.00	16656.0
265.90	2432.0
266.80	659.0
270.00	486.0
271.00	1677.0
272.00	1780.0
273.00	18728.0
274.00	54840.0
275.00	307712.0
276.00	43408.0
277.00	26240.0
277.90	3709.0
278.90	1086.0
281.00	1205.0
282.10	970.0
283.00	3366.0
284.10	2257.0
285.10	4492.0
286.10	740.0
287.90	491.0
289.10	1104.0
289.90	848.0
290.90	604.0
292.10	1577.0
293.00	6342.0
294.10	1873.0
295.00	2588.0
296.00	80584.0
297.00	11198.0
297.90	684.0
299.10	325.0
301.10	1326.0
302.10	1034.0
303.00	9449.0
303.90	2381.0
305.00	419.0
307.90	930.0
309.10	924.0
310.00	724.0
311.00	446.0
311.90	358.0
312.90	916.0
314.00	4065.0
315.00	7555.0
316.00	4812.0
317.00	1282.0
320.40	605.0
321.00	2668.0
322.10	1343.0
323.00	23320.0
324.00	4569.0
325.10	486.0
325.90	548.0
326.90	4991.0
328.00	2286.0
329.00	921.0
331.90	1214.0
333.00	2741.0
334.00	17928.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

335.00	4488.0
336.10	810.0
338.90	334.0
339.90	339.0
341.10	3941.0
342.00	960.0
346.00	5390.0
346.90	1561.0
350.10	351.0
351.00	377.0
352.00	7297.0
353.00	7834.0
354.00	8300.0
355.10	1956.0
355.80	329.0
359.00	404.0
363.90	594.0
365.00	37752.0
366.00	5307.0
370.00	1010.0
370.90	2242.0
372.00	13165.0
373.00	4253.0
374.00	325.0
376.70	314.0
383.00	3737.0
383.90	1454.0
390.00	1736.0
391.00	1743.0
392.00	1190.0
400.90	800.0
401.90	5652.0
403.00	7121.0
404.00	2883.0
404.90	645.0
415.00	338.0
420.10	335.0
420.90	5810.0
422.00	6806.0
423.00	47632.0
424.00	11917.0
424.90	1350.0
438.80	550.0
441.00	134464.0
442.00	774336.0
443.00	142784.0
444.10	16728.0
445.00	734.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

m/z	Abundance
36.00	1300.0
37.00	3123.0
38.00	7301.0
39.00	37016.0
40.00	3217.0
41.10	4552.0
42.00	2812.0
43.00	5869.0
44.00	17376.0
45.00	3536.0
45.90	1008.0
47.10	716.0
48.00	330.0
49.00	5069.0
50.00	131200.0
51.00	482368.0
52.00	27728.0
52.90	1882.0
54.00	748.0
55.10	4701.0
56.00	13766.0
57.00	33624.0
58.00	1409.0
59.00	674.0
60.10	1036.0
61.00	5906.0
62.00	7004.0
63.00	19976.0
64.00	2655.0
65.00	9452.0
66.00	1018.0
67.20	1082.0
68.10	9660.0
69.00	486656.0
70.00	3722.0
71.00	1187.0
72.10	634.0
73.10	7223.0
74.00	54608.0
75.00	82320.0
76.00	27680.0
77.00	588608.0
78.00	38800.0
79.00	40288.0
80.00	25576.0
81.00	40808.0
82.00	9666.0
83.00	9624.0
83.90	2132.0
85.00	10477.0
86.00	11125.0
87.00	5180.0
88.00	2471.0
89.10	1065.0
91.00	8903.0
91.90	9936.0
93.00	69232.0
94.00	5372.0
95.00	2152.0
96.00	3321.0
97.00	2172.0
98.00	53648.0
99.00	40704.0
100.00	4235.0
101.00	21504.0
102.00	1619.0
103.00	6475.0
104.00	14434.0
105.00	13694.0
106.00	4199.0
107.00	171520.0
108.00	27624.0
109.10	6030.0
110.00	306304.0
111.00	47224.0
112.00	5351.0
113.10	1948.0
113.90	761.0
115.10	1142.0
116.00	10520.0
117.00	160000.0
118.00	12017.0
119.10	1419.0
120.10	2123.0
121.00	1115.0
122.00	11474.0
123.00	17312.0
123.90	8784.0
125.00	7500.0
127.00	553088.0
128.00	42048.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

129.00	211584.0
130.00	18144.0
131.00	3882.0
132.10	2130.0
132.80	1092.0
133.90	6803.0
135.00	19792.0
136.00	7272.0
137.00	9175.0
137.90	1752.0
139.00	1764.0
140.00	2409.0
141.00	28560.0
142.00	9241.0
143.00	6880.0
144.10	2160.0
145.00	1826.0
146.00	4340.0
147.00	15625.0
148.00	39208.0
149.00	9332.0
150.00	2143.0
151.00	4229.0
152.00	2379.0
153.00	9529.0
154.10	7083.0
155.00	16456.0
156.10	24920.0
157.00	6515.0
158.00	4760.0
159.00	5728.0
160.00	10672.0
161.00	14811.0
162.00	4429.0
163.00	2356.0
164.00	1755.0
165.00	11260.0
166.00	8653.0
167.00	75704.0
168.00	41064.0
169.10	5795.0
170.10	2575.0
171.00	2416.0
172.00	6937.0
173.00	7756.0
174.00	13580.0
175.00	25576.0
176.00	7730.0
177.00	14218.0
178.00	4361.0
179.00	52176.0
180.00	36040.0
181.00	17456.0
182.00	2652.0
183.00	2217.0
184.00	4282.0
185.00	27136.0
186.00	191104.0
187.00	55472.0
188.00	5213.0
189.00	13257.0
190.00	1579.0
191.00	5719.0
192.00	17456.0
193.00	19104.0
194.00	4709.0
195.00	1767.0
196.00	43448.0
198.00	1412096.0
198.90	94040.0
200.00	7819.0
201.50	6150.0
203.00	11307.0
204.00	53400.0
205.00	90552.0
206.10	394688.0
207.00	54552.0
208.00	13843.0
209.00	5045.0
210.00	5597.0
211.00	13536.0
211.90	2270.0
212.90	940.0
214.20	672.0
215.00	3920.0
216.00	10072.0
217.00	103472.0
218.00	14931.0
218.90	1503.0
221.00	66528.0
221.90	12143.0
223.00	25048.0
224.00	206080.0
225.00	58824.0
225.90	6321.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

227.00	94200.0
228.00	15015.0
229.00	18376.0
229.90	2264.0
231.00	8781.0
231.90	1137.0
233.00	1729.0
234.00	6178.0
235.00	6167.0
236.00	4270.0
237.00	8824.0
238.10	1015.0
239.00	3861.0
240.00	3340.0
241.00	5337.0
242.00	12346.0
243.00	12889.0
244.00	185664.0
245.10	25608.0
246.00	36720.0
247.00	6013.0
248.00	1361.0
248.90	6632.0
249.70	1230.0
251.00	1620.0
252.00	1832.0
253.00	4764.0
255.00	880064.0
256.00	132288.0
257.00	11157.0
258.00	50928.0
259.00	8383.0
260.00	1444.0
261.00	2239.0
263.10	601.0
264.10	1158.0
265.00	20456.0
265.90	1963.0
266.90	784.0
267.90	534.0
270.00	785.0
270.90	1933.0
272.10	2510.0
273.00	27336.0
274.00	76848.0
275.00	425344.0
276.00	57184.0
277.00	35416.0
278.00	6074.0
279.00	912.0
281.10	1280.0
282.20	738.0
283.00	3629.0
284.00	2886.0
285.00	7359.0
285.90	1158.0
289.10	1454.0
289.90	1332.0
291.00	657.0
292.00	1728.0
293.00	7137.0
294.00	2480.0
295.10	3408.0
296.00	107464.0
297.00	14333.0
298.10	1356.0
299.10	366.0
301.10	1341.0
301.90	2320.0
303.00	13162.0
304.00	2537.0
305.10	644.0
307.20	503.0
308.00	2035.0
309.10	1345.0
310.00	1793.0
311.00	454.0
312.30	439.0
313.00	952.0
314.00	4948.0
315.00	11911.0
316.00	6747.0
317.20	1048.0
320.00	1062.0
321.00	3245.0
322.00	2102.0
323.00	37712.0
324.00	6760.0
325.10	923.0
325.90	1156.0
327.00	8131.0
328.00	3388.0
328.90	597.0
329.80	333.0
332.00	3448.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

333.00	4258.0
334.00	26768.0
335.00	6706.0
335.90	1179.0
339.00	696.0
340.00	1150.0
341.00	5248.0
342.00	1495.0
345.90	8087.0
346.90	2016.0
348.00	374.0
349.90	332.0
350.90	372.0
352.00	11055.0
353.00	9006.0
354.00	12427.0
355.00	2900.0
356.80	363.0
359.00	1059.0
359.90	383.0
361.10	419.0
365.00	56440.0
365.90	7686.0
366.80	809.0
370.00	990.0
370.90	3585.0
372.00	23448.0
373.00	4724.0
373.90	500.0
377.00	449.0
383.10	5522.0
384.00	1594.0
385.00	542.0
390.10	2406.0
391.00	2008.0
392.00	1350.0
394.70	318.0
396.80	388.0
400.70	1069.0
401.00	887.0
402.00	8877.0
403.00	12808.0
404.10	4864.0
404.90	882.0
415.10	626.0
421.00	9150.0
422.00	11982.0
423.00	74056.0
424.00	17432.0
425.00	2049.0
438.10	473.0
439.00	448.0
439.40	451.0
441.00	192192.0
442.00	1215488.0
443.00	248000.0
444.00	22632.0
445.10	1235.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

m/z	Abundance
36.00	1066.0
37.00	2318.0
38.00	4723.0
39.00	26792.0
40.00	3886.0
41.00	2950.0
42.10	3210.0
43.00	7361.0
44.00	17120.0
45.00	2884.0
45.90	1378.0
46.90	715.0
48.00	345.0
49.00	3278.0
50.00	89672.0
51.00	339968.0
52.00	18712.0
53.00	1595.0
53.90	883.0
55.00	3867.0
56.00	11227.0
57.00	24304.0
58.00	1536.0
59.00	1154.0
60.10	1210.0
61.00	4400.0
62.00	5320.0
63.00	13048.0
64.00	2521.0
65.00	7147.0
65.90	793.0
67.00	1411.0
69.00	346816.0
70.00	3733.0
71.20	980.0
71.90	405.0
73.00	5713.0
74.00	38616.0
75.00	57752.0
76.10	20512.0
77.00	415360.0
78.00	28136.0
79.00	29560.0
80.00	18608.0
81.00	30944.0
82.00	7836.0
83.00	6997.0
84.00	2283.0
85.00	6599.0
86.00	8996.0
87.00	4165.0
88.00	2041.0
88.90	992.0
90.10	355.0
91.00	7407.0
91.90	7646.0
93.00	48432.0
94.00	4204.0
94.90	962.0
96.10	2211.0
97.10	1982.0
98.00	41448.0
99.00	28624.0
100.00	2387.0
101.00	16128.0
101.80	814.0
103.00	5643.0
104.00	11478.0
105.00	10834.0
106.10	3343.0
107.00	126568.0
108.00	19280.0
109.00	3678.0
110.00	224768.0
111.00	35624.0
112.00	4209.0
112.90	1606.0
115.00	1366.0
115.90	6861.0
117.00	121568.0
118.00	8702.0
119.00	2065.0
120.00	1682.0
121.10	697.0
122.00	9363.0
123.00	13486.0
124.00	5927.0
125.00	5175.0
127.00	404544.0
128.00	32032.0
129.00	157696.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

130.00	13440.0
131.10	2948.0
132.00	1830.0
132.90	1373.0
134.00	5503.0
135.00	13597.0
136.00	5198.0
137.00	7411.0
137.90	1615.0
139.00	1557.0
140.00	2507.0
141.00	22064.0
142.10	5990.0
143.00	4568.0
143.90	799.0
145.20	1081.0
146.00	4512.0
147.00	12126.0
148.00	27360.0
149.00	6017.0
150.00	1431.0
151.10	3335.0
153.00	7704.0
154.00	6297.0
155.00	13411.0
156.00	21152.0
157.00	4881.0
158.00	4554.0
159.00	4418.0
160.00	7408.0
161.00	11708.0
161.90	3710.0
162.90	1418.0
164.00	1837.0
165.00	8935.0
166.00	7315.0
167.00	59408.0
168.00	31720.0
169.00	4786.0
170.10	1273.0
171.00	2154.0
172.00	4651.0
173.00	7326.0
174.00	11080.0
175.00	21088.0
176.10	7056.0
177.00	9429.0
178.10	3226.0
179.00	39472.0
180.00	26888.0
181.00	13466.0
182.00	2271.0
182.90	1022.0
183.90	3615.0
185.00	18808.0
186.00	149888.0
187.00	46096.0
188.10	4512.0
189.00	8646.0
190.00	2066.0
190.90	5053.0
192.00	14722.0
193.00	16274.0
193.90	3462.0
195.10	1799.0
196.00	35000.0
198.00	1154048.0
199.00	78720.0
200.00	6246.0
201.40	6013.0
202.90	8583.0
204.00	45680.0
205.00	75608.0
206.00	308480.0
207.00	40336.0
207.90	10972.0
209.00	2918.0
210.10	5207.0
211.00	12787.0
211.70	2908.0
213.10	1081.0
214.00	409.0
214.90	3496.0
216.00	6530.0
216.90	85696.0
218.00	9822.0
219.00	1143.0
219.80	841.0
221.00	55816.0
221.80	12979.0
223.00	19040.0
224.00	181312.0
225.00	48096.0
226.10	5120.0
227.00	74560.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

228.00	10574.0
228.90	15176.0
230.10	2305.0
231.00	6283.0
231.90	1122.0
233.00	947.0
233.90	5349.0
235.00	6426.0
236.00	4041.0
237.00	7220.0
238.00	872.0
239.00	2954.0
240.00	2470.0
241.00	4631.0
242.00	10349.0
243.10	10853.0
244.00	150208.0
245.00	18200.0
246.00	29080.0
246.90	5102.0
247.90	1886.0
248.90	5337.0
249.90	1324.0
251.10	1402.0
252.00	1518.0
253.00	3154.0
255.00	799808.0
256.00	107384.0
256.90	7476.0
258.00	41552.0
259.00	6506.0
259.90	1236.0
261.00	1410.0
262.70	432.0
263.70	633.0
264.10	711.0
265.00	16776.0
266.00	2453.0
267.10	505.0
268.00	517.0
269.10	431.0
270.00	1052.0
271.10	1602.0
272.00	2015.0
273.00	23256.0
274.00	60968.0
275.00	366272.0
276.00	45640.0
277.00	26072.0
278.00	5085.0
278.90	1156.0
279.90	608.0
281.00	1734.0
282.00	927.0
283.00	3161.0
284.10	2632.0
285.00	4539.0
286.00	978.0
288.00	481.0
288.90	1080.0
289.90	1702.0
290.80	509.0
291.90	1650.0
293.00	7703.0
294.00	1937.0
295.00	2621.0
296.00	106088.0
297.00	11998.0
298.00	983.0
300.90	1225.0
302.10	1380.0
303.00	11012.0
303.90	3196.0
304.90	1037.0
308.00	1388.0
309.00	1246.0
310.00	1132.0
310.90	356.0
311.80	313.0
312.90	1294.0
314.00	5763.0
315.00	11657.0
316.00	5184.0
317.00	1232.0
320.10	542.0
320.90	2852.0
322.00	1531.0
323.00	36696.0
324.00	5125.0
325.00	643.0
325.70	707.0
326.00	805.0
327.00	6849.0
328.00	3066.0
329.00	781.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219

330.90	329.0
332.00	2545.0
333.10	3195.0
334.00	22456.0
335.00	4868.0
335.80	512.0
340.00	513.0
341.00	4323.0
342.00	1645.0
346.00	7308.0
347.00	1204.0
350.00	410.0
350.90	419.0
352.00	11615.0
353.00	7301.0
354.00	10652.0
355.00	2660.0
359.00	1015.0
365.00	49424.0
366.00	6723.0
366.90	368.0
370.00	1186.0
370.90	2140.0
372.00	18056.0
373.00	5035.0
373.90	614.0
376.70	497.0
383.00	5982.0
384.00	1920.0
384.90	704.0
389.90	2502.0
391.00	1860.0
392.00	1329.0
401.00	1599.0
402.00	9035.0
403.00	13009.0
404.00	3539.0
404.90	863.0
414.90	954.0
416.10	356.0
420.90	9750.0
422.00	10554.0
423.00	77296.0
424.00	14852.0
425.10	1680.0
438.80	526.0
441.00	200320.0
442.00	1160704.0
443.00	250432.0
444.00	21144.0
444.90	1355.0

Instrument :
BNA_N
ClientSampleId :
DFTPP219