

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122818\
 Data File : BN004264.D
 Acq On : 28 Dec 2018 16:11
 Operator : JU/SJ
 Sample : SSTD3.257
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.257

Quant Time: Dec 28 16:46:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN122818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 28 15:21:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	2166	0.40	ng/ul	0.00
2) Naphthalene-d8	10.62	136	9545	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.46	164	5288	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	11255	0.40	ng/ul	-0.01
16) Chrysene-d12	21.40	240	14521	0.40	ng/ul	0.00
20) Perylene-d12	23.73	264	12109	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.87	266	7052	3.116	ng/ul	Ovalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122818\
Data File : BN004264.D
Acq On : 28 Dec 2018 16:11
Operator : JU/SJ
Sample : SST3.257
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.257

Quant Time: Dec 28 16:46:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN122818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 28 15:21:10 2018
Response via : Initial Calibration

