

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122818\
 Data File : BN004276.D
 Acq On : 28 Dec 2018 23:11
 Operator : JU/SJ
 Sample : J6541-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9F71

Manual Integrations
 APPROVED

Sohil
 1/2/2019 3:39:44 PM

Quant Time: Dec 31 09:26:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN122818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 28 22:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	2547	0.40	ng/ul	0.00
2) Naphthalene-d8	10.61	136	11691	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.46	164	9260m	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	16437m	0.40	ng/ul	-0.01
16) Chrysene-d12	21.40	240	17594m	0.40	ng/ul	0.00
20) Perylene-d12	23.71	264	14843m	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.20	152	4660	0.26	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	11430m	0.24	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.66	128	65980	2.004	ng/ul	100
5) 2-Methylnaphthalene	12.28	142	7650	0.335	ng/ul	99
7) Acenaphthylene	14.20	152	1821m	0.039	ng/ul	
8) Acenaphthene	14.52	153	3009	0.081	ng/ul	93
9) Fluorene	15.51	166	6462	0.153	ng/ul#	97
12) Phenanthrene	17.25	178	9493m	0.178	ng/ul	
13) Anthracene	17.34	178	4252m	0.088	ng/ul	
15) Fluoranthene	19.27	202	2968m	0.046	ng/ul	
17) Pyrene	19.63	202	27941	0.464	ng/ul	97
18) Benzo(a)anthracene	21.38	228	1982m	0.033	ng/ul	
19) Chrysene	21.43	228	3733m	0.058	ng/ul	
26) Benzo(g,h,i)perylene	26.81	276	1311	0.028	ng/ul#	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122818\
Data File : BN004276.D
Acq On : 28 Dec 2018 23:11
Operator : JU/SJ
Sample : J6541-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F71

Quant Time: Dec 31 09:26:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN122818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 28 22:16:51 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
1/2/2019 3:39:44 PM

