

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122921\
 Data File : BN018272.D
 Acq On : 29 Dec 2021 22:26
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 30 02:24:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 29 16:08:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.344	14.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.432	-8.0	113	0.00
4 S	2-Fluorophenol	0.400	0.456	-14.0	119	0.00
5 S	Phenol-d6	0.400	0.459	-14.8	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.405	-1.3	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.448	-12.0	121	0.00
9	Naphthalene	0.400	0.388	3.0	99	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	102	0.00
12	2-Methylnaphthalene	0.400	0.406	-1.5	102	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.550	-37.5#	165	-0.01
15 S	2-Fluorobiphenyl	0.400	0.388	3.0	101	0.00
16	Acenaphthylene	0.400	0.405	-1.3	119	0.00
17	Acenaphthene	0.400	0.386	3.5	104	0.00
18	Fluorene	0.400	0.409	-2.2	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	110	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.316	21.0	77	0.00
21	4-Bromophenyl-phenylether	0.400	0.381	4.8	114	0.00
22	Hexachlorobenzene	0.400	0.380	5.0	108	0.00
23	Atrazine	0.400	0.439	-9.7	155	0.00
24	Pentachlorophenol	0.400	0.566	-41.5#	214	0.00
25	Phenanthrene	0.400	0.381	4.8	109	0.00
26	Anthracene	0.400	0.382	4.5	125	0.00
27 SURR	Fluoranthene-d10	0.400	0.403	-0.8	121	0.00
28	Fluoranthene	0.400	0.408	-2.0	117	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	135	0.00
30	Pyrene	0.400	0.361	9.8	120	0.00
31 S	Terphenyl-d14	0.400	0.373	6.8	125	0.00
32	Benzo(a)anthracene	0.400	0.412	-3.0	164	0.00
33	Chrysene	0.400	0.381	4.8	133	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.534	-33.5#	224	0.00
35 I	Perylene-d12	0.400	0.400	0.0	143	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.442	-10.5	177	0.00
37	Benzo(b)fluoranthene	0.400	0.392	2.0	148	0.00
38	Benzo(k)fluoranthene	0.400	0.392	2.0	148	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	156	0.00
40	Dibenzo(a,h)anthracene	0.400	0.466	-16.5	199	0.00
41	Benzo(g,h,i)perylene	0.400	0.404	-1.0	157	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4