

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN123021\
 Data File : BN018291.D
 Acq On : 30 Dec 2021 11:55
 Operator : CG/JU
 Sample : PB141692BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB141692BL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/03/2022
 Supervised By :mohammad ahmed 01/06/2022

Quant Time: Dec 31 08:21:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 29 16:08:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	33677	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	125853	0.400	ng	0.00
13) Acenaphthene-d10	14.269	164	75283	0.400	ng	0.00
19) Phenanthrene-d10	17.018	188	118217	0.400	ng	0.01
29) Chrysene-d12	21.206	240	49215	0.400	ng	# 0.01
35) Perylene-d12	23.433	264	46127	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	20956	0.311	ng	0.03
5) Phenol-d6	6.859	99	9011m	0.138	ng	0.03
8) Nitrobenzene-d5	8.797	82	23058	0.384	ng	0.01
11) 2-Methylnaphthalene-d10	12.016	152	63311	0.299	ng	0.00
14) 2,4,6-Tribromophenol	15.779	330	3409	0.389	ng	0.01
15) 2-Fluorobiphenyl	12.899	172	81854	0.287	ng	0.00
27) Fluoranthene-d10	19.043	212	117513	0.312	ng	0.00
31) Terphenyl-d14	19.654	244	69571	0.556	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN123021\
Data File : BN018291.D
Acq On : 30 Dec 2021 11:55
Operator : CG/JU
Sample : PB141692BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB141692BL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 01/03/2022
Supervised By :mohammad ahmed 01/06/2022

Quant Time: Dec 31 08:21:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 29 16:08:47 2021
Response via : Initial Calibration

