

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010322\
 Data File : BN018314.D
 Acq On : 03 Jan 2022 19:04
 Operator : CG/JU
 Sample : PB141515BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK515

Quant Time: Jan 04 00:05:07 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 28 00:34:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	2276	0.400	ng/ul	0.00
4) Naphthalene-d8	10.387	136	8472	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.248	164	6572	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	16.990	188	16229	0.400	ng/ul	0.00
17) Chrysene-d12	21.190	240	18849	0.400	ng/ul	0.00
23) Perylene-d12	23.405	264	29433	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.067	96	12046	5.669	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.993	152	4297	0.292	ng/ul	0.00
18) Fluoranthene-d10	19.026	212	17150	0.307	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010322\
Data File : BN018314.D
Acq On : 03 Jan 2022 19:04
Operator : CG/JU
Sample : PB141515BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK515

Quant Time: Jan 04 00:05:07 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 28 00:34:45 2021
Response via : Initial Calibration

