

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010322\
 Data File : BN018317.D
 Acq On : 03 Jan 2022 20:51
 Operator : CG/JU
 Sample : M5091-14
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG3C4

Quant Time: Jan 04 00:06:09 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 28 00:34:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	1922	0.400	ng/ul	0.00
4) Naphthalene-d8	10.383	136	7322	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.250	164	6046	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	16.987	188	15923	0.400	ng/ul	0.00
17) Chrysene-d12	21.186	240	21011	0.400	ng/ul	0.00
23) Perylene-d12	23.401	264	29546	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.067	96	9454	5.269	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.989	152	5061	0.398	ng/ul	0.00
18) Fluoranthene-d10	19.023	212	26020	0.418	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.101	88	808	0.405	ng/ul#	50
15) Phenanthrene	17.029	178	2963	0.056	ng/ul#	88
16) Anthracene	17.118	178	956	0.020	ng/ul#	55
19) Fluoranthene	19.055	202	2061	0.025	ng/ul#	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN010322\
Data File : BN018317.D
Acq On : 03 Jan 2022 20:51
Operator : CG/JU
Sample : M5091-14
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BG3C4

Quant Time: Jan 04 00:06:09 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 28 00:34:45 2021
Response via : Initial Calibration

