

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011922\
 Data File : BN018429.D
 Acq On : 19 Jan 2022 22:48
 Operator : CG/JU
 Sample : PB142140BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS140

Quant Time: Jan 20 02:40:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 07 01:31:51 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 01/20/2022
 Supervised By :mohammad ahmed 01/21/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.587	152	324180	20.000	ng/u1	0.00
20) Naphthalene-d8	10.369	136	1440892	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.234	164	862652	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.975	188	1799948	20.000	ng/u1	0.00
79) Chrysene-d12	21.174	240	1490166	20.000	ng/u1	# 0.00
88) Perylene-d12	23.392	264	1427718	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.064	96	49164m	6.096	ng/uL	0.00
4) Pyridine-d5	3.464	84	802555	31.800	ng/u1	0.00
7) Phenol-d5	6.775	99	1171157	37.172	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.928	67	693635	34.561	ng/u1	0.00
11) 2-Chlorophenol-d4	7.122	132	909257	39.399	ng/u1	0.00
15) 4-Methylphenol-d8	8.310	113	948300	38.132	ng/u1	0.00
21) Nitrobenzene-d5	8.740	128	443024	41.281	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	494630	43.118	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.999	165	868872	41.689	ng/u1	0.00
31) 4-Chloroaniline-d4	10.510	131	1057355	33.011	ng/u1	0.00
46) Dimethylphthalate-d6	13.651	166	2709350	43.241	ng/u1	0.00
49) Acenaphthylene-d8	13.922	160	3176748	40.106	ng/u1	0.00
54) 4-Nitrophenol-d4	14.451	143	482231	40.331	ng/u1	0.00
60) Fluorene-d10	15.228	176	2097051	39.669	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.363	200	411350	40.388	ng/u1	0.00
73) Anthracene-d10	17.075	188	3072946	36.862	ng/u1	0.00
81) Pyrene-d10	19.375	212	3763508	44.026	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.251	264	3267722	45.988	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.093	88	105482	11.626	ng/uL#	96
5) Pyridine	3.481	79	813156	30.980	ng/u1	93
6) Benzaldehyde	6.728	77	650864	32.526	ng/u1	98
8) Phenol	6.805	94	1165714	35.500	ng/u1#	87
10) Bis(2-Chloroethyl)ether	7.022	93	917682	34.662	ng/u1	95
12) 2-Chlorophenol	7.157	128	905536	37.183	ng/u1	93
13) 2-Methylphenol	8.040	108	896021	36.271	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.122	45	1081947	30.794	ng/u1#	94
16) Acetophenone	8.404	105	1311816	34.437	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.404	70	707750	34.005	ng/u1#	91
18) 4-Methylphenol	8.375	108	934686	35.262	ng/u1	98
19) Hexachloroethane	8.657	117	395852	39.191	ng/u1#	76
22) Nitrobenzene	8.781	77	1115359	39.340	ng/u1	99
23) Isophorone	9.304	82	2168365	38.081	ng/u1	98
25) 2-Nitrophenol	9.493	139	520287	40.741	ng/u1	96
26) 2,4-Dimethylphenol	9.563	107	1109670	39.181	ng/u1	94
27) Bis(2-Chloroethoxy)met...	9.793	93	1278241	36.087	ng/u1	100
29) 2,4-Dichlorophenol	10.022	162	848838	40.210	ng/u1	98
30) Naphthalene	10.416	128	2809715	35.640	ng/u1	99
32) 4-Chloroaniline	10.534	127	1021085	31.214	ng/u1	100
33) Hexachlorobutadiene	10.716	225	491439	42.691	ng/u1	99
34) Caprolactam	11.281	113	315734m	38.048	ng/u1	
35) 4-Chloro-3-methylphenol	11.675	107	1100680	41.663	ng/u1	91
36) 2-Methylnaphthalene	12.040	142	1939764	36.876	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011922\
 Data File : BN018429.D
 Acq On : 19 Jan 2022 22:48
 Operator : CG/JU
 Sample : PB142140BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS140


Quant Time: Jan 20 02:40:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 07 01:31:51 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 01/20/2022
 Supervised By :mohammad ahmed 01/21/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.257	142	1951642	35.653	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.416	216	913042	41.264	ng/ul#	94
40) Hexachlorocyclopentadiene	12.404	237	528119	41.557	ng/ul	96
41) 2,4,6-Trichlorophenol	12.663	196	657169	43.791	ng/ul	95
42) 2,4,5-Trichlorophenol	12.734	196	713539	44.854	ng/ul	96
43) 1,1'-Biphenyl	13.063	154	2617609	38.963	ng/ul	98
44) 2-Chloronaphthalene	13.104	162	1976885	39.380	ng/ul	93
45) 2-Nitroaniline	13.310	65	738795	47.021	ng/ul	98
47) Dimethylphthalate	13.704	163	2614397	41.111	ng/ul	99
48) 2,6-Dinitrotoluene	13.816	165	549478	45.595	ng/ul#	85
50) Acenaphthylene	13.951	152	3005032	36.001	ng/ul	98
51) 3-Nitroaniline	14.139	138	498720	36.811	ng/ul	96
52) Acenaphthene	14.298	153	1994412	36.261	ng/ul	99
53) 2,4-Dinitrophenol	14.351	184	282885	41.676	ng/ul#	74
55) 4-Nitrophenol	14.469	109	494089	48.368	ng/ul#	78
56) Dibenzofuran	14.634	168	2897091	37.983	ng/ul#	86
57) 2,4-Dinitrotoluene	14.604	165	773502	44.610	ng/ul#	79
58) 2,3,4,6-Tetrachlorophenol	14.869	232	565234	44.078	ng/ul#	77
59) Diethylphthalate	15.075	149	2805114	42.727	ng/ul	96
61) Fluorene	15.286	166	2146947	36.550	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.281	204	989818	37.686	ng/ul#	77
63) 4-Nitroaniline	15.310	138	496911	41.118	ng/ul#	82
66) 4,6-Dinitro-2-methylph...	15.375	198	404670	38.979	ng/ul#	83
67) N-Nitrosodiphenylamine	15.498	169	2015603	36.128	ng/ul	97
68) 4-Bromophenyl-phenylether	16.175	248	698284	40.006	ng/ul#	80
69) Hexachlorobenzene	16.292	284	841991	40.241	ng/ul#	89
70) Atrazine	16.457	200	767739	41.964	ng/ul	92
71) Pentachlorophenol	16.633	266	483495	40.482	ng/ul	97
72) Phenanthrene	17.022	178	3742485	36.589	ng/ul	98
74) Anthracene	17.110	178	3462747	34.103	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.028	216	949802	36.956	ng/uL	99
76) Pentachlorobenzene	14.563	250	926873	38.573	ng/uL	97
77) Carbazole	17.380	167	3404343	38.092	ng/ul#	95
78) Di-n-butylphthalate	17.963	149	4825900	43.926	ng/ul#	98
80) Fluoranthene	19.045	202	4524643	43.668	ng/ul	97
82) Pyrene	19.404	202	4192210	39.710	ng/ul#	94
83) Butylbenzylphthalate	20.322	149	2272020	51.915	ng/ul#	95
84) 3,3'-Dichlorobenzidine	21.098	252	1308896	45.567	ng/ul#	98
85) Benzo(a)anthracene	21.157	228	4003715	41.045	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.110	149	3369110	50.613	ng/ul#	94
87) Chrysene	21.210	228	3703442	38.552	ng/ul	97
89) Di-n-octyl phthalate	21.986	149	5781737	56.515	ng/ul	100
90) Benzo(b)fluoranthene	22.727	252	4316439m	47.160	ng/ul	
91) Benzo(k)fluoranthene	22.774	252	3822901	45.038	ng/ul#	98
93) Benzo(a)pyrene	23.298	252	3682176	42.187	ng/ul	94
94) Indeno(1,2,3-cd)pyrene	25.651	276	3898565	40.387	ng/ul#	91
95) Dibenzo(a,h)anthracene	25.662	278	3207023	39.401	ng/ul#	97
96) Benzo(g,h,i)perylene	26.333	276	3252685	40.215	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.626 ng/uL
RT: 3.093 min Scan# 18
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion: 88 Resp: 10548
Ion Ratio Lower Upper
88 100
43 33.4 21.9 32.9
58 71.0 56.3 84.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#5
Pyridine
Concen: 30.980 ng/uL
RT: 3.481 min Scan# 84
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion: 79 Resp: 813156
Ion Ratio Lower Upper
79 100
52 72.2 62.9 94.3
51 38.4 27.8 41.8





#6
Benzaldehyde
Concen: 32.526 ng/u1
RT: 6.728 min Scan# 61
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion: 77 Resp: 650864
Ion Ratio Lower Upper
77 100
105 91.5 73.5 110.3
106 90.3 69.9 104.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#8
Phenol
Concen: 35.500 ng/u1
RT: 6.805 min Scan# 649
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion: 94 Resp: 1165714
Ion Ratio Lower Upper
94 100
65 32.8 21.8 32.6#
66 46.7 30.2 45.4#





#10
Bis(2-Chloroethyl)ether
Concen: 34.662 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion: 93 Resp: 91768
Ion Ratio Lower Upper
93 100
63 73.8 63.4 95.0
95 31.8 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#12
2-Chlorophenol
Concen: 37.183 ng/ul
RT: 7.157 min Scan# 709
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48



Tgt Ion:128 Resp: 905536
Ion Ratio Lower Upper
128 100
64 48.8 44.2 66.2
130 33.8 25.0 37.6





#13

2-Methylphenol

Concen: 36.271 ng/ul

RT: 8.040 min Scan# 8

Delta R.T. -0.006 min

Lab File: BN018429.D

Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :

SLCS140

Tgt Ion:108 Resp: 89602

Ion Ratio Lower Upper





108 100

107 87.9 69.5 104.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.794 ng/ul

RT: 8.122 min Scan# 873

Delta R.T. -0.012 min

Lab File: BN018429.D

Acq: 19 Jan 2022 22:48




Tgt Ion: 45 Resp: 1081947


Ion Ratio Lower Upper

45 100

77 17.1 11.9 17.9


79 13.7 8.6 12.8#





#16
Acetophenone
Concen: 34.437 ng/ul
RT: 8.404 min Scan# 921
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:105 Resp: 1311810
Ion Ratio Lower Upper
105 100
77 85.8 64.8 97.2
51 31.4 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#17
N-Nitroso-di-n-propylamine
Concen: 34.005 ng/ul
RT: 8.404 min Scan# 921
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48



Tgt Ion: 70 Resp: 707750
Ion Ratio Lower Upper
70 100
42 60.0 42.4 63.6
101 9.7 6.6 10.0
130 20.5 13.6 20.4





#18

4-Methylphenol

Concen: 35.262 ng/ul

RT: 8.375 min Scan# 915

Delta R.T. -0.000 min

Lab File: BN018429.D

Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :

SLCS140

Tgt Ion:108 Resp: 934680

Ion Ratio Lower Upper





108 100

107 114.3 89.4 134.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#19

Hexachloroethane

Concen: 39.191 ng/ul

RT: 8.657 min Scan# 964

Delta R.T. -0.006 min

Lab File: BN018429.D

Acq: 19 Jan 2022 22:48




Tgt Ion:117 Resp: 395852


Ion Ratio Lower Upper

117 100

201 81.6 85.9 128.9#


199 50.3 54.5 81.7#





#22
Nitrobenzene
Concen: 39.340 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion: 77 Resp: 1115359
Ion Ratio Lower Upper
77 100
123 40.6 32.0 48.0
65 14.6 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#23
Isophorone
Concen: 38.081 ng/ul
RT: 9.304 min Scan# 1074
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion: 82 Resp: 2168365
Ion Ratio Lower Upper
82 100
95 8.0 5.7 8.5
138 15.6 11.9 17.9





#25
2-Nitrophenol
Concen: 40.741 ng/ul
RT: 9.493 min Scan# 1105
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:139 Resp: 52028
Ion Ratio Lower Upper
139 100
65 59.1 47.7 71.5
109 38.9 27.0 40.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#26
2,4-Dimethylphenol
Concen: 39.181 ng/ul
RT: 9.563 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:107 Resp: 1109670
Ion Ratio Lower Upper
107 100
121 47.9 38.9 58.3
122 78.4 68.6 103.0





#27
Bis(2-Chloroethoxy)methane
Concen: 36.087 ng/ul
RT: 9.793 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion: 93 Resp: 127824
Ion Ratio Lower Upper
93 100
95 32.1 25.7 38.5
123 10.5 8.3 12.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#29
2,4-Dichlorophenol
Concen: 40.210 ng/ul
RT: 10.022 min Scan# 1196
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:162 Resp: 848838
Ion Ratio Lower Upper
162 100
164 63.2 51.8 77.6
98 40.2 31.0 46.6





#30
Naphthalene
Concen: 35.640 ng/ul
RT: 10.416 min Scan# 1263
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:128 Resp: 2809715
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 13.7 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#32
4-Chloroaniline
Concen: 31.214 ng/ul
RT: 10.534 min Scan# 1283
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Tgt Ion:127 Resp: 1021085
Ion Ratio Lower Upper
127 100
129 31.3 25.0 37.6






#33
Hexachlorobutadiene
Concen: 42.691 ng/ul
RT: 10.716 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :

SLCS140






Tgt Ion:225 Resp: 491439
Ion Ratio Lower Upper
225 100
223 62.8 49.7 74.5
227 63.1 49.5 74.3


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#34
Caprolactam
Concen: 38.048 ng/ul m
RT: 11.281 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:113 Resp: 315734
Ion Ratio Lower Upper
113 100
55 160.2 147.4 221.2
56 126.3 107.6 161.4





#35
4-Chloro-3-methylphenol
Concen: 41.663 ng/ul
RT: 11.675 min Scan# 1477
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:107 Resp: 1100680
Ion Ratio Lower Upper
107 100
144 23.2 19.4 29.0
142 67.7 62.3 93.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#36
2-Methylnaphthalene
Concen: 36.876 ng/ul
RT: 12.040 min Scan# 1539
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:142 Resp: 1939764
Ion Ratio Lower Upper
142 100
141 90.0 70.2 105.4





#37
1-Methylnaphthalene
Concen: 35.653 ng/ul
RT: 12.257 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:142 Resp: 195164
Ion Ratio Lower Upper
142 100
141 90.8 72.2 108.4
116 4.7 3.0 4.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 41.264 ng/ul
RT: 12.416 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:216 Resp: 913042
Ion Ratio Lower Upper
216 100
214 79.6 64.4 96.6
179 24.1 15.1 22.7#
108 26.3 12.0 18.0#





#40
Hexachlorocyclopentadiene
Concen: 41.557 ng/ul
RT: 12.404 min Scan# 1601
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:237 Resp: 528119
Ion Ratio Lower Upper
237 100
235 64.5 49.1 73.7
272 12.4 10.6 16.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#41
2,4,6-Trichlorophenol
Concen: 43.791 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48



Tgt Ion:196 Resp: 657169
Ion Ratio Lower Upper
196 100
198 98.9 74.5 111.7
200 30.5 24.8 37.2






#42
2,4,5-Trichlorophenol
Concen: 44.854 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :

SLCS140






Tgt Ion:196 Resp: 713539
Ion Ratio Lower Upper
196 100
198 89.0 75.3 112.9
200 30.8 24.2 36.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#43
1,1'-Biphenyl
Concen: 38.963 ng/ul
RT: 13.063 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:154 Resp: 2617609
Ion Ratio Lower Upper
154 100
153 41.2 32.4 48.6
76 13.9 12.0 18.0





#44
2-Chloronaphthalene
Concen: 39.380 ng/ul
RT: 13.104 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:162 Resp: 1976885
Ion Ratio Lower Upper
162 100
164 32.3 26.8 40.2
127 43.0 28.8 43.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#45
2-Nitroaniline
Concen: 47.021 ng/ul
RT: 13.310 min Scan# 1755
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion: 65 Resp: 738795
Ion Ratio Lower Upper
65 100
92 66.4 53.0 79.6
138 88.3 73.8 110.8





#47
Dimethylphthalate
Concen: 41.111 ng/ul
RT: 13.704 min Scan# 1821
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:163 Resp: 261439
Ion Ratio Lower Upper
163 100
194 3.8 3.6 5.4
164 10.0 8.2 12.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#48
2,6-Dinitrotoluene
Concen: 45.595 ng/ul
RT: 13.816 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Tgt Ion:165 Resp: 549478
Ion Ratio Lower Upper
165 100
89 75.3 49.3 73.9#
121 23.5 16.8 25.2





#50
Acenaphthylene
Concen: 36.001 ng/ul
RT: 13.951 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :





SLCS140

Tgt Ion:152 Resp: 300503
Ion Ratio Lower Upper
152 100
151 21.1 16.2 24.2
153 13.6 10.6 15.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#51
3-Nitroaniline
Concen: 36.811 ng/ul
RT: 14.139 min Scan# 1896
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:138 Resp: 498720
Ion Ratio Lower Upper
138 100
108 12.9 10.6 15.8
92 129.8 107.9 161.9





#52
Acenaphthene
Concen: 36.261 ng/ul
RT: 14.298 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:153 Resp: 1994412
Ion Ratio Lower Upper
153 100
152 46.4 37.5 56.3
154 86.0 69.8 104.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#53
2,4-Dinitrophenol
Concen: 41.676 ng/ul
RT: 14.351 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:184 Resp: 282885
Ion Ratio Lower Upper
184 100
63 95.3 58.6 87.8#
154 76.1 46.2 69.2#





#55
4-Nitrophenol
Concen: 48.368 ng/ul
RT: 14.469 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:109 Resp: 494089
Ion Ratio Lower Upper
109 100
139 103.1 104.0 156.0
65 121.6 117.5 176.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#56
Dibenzofuran
Concen: 37.983 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48



Tgt Ion:168 Resp: 2897091
Ion Ratio Lower Upper
168 100
139 42.9 27.8 41.6#









#57
2,4-Dinitrotoluene
Concen: 44.610 ng/ul
RT: 14.604 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :
BNA_N
ClientSampleId :
SLCS140

Tgt Ion:165 Resp: 773502
Ion Ratio Lower Upper
165 100
63 59.4 36.0 54.0#
182 2.9 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 44.078 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Tgt Ion:232 Resp: 565234
Ion Ratio Lower Upper
232 100
131 57.4 31.4 47.0#
130 3.3 1.8 2.6#
166 34.5 21.8 32.8#





#59
Diethylphthalate
Concen: 42.727 ng/ul
RT: 15.075 min Scan# 2054
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :





SLCS140

Tgt Ion:149 Resp: 2805114
Ion Ratio Lower Upper
149 100
177 20.7 18.6 28.0
150 11.7 10.0 15.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#61
Fluorene
Concen: 36.550 ng/ul
RT: 15.286 min Scan# 2091
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:166 Resp: 2146947
Ion Ratio Lower Upper
166 100
165 104.8 79.0 118.4
167 13.9 11.0 16.4





#62
4-Chlorophenyl-phenylether
Concen: 37.686 ng/ul
RT: 15.281 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument : BNA_N
ClientSampleId : SLCS140



Tgt Ion:204 Resp: 989818
Ion Ratio Lower Upper
204 100
206 30.4 26.8 40.2
141 78.5 43.1 64.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#63
4-Nitroaniline
Concen: 41.118 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:138 Resp: 496911
Ion Ratio Lower Upper
138 100
92 57.4 45.6 68.4
108 112.7 68.0 102.0





#66
4,6-Dinitro-2-methylphenol
Concen: 38.979 ng/ul
RT: 15.375 min Scan# 2106
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:198 Resp: 404670
Ion Ratio Lower Upper
198 100
182 3.9 3.4 5.2
77 36.7 21.1 31.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#67
N-Nitrosodiphenylamine
Concen: 36.128 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Tgt Ion:169 Resp: 2015603
Ion Ratio Lower Upper
169 100
168 66.2 52.2 78.4
167 37.5 27.4 41.2






#68
4-Bromophenyl-phenylether
Concen: 40.006 ng/ul
RT: 16.175 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :

SLCS140






Tgt Ion:248 Resp: 698284
Ion Ratio Lower Upper
248 100
250 95.8 72.3 108.5
141 99.0 53.7 80.5


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#69
Hexachlorobenzene
Concen: 40.241 ng/ul
RT: 16.292 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:284 Resp: 841991
Ion Ratio Lower Upper
284 100
142 42.1 24.3 36.5#
249 29.2 23.8 35.6





#70
Atrazine
Concen: 41.964 ng/ul
RT: 16.457 min Scan# 2290
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:200 Resp: 767739
Ion Ratio Lower Upper
200 100
173 27.5 20.2 30.2
215 43.2 39.7 59.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#71
Pentachlorophenol
Concen: 40.482 ng/ul
RT: 16.633 min Scan# 2320
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:266 Resp: 483495
Ion Ratio Lower Upper
266 100
264 61.3 52.0 78.0
268 66.1 53.5 80.3





#72
Phenanthrene
Concen: 36.589 ng/u1
RT: 17.022 min Scan# 2386
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:178 Resp: 3742485
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.4
176 19.3 14.6 21.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#74
Anthracene
Concen: 34.103 ng/u1
RT: 17.110 min Scan# 2401
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:178 Resp: 3462747
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.2 14.7 22.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.956 ng/uL
RT: 13.028 min Scan# 1707
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:216 Resp: 949802
Ion Ratio Lower Upper
216 100
214 79.3 63.1 94.7
218 48.7 39.8 59.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#76
Pentachlorobenzene
Concen: 38.573 ng/uL
RT: 14.563 min Scan# 1968
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:250 Resp: 926873
Ion Ratio Lower Upper
250 100
248 59.6 49.5 74.3
252 64.1 52.7 79.1





#77
 Carbazole
 Concen: 38.092 ng/uI
 RT: 17.380 min Scan# 2447
 Delta R.T. -0.006 min
 Lab File: BN018429.D
 Acq: 19 Jan 2022 22:48

Instrument :
 BNA_N
 ClientSampleId :
 SLCS140






Tgt Ion:167 Resp: 340434


Ion	Ratio	Lower	Upper
167	100		
166	22.7	16.9	25.3
139	15.1	9.6	14.4

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 01/20/2022
 Supervised By :mohammad ahmed 01/21/2022






#78
 Di-n-butylphthalate
 Concen: 43.926 ng/uI
 RT: 17.963 min Scan# 2546
 Delta R.T. -0.006 min
 Lab File: BN018429.D
 Acq: 19 Jan 2022 22:48



Tgt Ion:149 Resp: 4825900


Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.3	10.9
104	6.4	4.2	6.2





#80
Fluoranthene
Concen: 43.668 ng/u1
RT: 19.045 min Scan# 2729
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:202 Resp: 452464
Ion Ratio Lower Upper
202 100
101 11.7 8.5 12.7
100 9.0 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#82
Pyrene
Concen: 39.710 ng/u1
RT: 19.404 min Scan# 2791
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:202 Resp: 4192210
Ion Ratio Lower Upper
202 100
101 14.2 9.6 14.4
100 11.9 7.8 11.8#





#83
Butylbenzylphthalate
Concen: 51.915 ng/ul
RT: 20.322 min Scan# 2948
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:149 Resp: 2272020
Ion Ratio Lower Upper
149 100
91 81.9 63.7 95.5
206 18.1 19.1 28.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#84
3,3'-Dichlorobenzidine
Concen: 45.567 ng/ul
RT: 21.098 min Scan# 3079
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:252 Resp: 1308896
Ion Ratio Lower Upper
252 100
254 65.5 51.5 77.3
126 13.2 7.9 11.9





#85
Benzo(a)anthracene
Concen: 41.045 ng/ul
RT: 21.157 min Scan# 3089
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument : BNA_N
ClientSampleId : BN018429.D
SLCS140



Tgt Ion:228 Resp: 4003715
Ion Ratio Lower Upper
228 100
229 20.5 15.4 23.2
226 25.9 21.0 31.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 50.613 ng/ul
RT: 21.110 min Scan# 3081
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:149 Resp: 3369110
Ion Ratio Lower Upper
149 100
167 25.8 23.4 35.0
279 3.5 4.1 6.1#





#87
Chrysene
Concen: 38.552 ng/u1
RT: 21.210 min Scan# 3098
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140



Tgt Ion:228 Resp: 370344
Ion Ratio Lower Upper
228 100
226 30.5 23.5 35.3
229 20.6 15.0 22.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#89
Di-n-octyl phthalate
Concen: 56.515 ng/u1
RT: 21.986 min Scan# 3230
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48


Tgt Ion:149 Resp: 5781737





#90
Benzo(b)fluoranthene
Concen: 47.160 ng/ul m
RT: 22.727 min Scan# 3356
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument :
BNA_N
ClientSampleId :
SLCS140




Tgt Ion:252 Resp: 4316439
Ion Ratio Lower Upper
252 100
253 21.3 17.8 26.6
125 11.4 7.2 10.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#91
Benzo(k)fluoranthene
Concen: 45.038 ng/ul
RT: 22.774 min Scan# 3364
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Tgt Ion:252 Resp: 3822901
Ion Ratio Lower Upper
252 100
253 21.3 17.3 25.9
125 10.5 6.9 10.3





#93
Benzo(a)pyrene
Concen: 42.187 ng/ul
RT: 23.298 min Scan# 3453
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48




Instrument : BNA_N
ClientSampleId : SLCS140




Tgt Ion:252 Resp: 3682170
Ion Ratio Lower Upper
252 100
253 21.1 19.3 28.9
125 11.9 8.0 12.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 40.387 ng/ul
RT: 25.651 min Scan# 3853
Delta R.T. -0.000 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48



Tgt Ion:276 Resp: 3898565
Ion Ratio Lower Upper
276 100
138 28.8 16.6 24.8#
277 24.8 19.1 28.7





#95
Dibenzo(a,h)anthracene
Concen: 39.401 ng/ul
RT: 25.662 min Scan# 3855
Delta R.T. -0.006 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Instrument :

BNA_N

ClientSampleId :





SLCS140

Tgt Ion:278 Resp: 320702
Ion Ratio Lower Upper
278 100
139 17.6 11.7 17.5
279 23.7 19.1 28.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 01/20/2022
Supervised By :mohammad ahmed 01/21/2022



#96
Benzo(g,h,i)perylene
Concen: 40.215 ng/ul
RT: 26.333 min Scan# 3969
Delta R.T. -0.012 min
Lab File: BN018429.D
Acq: 19 Jan 2022 22:48

Tgt Ion:276 Resp: 3252685
Ion Ratio Lower Upper
276 100
138 23.3 15.6 23.4
277 22.6 17.8 26.6

