

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011923\
 Data File : BN023716.D
 Acq On : 19 Jan 2023 13:23
 Operator : CG/JU
 Sample : PB150214BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB150214BL

Quant Time: Jan 20 04:16:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 04:14:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.006	152	5279	0.400	ng	0.01
7) Naphthalene-d8	10.829	136	13588	0.400	ng	0.01
13) Acenaphthene-d10	14.656	164	6419	0.400	ng	0.01
19) Phenanthrene-d10	17.414	188	13624	0.400	ng	0.01
29) Chrysene-d12	21.602	240	11142	0.400	ng	0.01
35) Perylene-d12	24.087	264	9797	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.536	112	3714	0.335	ng	0.01
5) Phenol-d6	7.161	99	4265	0.323	ng	0.01
8) Nitrobenzene-d5	9.174	82	3024	0.281	ng	0.01
11) 2-Methylnaphthalene-d10	12.420	152	5936	0.316	ng	0.01
14) 2,4,6-Tribromophenol	16.161	330	678	0.249	ng	0.01
15) 2-Fluorobiphenyl	13.287	172	8079	0.295	ng	0.01
27) Fluoranthene-d10	19.439	212	9386	0.294	ng	0.00
31) Terphenyl-d14	20.035	244	6329	0.360	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011923\
Data File : BN023716.D
Acq On : 19 Jan 2023 13:23
Operator : CG/JU
Sample : PB150214BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB150214BL

Quant Time: Jan 20 04:16:01 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN011823.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 04:14:02 2023
Response via : Initial Calibration

