

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012023\
 Data File : BN023725.D
 Acq On : 20 Jan 2023 13:33
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020387

Quant Time: Jan 20 21:56:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 03:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	226456	20.000	ng/u1	0.00
20) Naphthalene-d8	10.704	136	922010	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.545	164	605670	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.286	188	1267506	20.000	ng/u1	0.00
79) Chrysene-d12	21.486	240	1136497	20.000	ng/u1	0.00
88) Perylene-d12	23.909	264	1200614	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	42034	7.988	ng/uL	0.04
4) Pyridine-d5	3.699	84	288662	18.282	ng/u1	0.03
7) Phenol-d5	7.057	99	373635	18.616	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.222	67	238106	19.339	ng/u1	0.00
11) 2-Chlorophenol-d4	7.422	132	299784	19.503	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	289388	18.065	ng/u1	0.00
21) Nitrobenzene-d5	9.063	128	142132	20.443	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	151649	20.604	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	275394	19.617	ng/u1	0.00
31) 4-Chloroaniline-d4	10.845	131	407312	18.445	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	818143	18.159	ng/u1	0.00
49) Acenaphthylene-d8	14.233	160	938006	18.790	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	147716	16.105	ng/u1	0.00
60) Fluorene-d10	15.533	176	686136	17.822	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	75635	9.892	ng/u1	0.00
73) Anthracene-d10	17.386	188	1041574	18.364	ng/u1	0.00
81) Pyrene-d10	19.680	212	1205472	18.546	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.756	264	1099988	18.527	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	45200	8.136	ng/uL	99
5) Pyridine	3.722	79	299399	18.523	ng/u1	94
6) Benzaldehyde	7.028	77	167854	21.595	ng/u1	99
8) Phenol	7.081	94	382306	18.713	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.316	93	309341	19.163	ng/u1	98
12) 2-Chlorophenol	7.451	128	307646	19.368	ng/u1	99
13) 2-Methylphenol	8.340	108	277588	18.055	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.428	45	421510	19.524	ng/u1	100
16) Acetophenone	8.722	105	466152	18.827	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.710	70	241483	19.370	ng/u1	99
18) 4-Methylphenol	8.675	108	312934	18.482	ng/u1	98
19) Hexachloroethane	8.975	117	127170	20.017	ng/u1	94
22) Nitrobenzene	9.104	77	360574	20.722	ng/u1	100
23) Isophorone	9.634	82	647548	20.576	ng/u1	100
25) 2-Nitrophenol	9.816	139	169117	20.672	ng/u1	96
26) 2,4-Dimethylphenol	9.881	107	336499	20.267	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.116	93	398636	19.529	ng/u1	99
29) 2,4-Dichlorophenol	10.351	162	274528	19.543	ng/u1	98
30) Naphthalene	10.757	128	940324	18.972	ng/u1	98
32) 4-Chloroaniline	10.869	127	404255	18.279	ng/u1	99
33) Hexachlorobutadiene	11.040	225	165814	19.085	ng/u1	96
34) Caprolactam	11.645	113	74175	15.722	ng/u1	97
35) 4-Chloro-3-methylphenol	11.992	107	287170	18.749	ng/u1	97
36) 2-Methylnaphthalene	12.363	142	648418	19.625	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012023\
 Data File : BN023725.D
 Acq On : 20 Jan 2023 13:33
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020387

Quant Time: Jan 20 21:56:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 03:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.587	142	655680	19.424	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.734	216	326972	18.253	ng/ul	97
40) Hexachlorocyclopentadiene	12.710	237	209403	18.548	ng/ul	99
41) 2,4,6-Trichlorophenol	12.975	196	213080	18.046	ng/ul	93
42) 2,4,5-Trichlorophenol	13.045	196	233237	18.062	ng/ul	88
43) 1,1'-Biphenyl	13.375	154	853530	18.266	ng/ul	99
44) 2-Chloronaphthalene	13.422	162	666756	18.392	ng/ul	98
45) 2-Nitroaniline	13.628	65	197334	19.738	ng/ul	95
47) Dimethylphthalate	14.004	163	816430	18.033	ng/ul	99
48) 2,6-Dinitrotoluene	14.128	165	160604	18.765	ng/ul	98
50) Acenaphthylene	14.263	152	1038638	18.712	ng/ul	99
51) 3-Nitroaniline	14.451	138	155676	18.151	ng/ul	99
52) Acenaphthene	14.604	153	703739	17.983	ng/ul	99
53) 2,4-Dinitrophenol	14.657	184	31525	5.636	ng/ul	94
55) 4-Nitrophenol	14.757	109	121667	16.425	ng/ul	100
56) Dibenzofuran	14.939	168	969457	17.798	ng/ul	97
57) 2,4-Dinitrotoluene	14.910	165	233039	18.651	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.169	232	189822	17.155	ng/ul	97
59) Diethylphthalate	15.363	149	816813	18.011	ng/ul	99
61) Fluorene	15.586	166	788934	17.886	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.580	204	379747	17.684	ng/ul	99
63) 4-Nitroaniline	15.616	138	146849	18.596	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.669	198	75519	10.008	ng/ul	94
67) N-Nitrosodiphenylamine	15.798	169	678365	19.071	ng/ul	100
68) 4-Bromophenyl-phenylether	16.480	248	235546	18.540	ng/ul	99
69) Hexachlorobenzene	16.592	284	270493	18.168	ng/ul	98
70) Atrazine	16.751	200	240582	18.072	ng/ul	99
71) Pentachlorophenol	16.933	266	126547	13.318	ng/ul	98
72) Phenanthrene	17.327	178	1240739	18.269	ng/ul	100
74) Anthracene	17.422	178	1255075	18.568	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.339	216	318855	18.685	ng/ul	93
76) Pentachlorobenzene	14.857	250	322213	18.850	ng/ul	95
77) Carbazole	17.692	167	1138527	18.551	ng/ul	99
78) Di-n-butylphthalate	18.257	149	1383016	18.980	ng/ul	99
80) Fluoranthene	19.345	202	1392541	18.317	ng/ul	99
82) Pyrene	19.710	202	1492370	18.585	ng/ul	100
83) Butylbenzylphthalate	20.616	149	594342	18.696	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.404	252	464116	18.694	ng/ul#	95
85) Benzo(a)anthracene	21.474	228	1386272	18.007	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.392	149	897288	19.440	ng/ul	98
87) Chrysene	21.527	228	1306191	18.096	ng/ul	96
89) Di-n-octyl phthalate	22.327	149	1445168	17.441	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	1411030	18.450	ng/ul	99
91) Benzo(k)fluoranthene	23.221	252	1412481	18.673	ng/ul	99
93) Benzo(a)pyrene	23.809	252	1223948	18.453	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.439	276	1457138	17.777	ng/ul	99
95) Dibenzo(a,h)anthracene	26.456	278	1221837	18.027	ng/ul	98
96) Benzo(g,h,i)perylene	27.227	276	1133977	17.371	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 8.136 ng/uL
 RT: 3.311 min Scan# 31
 Delta R.T. 0.037 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 SST020387


Tgt Ion: 88 Resp: 45200
 Ion Ratio Lower Upper
 88 100
 43 30.6 24.6 36.8
 58 71.0 56.2 84.4



#5
 Pyridine
 Concen: 18.523 ng/uL
 RT: 3.722 min Scan# 108
 Delta R.T. 0.036 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion: 79 Resp: 299399
 Ion Ratio Lower Upper
 79 100
 52 69.3 59.9 89.9
 51 33.8 28.6 42.8









#6
Benzaldehyde
Concen: 21.595 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.007 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion: 77 Resp: 167854
Ion Ratio Lower Upper
77 100
105 93.2 75.1 112.7
106 88.9 72.8 109.2



#8
Phenol
Concen: 18.713 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion: 94 Resp: 382306
Ion Ratio Lower Upper
94 100
65 28.9 23.7 35.5
66 38.0 29.9 44.9





#10

Bis(2-Chloroethyl)ether

Concen: 19.163 ng/ul

RT: 7.316 min Scan# 719

Delta R.T. 0.007 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion: 93 Resp: 309341

Ion Ratio Lower Upper

93 100

63 77.1 61.2 91.8

95 30.9 26.6 40.0



#12

2-Chlorophenol

Concen: 19.368 ng/ul

RT: 7.451 min Scan# 742

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:128 Resp: 307646


Ion Ratio Lower Upper

128 100

64 46.7 37.4 56.2

130 31.1 25.6 38.4









#13
2-Methylphenol
Concen: 18.055 ng/ul
RT: 8.340 min Scan# 893
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:108 Resp: 277588
Ion Ratio Lower Upper
108 100
107 87.9 72.3 108.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.524 ng/ul
RT: 8.428 min Scan# 908
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion: 45 Resp: 421510
Ion Ratio Lower Upper
45 100
77 14.3 11.4 17.2
79 10.9 8.8 13.2









#16
 Acetophenone
 Concen: 18.827 ng/ul
 RT: 8.722 min Scan# 91
 Delta R.T. 0.001 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 SST020387


Tgt Ion:105 Resp: 466152
 Ion Ratio Lower Upper
 105 100
 77 80.1 64.2 96.2
 51 27.9 22.0 33.0



#17
 N-Nitroso-di-n-propylamine
 Concen: 19.370 ng/ul
 RT: 8.710 min Scan# 956
 Delta R.T. 0.001 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion: 70 Resp: 241483
 Ion Ratio Lower Upper
 70 100
 42 60.6 47.8 71.6
 101 9.3 7.8 11.6
 130 19.7 16.3 24.5









#18
4-Methylphenol
Concen: 18.482 ng/ul
RT: 8.675 min Scan# 91
Delta R.T. 0.007 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:108 Resp: 312934
Ion Ratio Lower Upper
108 100
107 110.9 90.7 136.1



#19
Hexachloroethane
Concen: 20.017 ng/ul
RT: 8.975 min Scan# 1001
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:117 Resp: 127170
Ion Ratio Lower Upper
117 100
201 89.7 76.9 115.3
199 58.8 44.6 67.0









#22
Nitrobenzene
Concen: 20.722 ng/ul
RT: 9.104 min Scan# 1023
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion: 77 Resp: 360574
Ion Ratio Lower Upper
77 100
123 44.6 35.6 53.4
65 13.8 11.0 16.4



#23
Isophorone
Concen: 20.576 ng/ul
RT: 9.634 min Scan# 1113
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion: 82 Resp: 647548
Ion Ratio Lower Upper
82 100
95 6.9 5.5 8.3
138 17.2 13.8 20.8









#25
 2-Nitrophenol
 Concen: 20.672 ng/ul
 RT: 9.816 min Scan# 1144
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Instrument :
 BNA_N
 ClientSampleId :
 SST020387

Tgt Ion:139 Resp: 169117




Ion	Ratio	Lower	Upper
139	100		
65	53.9	40.3	60.5
109	29.6	24.6	37.0




#26
 2,4-Dimethylphenol
 Concen: 20.267 ng/ul
 RT: 9.881 min Scan# 1155
 Delta R.T. 0.001 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:107 Resp: 336499

Ion	Ratio	Lower	Upper
107	100		
121	51.0	40.4	60.6
122	85.8	67.0	100.4





#27

Bis(2-Chloroethoxy)methane

Concen: 19.529 ng/ul

RT: 10.116 min Scan# 1195

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion: 93 Resp: 398636

Ion Ratio Lower Upper

93 100

95 31.7 25.5 38.3

123 11.1 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 19.543 ng/ul

RT: 10.351 min Scan# 1235

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion: 162 Resp: 274528


Ion Ratio Lower Upper

162 100

164 64.7 53.8 80.6

98 36.5 28.8 43.2





#30

Naphthalene

Concen: 18.972 ng/ul

RT: 10.757 min Scan# 11

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:128 Resp: 940324

Ion Ratio Lower Upper

128 100

129 10.6 9.1 13.7

127 13.1 10.8 16.2



#32

4-Chloroaniline

Concen: 18.279 ng/ul

RT: 10.869 min Scan# 1323

Delta R.T. -0.005 min

Lab File: BN023725.D




Acq: 20 Jan 2023 13:33


Tgt Ion:127 Resp: 404255

Ion Ratio Lower Upper

127 100

129 32.5 26.4 39.6









#33
Hexachlorobutadiene
Concen: 19.085 ng/ul
RT: 11.040 min Scan# 11
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:225 Resp: 165814
Ion Ratio Lower Upper
225 100
223 59.7 51.7 77.5
227 62.4 50.8 76.2



#34
Caprolactam
Concen: 15.722 ng/ul
RT: 11.645 min Scan# 1455
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:113 Resp: 74175
Ion Ratio Lower Upper
113 100
55 161.0 129.1 193.7
56 125.8 94.2 141.4









#35
4-Chloro-3-methylphenol
Concen: 18.749 ng/ul
RT: 11.992 min Scan# 1514
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:107 Resp: 287170
Ion Ratio Lower Upper
107 100
144 26.6 21.9 32.9
142 83.7 64.0 96.0



#36
2-Methylnaphthalene
Concen: 19.625 ng/ul
RT: 12.363 min Scan# 1577
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:142 Resp: 648418
Ion Ratio Lower Upper
142 100
141 88.7 73.6 110.4









#37
1-Methylnaphthalene
Concen: 19.424 ng/ul
RT: 12.587 min Scan# 1615
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:142 Resp: 655680
Ion Ratio Lower Upper
142 100
141 92.3 73.8 110.6
116 3.9 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.253 ng/ul
RT: 12.734 min Scan# 1640
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:216 Resp: 326972
Ion Ratio Lower Upper
216 100
214 81.5 62.4 93.6
179 20.7 16.2 24.2
108 19.4 14.6 21.8









#40
Hexachlorocyclopentadiene
Concen: 18.548 ng/ul
RT: 12.710 min Scan# 1636
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:237 Resp: 209403
Ion Ratio Lower Upper
237 100
235 64.5 51.0 76.4
272 13.7 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 18.046 ng/ul
RT: 12.975 min Scan# 1681
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:196 Resp: 213080
Ion Ratio Lower Upper
196 100
198 96.1 71.4 107.0
200 30.5 22.5 33.7









#42
2,4,5-Trichlorophenol
Concen: 18.062 ng/ul
RT: 13.045 min Scan# 1693
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:196 Resp: 233237
Ion Ratio Lower Upper
196 100
198 98.8 68.1 102.1
200 32.0 24.1 36.1



#43
1,1'-Biphenyl
Concen: 18.266 ng/ul
RT: 13.375 min Scan# 1749
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:154 Resp: 853530
Ion Ratio Lower Upper
154 100
153 41.1 32.5 48.7
76 14.7 11.0 16.6





#44

2-Chloronaphthalene

Concen: 18.392 ng/ul

RT: 13.422 min Scan# 1757

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:162 Resp: 666756

Ion Ratio Lower Upper

162 100

164 32.2 26.2 39.2

127 37.9 31.8 47.8



#45

2-Nitroaniline

Concen: 19.738 ng/ul

RT: 13.628 min Scan# 1792

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion: 65 Resp: 197334


Ion Ratio Lower Upper

65 100

92 64.8 53.5 80.3

138 99.2 84.0 126.0









#47
Dimethylphthalate
Concen: 18.033 ng/ul
RT: 14.004 min Scan# 1856
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:163 Resp: 816430
Ion Ratio Lower Upper
163 100
194 4.3 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 18.765 ng/ul
RT: 14.128 min Scan# 1877
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:165 Resp: 160604
Ion Ratio Lower Upper
165 100
89 63.0 49.9 74.9
121 21.4 19.1 28.7









#50
Acenaphthylene
Concen: 18.712 ng/ul
RT: 14.263 min Scan# 1900
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:152 Resp: 1038638
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.2 10.2 15.4



#51
3-Nitroaniline
Concen: 18.151 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:138 Resp: 155676
Ion Ratio Lower Upper
138 100
108 10.5 8.1 12.1
92 117.4 93.2 139.8









#52
 Acenaphthene
 Concen: 17.983 ng/ul
 RT: 14.604 min Scan# 1958
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020387


Tgt Ion:	153	Resp:	703739
Ion	Ratio	Lower	Upper
153	100		
152	49.1	38.5	57.7
154	90.0	71.5	107.3



#53
 2,4-Dinitrophenol
 Concen: 5.636 ng/ul
 RT: 14.657 min Scan# 1967
 Delta R.T. 0.001 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:	184	Resp:	31525
Ion	Ratio	Lower	Upper
184	100		
63	80.4	58.2	87.4
154	63.8	49.4	74.0









#55
4-Nitrophenol
Concen: 16.425 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:109 Resp: 121667
Ion Ratio Lower Upper
109 100
139 131.1 104.8 157.2
65 133.1 107.3 160.9



#56
Dibenzofuran
Concen: 17.798 ng/ul
RT: 14.939 min Scan# 2015
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:168 Resp: 969457
Ion Ratio Lower Upper
168 100
139 38.8 29.6 44.4









#57
 2,4-Dinitrotoluene
 Concen: 18.651 ng/ul
 RT: 14.910 min Scan# 2010
 Delta R.T. 0.001 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 SST020387


Tgt Ion:165 Resp: 233039
 Ion Ratio Lower Upper
 165 100
 63 45.3 37.2 55.8
 182 2.6 2.2 3.2



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 17.155 ng/ul
 RT: 15.169 min Scan# 2054
 Delta R.T. 0.001 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:232 Resp: 189822
 Ion Ratio Lower Upper
 232 100
 131 46.5 39.5 59.3
 130 2.5 2.2 3.4
 166 30.0 25.1 37.7









#59
 Diethylphthalate
 Concen: 18.011 ng/ul
 RT: 15.363 min Scan# 2087
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument : BNA_N
 ClientSampleId : SSTD020387


Tgt Ion:	149	Resp:	816813
Ion	Ratio	Lower	Upper
149	100		
177	21.8	17.9	26.9
150	12.3	9.9	14.9



#61
 Fluorene
 Concen: 17.886 ng/ul
 RT: 15.586 min Scan# 2125
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:	166	Resp:	788934
Ion	Ratio	Lower	Upper
166	100		
165	96.7	77.8	116.6
167	13.1	10.4	15.6









#62
4-Chlorophenyl-phenylether
Concen: 17.684 ng/ul
RT: 15.580 min Scan# 2124
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:204 Resp: 379747
Ion Ratio Lower Upper
204 100
206 33.5 27.0 40.6
141 63.0 51.0 76.4



#63
4-Nitroaniline
Concen: 18.596 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:138 Resp: 146849
Ion Ratio Lower Upper
138 100
92 52.3 41.3 61.9
108 79.9 64.2 96.2









#66
4,6-Dinitro-2-methylphenol
Concen: 10.008 ng/ul
RT: 15.669 min Scan# 2139
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:198 Resp: 75519
Ion Ratio Lower Upper
198 100
182 5.3 3.6 5.4
77 30.0 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 19.071 ng/ul
RT: 15.798 min Scan# 2161
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:169 Resp: 678365
Ion Ratio Lower Upper
169 100
168 67.0 53.4 80.2
167 35.9 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 18.540 ng/ul
RT: 16.480 min Scan# 2277
Delta R.T. 0.001 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33




Instrument :
BNA_N
ClientSampleId :
SSTD020387


Tgt Ion:248 Resp: 235546
Ion Ratio Lower Upper
248 100
250 94.0 76.4 114.6
141 71.8 58.0 87.0



#69
Hexachlorobenzene
Concen: 18.168 ng/ul
RT: 16.592 min Scan# 2296
Delta R.T. -0.005 min
Lab File: BN023725.D
Acq: 20 Jan 2023 13:33

Tgt Ion:284 Resp: 270493
Ion Ratio Lower Upper
284 100
142 38.2 28.9 43.3
249 30.4 24.2 36.2





#70

Atrazine

Concen: 18.072 ng/ul

RT: 16.751 min Scan# 21

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:200 Resp: 240582

Ion Ratio Lower Upper

200 100

173 25.6 21.4 32.2

215 48.4 39.0 58.6



#71

Pentachlorophenol

Concen: 13.318 ng/ul

RT: 16.933 min Scan# 2354

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:266 Resp: 126547


Ion Ratio Lower Upper

266 100

264 66.4 51.4 77.0

268 66.5 54.5 81.7









#72
 Phenanthrene
 Concen: 18.269 ng/ul
 RT: 17.327 min Scan# 2421
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020387

Tgt Ion:178 Resp: 1240739




Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.3	18.5
176	19.3	15.2	22.8




#74
 Anthracene
 Concen: 18.568 ng/ul
 RT: 17.422 min Scan# 2437
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:178 Resp: 1255075

Ion	Ratio	Lower	Upper
178	100		
179	15.0	12.6	19.0
176	18.1	15.1	22.7









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 18.685 ng/uL
 RT: 13.339 min Scan# 11
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 SST020387


Tgt Ion:	216	Resp:	318855
Ion Ratio	Lower	Upper	
216	100		
214	85.0	61.0	91.4
218	48.6	39.8	59.8



#76
 Pentachlorobenzene
 Concen: 18.850 ng/uL
 RT: 14.857 min Scan# 2001
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:	250	Resp:	322213
Ion Ratio	Lower	Upper	
250	100		
248	56.9	50.3	75.5
252	62.9	48.9	73.3









#77
 Carbazole
 Concen: 18.551 ng/ul
 RT: 17.692 min Scan# 2483
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020387


Tgt Ion:167 Resp: 1138527
 Ion Ratio Lower Upper
 167 100
 166 22.7 17.4 26.2
 139 13.7 11.0 16.6



#78
 Di-n-butylphthalate
 Concen: 18.980 ng/ul
 RT: 18.257 min Scan# 2579
 Delta R.T. -0.005 min
 Lab File: BN023725.D
 Acq: 20 Jan 2023 13:33

Tgt Ion:149 Resp: 1383016
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.2
 104 5.4 4.3 6.5





#80

Fluoranthene

Concen: 18.317 ng/ul

RT: 19.345 min Scan# 21

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:202 Resp: 1392541

Ion Ratio Lower Upper

202 100

101 11.4 9.5 14.3

100 8.8 7.3 10.9



#82

Pyrene

Concen: 18.585 ng/ul

RT: 19.710 min Scan# 2826

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:202 Resp: 1492370


Ion Ratio Lower Upper

202 100

101 13.6 10.9 16.3

100 11.1 8.9 13.3





#83

Butylbenzylphthalate

Concen: 18.696 ng/ul

RT: 20.616 min Scan# 2980

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:149 Resp: 594342

Ion Ratio Lower Upper

149 100

91 74.9 57.9 86.9

206 21.2 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 18.694 ng/ul

RT: 21.404 min Scan# 3114

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:252 Resp: 464116


Ion Ratio Lower Upper

252 100

254 67.8 51.2 76.8

126 13.6 8.8 13.2#





#85

Benzo(a)anthracene

Concen: 18.007 ng/ul

RT: 21.474 min Scan# 3126

Delta R.T. 0.007 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:228 Resp: 1386272

Ion Ratio Lower Upper

228 100

229 19.1 15.0 22.4

226 27.5 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.440 ng/ul

RT: 21.392 min Scan# 3112

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:149 Resp: 897288


Ion Ratio Lower Upper

149 100

167 26.6 22.0 33.0

279 4.1 3.4 5.2





#87

Chrysene

Concen: 18.096 ng/ul

RT: 21.527 min Scan# 3135

Delta R.T. 0.007 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:228 Resp: 1306191

Ion Ratio Lower Upper

228 100

226 28.5 24.6 37.0

229 19.1 16.8 25.2



#89

Di-n-octyl phthalate

Concen: 17.441 ng/ul




RT: 22.327 min Scan# 3271


Delta R.T. 0.007 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Tgt Ion:149 Resp: 1445168





#90

Benzo(b)fluoranthene

Concen: 18.450 ng/ul

RT: 23.168 min Scan# 3414

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:252 Resp: 1411030

Ion Ratio Lower Upper

252 100

253 22.0 17.2 25.8

125 10.3 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 18.673 ng/ul

RT: 23.221 min Scan# 3423

Delta R.T. 0.007 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:252 Resp: 1412481


Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8

125 10.2 8.6 12.8





#93

Benzo(a)pyrene

Concen: 18.453 ng/ul

RT: 23.809 min Scan# 3523

Delta R.T. 0.013 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:252 Resp: 1223948

Ion Ratio Lower Upper

252 100

253 20.6 16.6 25.0

125 12.2 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.777 ng/ul

RT: 26.439 min Scan# 3970

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33




Tgt Ion:276 Resp: 1457138


Ion Ratio Lower Upper

276 100

138 25.8 20.6 30.8

277 24.8 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 18.027 ng/ul

RT: 26.456 min Scan# 3973

Delta R.T. -0.005 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Instrument :

BNA_N

ClientSampleId :

SSTD020387





Tgt Ion:278 Resp: 1221837

Ion Ratio Lower Upper

278 100

139 17.2 14.1 21.1

279 24.2 18.0 27.0



#96

Benzo(g,h,i)perylene

Concen: 17.371 ng/ul

RT: 27.227 min Scan# 4104

Delta R.T. 0.001 min

Lab File: BN023725.D

Acq: 20 Jan 2023 13:33

Tgt Ion:276 Resp: 1133977

Ion Ratio Lower Upper

276 100

138 22.2 17.0 25.6

277 22.6 17.6 26.4

