


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012023\
Data File : BN023724.D
Acq On : 20 Jan 2023 12:56
Operator : CG/JU
Sample : DFTPP270
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP270

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
Title : SVOA CALIBRATION
Last Update : Fri Jan 20 23:02:18 2023



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2482

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	41.5	191570	PASS
68	69	0.00	2	1.8	3576	PASS
69	198	0.00	100	43.3	199893	PASS
70	69	0.00	2	0.6	1120	PASS
127	198	10	80	51.1	235989	PASS
197	198	0.00	2	0.6	2542	PASS
198	198	100	100	100.0	462144	PASS
199	198	5	9	6.7	30760	PASS
275	198	10	60	26.0	120168	PASS
365	198	1	100	3.2	14811	PASS
441	198	0.01	100	13.5	62456	PASS
442	198	50	100	85.9	397120	PASS
443	442	15	24	20.4	81136	PASS

m/z	Abundance
39.00	202.0
40.00	1172.0
41.00	362.0
42.00	173.0
43.00	314.0
44.00	1242.0
44.90	160.0
46.90	173.0
48.90	523.0
51.00	217.0
55.00	190.0
57.00	187.0
73.00	730.0
74.90	218.0
76.90	384.0
78.10	152.0
78.90	172.0
83.90	387.0
85.90	292.0
91.10	224.0
95.90	388.0
133.10	433.0
135.10	154.0
147.00	345.0
176.80	161.0
190.90	353.0
192.80	319.0
207.00	3467.0
208.00	815.0
209.00	373.0
220.90	223.0
281.00	506.0
282.00	207.0
355.00	174.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

m/z	Abundance
36.00	256.0
37.00	904.0
38.00	2288.0
39.00	13962.0
40.00	1533.0
41.00	651.0
42.90	309.0
43.90	1357.0
45.00	454.0
46.90	181.0
47.90	222.0
48.90	1436.0
50.00	45888.0
51.00	172096.0
52.00	9354.0
53.00	512.0
55.00	1048.0
55.90	5046.0
57.00	11552.0
57.90	568.0
59.00	230.0
60.00	200.0
60.90	2249.0
62.00	2773.0
63.00	6511.0
64.00	973.0
65.00	3407.0
65.90	341.0
67.00	349.0
68.00	3322.0
69.00	178368.0
70.00	913.0
71.90	154.0
73.00	2246.0
74.00	19808.0
75.00	29656.0
76.00	10374.0
77.00	208320.0
78.00	15144.0
79.00	13470.0
80.00	10735.0
81.00	14782.0
82.00	3579.0
83.00	2753.0
84.00	623.0
84.90	2551.0
86.00	4495.0
87.00	1845.0
88.00	711.0
88.90	397.0
90.90	3009.0
92.00	3739.0
93.00	24256.0
93.90	1745.0
94.90	377.0
96.00	1391.0
96.90	429.0
97.90	18768.0
99.00	16704.0
99.90	1437.0
101.00	7735.0
101.90	391.0
103.00	2488.0
103.90	4745.0
105.00	4561.0
105.90	1685.0
107.00	57920.0
108.00	9539.0
109.00	1858.0
110.00	102880.0
111.00	15648.0
111.90	2027.0
113.00	589.0
114.90	286.0
116.00	3239.0
116.90	59592.0
118.00	4833.0
118.90	549.0
120.00	603.0
120.90	433.0
122.00	3986.0
123.00	6026.0
123.90	2590.0
125.00	2516.0
126.00	812.0
127.00	194432.0
128.00	15397.0
129.00	75392.0
130.00	6453.0
130.90	1284.0
132.00	596.0
132.80	555.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

133.90	2460.0
135.00	6175.0
135.90	2685.0
137.00	4030.0
138.00	885.0
138.90	433.0
140.10	1076.0
140.90	10256.0
142.00	3361.0
142.90	2233.0
144.00	586.0
144.90	595.0
146.00	1737.0
147.00	5914.0
148.00	13436.0
149.00	2966.0
149.90	769.0
151.00	1489.0
151.90	713.0
153.00	3003.0
154.00	2352.0
155.00	5748.0
156.00	8194.0
157.00	1574.0
157.90	1904.0
159.00	1488.0
159.90	2845.0
161.00	4735.0
161.90	1358.0
162.90	545.0
164.00	593.0
165.00	3427.0
165.90	2680.0
167.00	22664.0
167.90	15256.0
169.00	2128.0
169.90	664.0
171.00	835.0
171.90	1823.0
172.90	2610.0
173.90	4070.0
175.00	8542.0
176.00	1958.0
176.90	3819.0
177.90	1086.0
178.90	14804.0
180.00	11575.0
181.00	4799.0
182.00	807.0
182.80	492.0
184.00	1018.0
185.00	7219.0
186.00	52696.0
187.00	15785.0
188.00	1826.0
188.90	3896.0
190.00	656.0
191.00	1949.0
192.00	5156.0
193.00	5315.0
194.00	1072.0
195.00	459.0
196.00	10901.0
197.00	2075.0
197.90	361216.0
199.00	25424.0
200.00	2111.0
201.50	1715.0
203.00	2659.0
204.00	13073.0
205.00	23600.0
206.00	98216.0
207.00	15376.0
208.00	3846.0
208.90	1535.0
210.20	1866.0
210.90	3857.0
211.80	944.0
212.90	248.0
214.90	961.0
215.90	2280.0
217.00	25112.0
218.00	3453.0
219.20	262.0
221.00	14318.0
221.90	3186.0
222.90	5847.0
224.00	54392.0
225.00	12769.0
226.00	1783.0
227.00	22040.0
228.00	3440.0
229.00	4250.0
230.00	700.0
231.00	1670.0
232.00	354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

233.00	309.0
234.00	1532.0
235.00	1702.0
235.90	1061.0
236.90	2774.0
238.00	451.0
238.90	786.0
240.00	834.0
241.00	1336.0
242.00	2753.0
243.00	2767.0
244.00	39576.0
245.00	6208.0
245.90	7954.0
247.00	1730.0
247.90	371.0
249.00	1404.0
249.80	338.0
250.80	368.0
251.90	349.0
252.90	932.0
255.00	190464.0
256.00	29232.0
257.00	2782.0
258.00	11381.0
258.90	1966.0
260.00	395.0
261.00	327.0
262.90	176.0
263.90	317.0
264.90	4631.0
266.00	835.0
266.80	215.0
270.00	207.0
270.90	850.0
272.00	623.0
273.00	6113.0
273.90	16187.0
275.00	82488.0
276.00	13085.0
276.90	6920.0
278.00	1583.0
278.90	400.0
281.00	892.0
281.90	341.0
283.00	776.0
284.00	614.0
285.00	1122.0
286.00	178.0
288.90	274.0
289.90	272.0
291.00	227.0
292.00	377.0
293.00	1646.0
294.00	693.0
294.90	616.0
296.00	24616.0
297.00	3695.0
297.80	242.0
301.10	192.0
302.00	384.0
303.00	2706.0
304.00	663.0
304.90	171.0
307.70	230.0
308.00	244.0
308.90	261.0
309.90	246.0
312.80	263.0
314.00	1141.0
315.00	2582.0
315.90	1499.0
316.90	313.0
320.90	681.0
322.00	389.0
323.00	5976.0
324.00	1551.0
324.90	185.0
326.90	1550.0
327.90	828.0
329.00	185.0
331.90	473.0
333.10	1037.0
334.00	4175.0
334.90	1179.0
335.90	154.0
339.00	157.0
341.00	951.0
341.90	262.0
345.90	1547.0
346.90	398.0
352.00	2271.0
352.90	1522.0
354.00	2342.0
355.00	524.0
364.90	9801.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

365.90	1503.0
366.90	186.0
370.00	205.0
370.90	661.0
371.90	3633.0
372.90	1009.0
383.00	1026.0
384.00	378.0
389.90	367.0
391.00	424.0
392.00	175.0
400.90	296.0
402.00	1381.0
403.00	1945.0
403.90	892.0
404.90	185.0
414.90	190.0
421.00	1681.0
421.90	1722.0
423.00	12079.0
424.00	4500.0
425.00	765.0
440.90	32720.0
441.90	222848.0
443.00	48344.0
444.00	3941.0
444.80	317.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

m/z	Abundance
36.00	237.0
37.00	945.0
38.00	3029.0
39.00	17584.0
40.00	1682.0
41.00	709.0
43.00	395.0
44.00	1241.0
45.00	478.0
48.10	289.0
49.00	1987.0
50.00	59848.0
51.00	225152.0
52.00	11983.0
53.00	501.0
55.00	1498.0
56.00	6506.0
57.00	14151.0
57.90	763.0
58.90	288.0
59.90	208.0
61.00	2979.0
62.00	3811.0
63.00	9219.0
64.00	1303.0
65.00	4193.0
66.00	443.0
67.10	483.0
68.00	4007.0
69.00	235008.0
70.00	1327.0
73.00	2560.0
74.00	25768.0
75.00	39344.0
76.00	12828.0
77.00	277440.0
78.00	19440.0
79.00	17736.0
80.00	13178.0
81.00	19120.0
82.00	4740.0
83.00	4358.0
83.90	927.0
85.00	3172.0
86.00	5654.0
87.00	2578.0
87.90	1056.0
89.00	501.0
91.00	4159.0
92.00	4748.0
93.00	33360.0
93.90	2230.0
94.90	547.0
96.00	1387.0
97.20	554.0
98.00	25744.0
99.00	20016.0
100.00	1826.0
100.90	9785.0
102.00	687.0
103.00	3000.0
104.00	6328.0
104.90	6071.0
106.00	1991.0
106.90	82448.0
108.00	12615.0
109.00	2386.0
110.00	139584.0
111.00	22272.0
112.00	3072.0
113.00	914.0
114.20	173.0
115.00	376.0
116.00	4621.0
116.90	80976.0
118.00	5852.0
118.90	766.0
119.90	889.0
120.90	413.0
122.00	5468.0
123.00	8421.0
123.90	4031.0
124.90	3408.0
127.00	278144.0
128.00	20312.0
129.00	110152.0
130.00	9091.0
131.00	1784.0
132.00	1020.0
132.90	626.0
133.90	3200.0
135.00	9581.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

136.00	3703.0
136.90	5345.0
138.00	1108.0
138.90	569.0
139.90	1460.0
141.00	14470.0
141.90	4712.0
143.00	3180.0
143.90	903.0
145.00	643.0
145.90	2487.0
147.00	7680.0
148.00	19352.0
149.00	4280.0
150.00	1134.0
151.00	1858.0
151.80	1179.0
153.00	4160.0
154.00	3315.0
155.00	8015.0
156.00	11421.0
157.00	2249.0
157.90	2367.0
158.90	1854.0
160.00	4505.0
161.00	6785.0
162.00	1853.0
162.90	539.0
164.00	675.0
164.90	5597.0
166.00	3992.0
167.00	31336.0
168.00	20104.0
169.00	2812.0
170.00	982.0
170.90	1290.0
171.90	2822.0
173.00	3440.0
174.00	5642.0
175.00	11373.0
176.00	2849.0
176.90	5249.0
177.90	1935.0
179.00	21264.0
180.00	14132.0
181.00	6562.0
182.00	1101.0
182.90	423.0
183.90	1820.0
185.00	10114.0
186.00	78840.0
187.00	23728.0
188.00	2311.0
188.90	5177.0
190.00	812.0
191.00	2756.0
192.00	7745.0
193.00	7957.0
194.00	1672.0
195.10	765.0
196.00	16079.0
196.90	3217.0
197.90	534144.0
198.90	35168.0
200.00	3111.0
201.50	2543.0
203.00	3614.0
204.00	20416.0
205.00	34136.0
206.00	146624.0
207.00	22768.0
208.00	5374.0
209.00	2138.0
210.00	2360.0
211.00	6040.0
212.90	447.0
214.10	214.0
215.00	1865.0
216.00	3248.0
217.00	39112.0
218.00	5377.0
219.00	448.0
221.00	22824.0
221.80	5672.0
223.00	8792.0
224.00	77280.0
225.00	20296.0
226.00	2913.0
227.00	32816.0
227.90	4699.0
229.00	7273.0
229.90	1011.0
231.00	2837.0
231.90	436.0
232.90	551.0
233.90	2197.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

234.90	2558.0
236.00	1798.0
236.90	3589.0
238.00	463.0
239.00	1206.0
240.00	1168.0
241.00	2195.0
242.00	4054.0
243.00	4622.0
244.00	63728.0
245.00	8782.0
245.90	13447.0
247.00	2663.0
247.90	637.0
248.90	2392.0
249.90	491.0
250.90	449.0
251.90	510.0
253.00	1438.0
254.00	2437.0
255.00	313024.0
256.00	45928.0
257.00	4157.0
258.00	18248.0
259.00	3012.0
260.00	558.0
261.00	477.0
262.80	252.0
264.00	567.0
265.00	8098.0
265.90	1299.0
266.90	240.0
269.90	315.0
270.90	745.0
271.90	1040.0
272.90	9683.0
274.00	25392.0
275.00	135616.0
276.00	20008.0
277.00	12758.0
278.00	2499.0
278.90	550.0
281.00	689.0
281.90	462.0
282.90	1541.0
284.00	869.0
285.00	1993.0
286.00	423.0
289.00	349.0
289.90	394.0
290.90	385.0
292.00	630.0
293.00	2786.0
293.90	702.0
295.00	838.0
296.00	40992.0
297.00	5792.0
298.00	579.0
300.80	470.0
302.00	607.0
303.00	4387.0
304.00	1318.0
305.00	199.0
307.90	595.0
309.00	316.0
309.90	503.0
312.90	457.0
314.00	1976.0
314.90	4349.0
316.00	2373.0
317.00	469.0
319.90	239.0
320.90	1198.0
321.90	660.0
323.00	12771.0
324.00	2470.0
325.90	289.0
326.90	2720.0
327.90	1277.0
329.00	197.0
332.00	969.0
333.00	1297.0
334.00	8687.0
334.90	2116.0
336.00	278.0
339.00	212.0
339.90	193.0
341.00	1734.0
341.90	482.0
345.90	2672.0
346.90	559.0
350.90	207.0
352.00	3592.0
353.00	2914.0
354.00	4323.0
354.90	787.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

356.00	211.0
358.90	357.0
364.00	217.0
365.00	16776.0
365.90	2600.0
366.90	160.0
369.90	456.0
370.90	939.0
372.00	6251.0
373.00	1564.0
373.80	201.0
376.90	228.0
383.00	1758.0
384.00	577.0
384.90	193.0
389.90	801.0
391.00	629.0
391.90	500.0
400.90	334.0
401.90	2484.0
403.00	4094.0
404.00	1489.0
404.90	459.0
410.00	164.0
414.90	227.0
420.90	3479.0
421.90	3651.0
423.00	23712.0
424.00	7251.0
425.00	917.0
438.80	217.0
440.90	67656.0
442.00	429568.0
443.00	90696.0
444.00	8586.0
444.90	540.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

m/z	Abundance
35.00	212.0
37.00	922.0
38.00	2636.0
39.00	13651.0
40.00	1753.0
41.00	550.0
42.00	165.0
43.00	360.0
44.00	1262.0
44.90	480.0
47.00	247.0
48.00	170.0
49.00	1685.0
50.00	48256.0
51.00	178112.0
52.00	9349.0
53.00	387.0
55.00	1216.0
56.00	5113.0
57.00	11881.0
58.00	591.0
59.00	222.0
59.90	177.0
61.00	2315.0
62.00	2365.0
63.00	7416.0
64.00	1043.0
65.00	3340.0
66.00	334.0
67.00	483.0
68.00	3399.0
69.00	186304.0
70.00	1121.0
73.00	2465.0
74.00	21000.0
75.00	30848.0
76.00	11694.0
77.00	222080.0
78.00	15297.0
79.00	13797.0
80.00	11125.0
81.00	15021.0
82.00	3875.0
83.00	3118.0
84.00	731.0
85.00	2826.0
86.00	4927.0
87.00	1969.0
87.80	883.0
89.00	414.0
90.90	3413.0
91.90	3980.0
93.00	26240.0
94.00	1761.0
95.00	390.0
96.00	1507.0
96.90	478.0
98.00	21768.0
99.00	16270.0
100.00	1253.0
100.90	8437.0
102.00	484.0
102.90	2736.0
103.90	5408.0
105.00	5604.0
106.00	1695.0
107.00	65232.0
108.00	10049.0
108.90	1743.0
110.00	117152.0
111.00	17832.0
111.90	2365.0
112.90	871.0
115.00	354.0
116.00	3824.0
117.00	68088.0
117.90	4496.0
118.90	664.0
119.90	836.0
120.80	417.0
122.00	4491.0
123.00	7573.0
124.00	3049.0
125.00	3085.0
127.00	235392.0
128.00	17648.0
129.00	90880.0
130.00	7455.0
131.00	1634.0
132.00	948.0
133.00	722.0
133.90	2640.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

134.90	7769.0
136.00	2872.0
137.00	4434.0
137.90	860.0
138.90	442.0
139.90	1025.0
141.00	12080.0
141.90	3690.0
143.00	2884.0
144.00	684.0
145.00	643.0
146.00	2157.0
147.00	6296.0
148.00	16097.0
149.00	3364.0
149.90	855.0
151.10	1547.0
151.80	1050.0
153.00	4492.0
154.00	3142.0
155.00	6758.0
156.00	9597.0
157.00	1596.0
157.90	2191.0
159.00	1619.0
160.00	3934.0
161.00	5896.0
161.90	1680.0
162.90	506.0
163.90	554.0
164.90	4478.0
166.00	4085.0
167.00	27944.0
168.00	17224.0
169.00	2487.0
170.10	772.0
171.00	1136.0
172.00	2303.0
172.90	2901.0
174.00	5429.0
175.00	9904.0
176.00	2628.0
177.00	4704.0
178.00	1504.0
178.90	17936.0
180.00	13558.0
181.00	5609.0
181.90	1158.0
182.80	449.0
183.90	1662.0
185.00	9444.0
186.00	67872.0
187.00	19088.0
188.00	2376.0
188.90	4275.0
189.90	731.0
191.00	2368.0
191.90	6315.0
193.00	7614.0
194.00	1568.0
194.90	800.0
196.00	13904.0
196.80	2334.0
197.90	491072.0
198.90	31688.0
199.90	2674.0
201.50	2160.0
203.00	3623.0
204.00	18992.0
205.00	32120.0
206.00	130328.0
207.00	20208.0
208.00	5220.0
209.00	1909.0
210.00	1950.0
211.00	4789.0
213.00	418.0
215.00	1383.0
216.00	2769.0
216.90	35944.0
218.00	4648.0
219.00	460.0
221.10	19240.0
221.90	4642.0
222.90	7936.0
224.00	71440.0
225.00	19776.0
226.00	2381.0
227.00	31848.0
228.00	4384.0
229.00	7219.0
230.00	967.0
231.00	2427.0
232.00	429.0
232.90	449.0
234.00	2200.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

235.00	2680.0
236.00	1723.0
237.00	3041.0
237.90	530.0
239.00	1354.0
240.00	1063.0
240.90	1556.0
242.00	4179.0
243.00	4168.0
244.00	62864.0
245.00	7636.0
246.00	12692.0
247.00	2156.0
248.00	497.0
248.90	1885.0
249.90	409.0
251.00	447.0
252.00	623.0
252.90	1264.0
254.00	2056.0
255.00	288576.0
256.00	46536.0
257.00	3419.0
258.00	17528.0
258.90	2800.0
260.00	531.0
260.90	578.0
263.00	210.0
263.90	363.0
265.00	7337.0
266.00	1092.0
266.90	255.0
269.80	316.0
270.90	948.0
271.90	821.0
273.00	9753.0
274.00	24000.0
275.00	142400.0
276.00	18432.0
277.00	12236.0
277.90	2123.0
278.90	477.0
281.00	737.0
281.90	552.0
282.90	1509.0
283.90	879.0
285.00	2100.0
285.80	397.0
288.00	228.0
288.90	602.0
289.90	409.0
290.90	331.0
291.90	571.0
292.90	2449.0
294.00	858.0
295.00	893.0
296.00	42904.0
297.00	5698.0
298.00	592.0
301.00	485.0
302.00	813.0
303.00	4939.0
304.00	1467.0
308.00	650.0
309.00	279.0
310.00	490.0
310.90	179.0
313.00	250.0
314.00	1854.0
315.00	4818.0
316.00	2359.0
316.90	388.0
319.90	190.0
320.90	1284.0
322.00	672.0
323.00	14340.0
324.00	2619.0
324.90	153.0
325.80	357.0
326.90	2847.0
328.00	1379.0
328.90	344.0
332.00	995.0
333.00	1394.0
334.00	8270.0
335.00	2237.0
335.90	337.0
339.00	221.0
340.00	212.0
340.90	1790.0
341.90	532.0
345.90	2888.0
346.90	581.0
350.90	181.0
352.00	3516.0
352.90	2840.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270

353.90	4286.0
354.90	1160.0
358.90	425.0
364.90	17856.0
365.90	3081.0
370.10	532.0
370.90	1129.0
372.00	7164.0
373.00	1916.0
373.90	211.0
383.00	1841.0
384.00	539.0
384.90	249.0
389.90	1071.0
391.00	659.0
392.00	491.0
401.00	458.0
401.90	3121.0
403.00	3643.0
404.00	1472.0
404.80	333.0
414.90	205.0
420.90	3920.0
422.00	3536.0
423.00	27160.0
424.00	7659.0
425.00	721.0
438.90	155.0
439.90	231.0
441.00	86992.0
442.00	538944.0
443.00	104368.0
444.00	10320.0
445.00	646.0

Instrument :
BNA_N
ClientSampleId :
DFTPP270