

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012422\
 Data File : BN018452.D
 Acq On : 24 Jan 2022 15:42
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jan 24 17:54:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 24 17:52:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.965	152	35366	0.400	ng	0.00
7) Naphthalene-d8	10.761	136	150677	0.400	ng	0.00
13) Acenaphthene-d10	14.573	164	81510	0.400	ng	0.00
19) Phenanthrene-d10	17.309	188	149578	0.400	ng	0.00
29) Chrysene-d12	21.482	240	112660	0.400	ng	#-0.01
35) Perylene-d12	23.895	264	92955	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.512	112	145796	1.676	ng	0.00
5) Phenol-d6	7.107	99	160339	1.838	ng	0.00
8) Nitrobenzene-d5	9.114	82	149646	1.560	ng	0.00
11) 2-Methylnaphthalene-d10	12.345	152	422083	1.790	ng	0.00
14) 2,4,6-Tribromophenol	16.056	330	47285	1.467	ng	0.00
15) 2-Fluorobiphenyl	13.215	172	487164	1.655	ng	0.00
27) Fluoranthene-d10	19.341	212	766495	1.794	ng	0.00
31) Terphenyl-d14	19.936	244	545337	2.236	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.391	88	32311	1.624	ng	# 58
3) n-Nitrosodimethylamine	3.723	42	49607	1.556	ng	97
6) bis(2-Chloroethyl)ether	7.375	93	160222	1.697	ng	98
9) Naphthalene	10.812	128	612243	1.600	ng	100
10) Hexachlorobutadiene	11.116	225	120826	1.699	ng	# 100
12) 2-Methylnaphthalene	12.420	142	418338	1.784	ng	98
16) Acenaphthylene	14.294	152	537640	1.661	ng	100
17) Acenaphthene	14.636	154	362324	1.603	ng	96
18) Fluorene	15.613	166	434923	1.713	ng	100
20) 4,6-Dinitro-2-methylph...	15.689	198	12930	1.230	ng	# 80
21) 4-Bromophenyl-phenylether	16.506	248	135840	1.652	ng	96
22) Hexachlorobenzene	16.628	284	149374	1.365	ng	# 93
23) Atrazine	16.774	200	94705	1.731	ng	97
24) Pentachlorophenol	16.969	266	48108	1.807	ng	99
25) Phenanthrene	17.358	178	668178	1.589	ng	100
26) Anthracene	17.443	178	583413	1.679	ng	99
28) Fluoranthene	19.370	202	730509	1.636	ng	# 98
30) Pyrene	19.728	202	755649	1.895	ng	100
32) Benzo(a)anthracene	21.471	228	596096	1.796	ng	99
33) Chrysene	21.524	228	594603	1.578	ng	98
34) Bis(2-ethylhexyl)phtha...	21.408	149	433584	2.179	ng	98
36) Indeno(1,2,3-cd)pyrene	26.397	276	299833	1.015	ng	# 94
37) Benzo(b)fluoranthene	23.159	252	564626	1.874	ng	# 98
38) Benzo(k)fluoranthene	23.210	252	490890	1.666	ng	# 97
39) Benzo(a)pyrene	23.785	252	423547	1.521	ng	# 97
40) Dibenzo(a,h)anthracene	26.421	278	199380	0.902	ng	# 95
41) Benzo(g,h,i)perylene	27.164	276	297286	1.100	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012422\
Data File : BN018452.D
Acq On : 24 Jan 2022 15:42
Operator : CG/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Jan 24 17:54:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN012422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 24 17:52:32 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.965 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 35366
Ion Ratio Lower Upper
152 100
150 155.0 121.7 182.5
115 58.2 43.8 65.8



#2
1,4-Dioxane
Concen: 1.624 ng
RT: 3.391 min Scan# 34
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion: 88 Resp: 32311
Ion Ratio Lower Upper
88 100
43 29.3 69.7 104.5#
58 79.1 83.8 125.8#









#3
n-Nitrosodimethylamine
Concen: 1.556 ng
RT: 3.723 min Scan# 70
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 49607
Ion Ratio Lower Upper
42 100
74 140.2 108.6 163.0
44 7.2 5.8 8.8



#4
2-Fluorophenol
Concen: 1.676 ng
RT: 5.512 min Scan# 264
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion: 112 Resp: 145796
Ion Ratio Lower Upper
112 100
64 56.8 48.0 72.0
63 28.8 25.3 37.9









#5
Phenol-d6
Concen: 1.838 ng
RT: 7.107 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 160339
Ion Ratio Lower Upper
99 100
42 17.7 16.3 24.5
71 30.3 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 1.697 ng
RT: 7.375 min Scan# 466
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion: 93 Resp: 160222
Ion Ratio Lower Upper
93 100
63 75.0 61.7 92.5
95 32.3 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.761 min Scan# 761
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:136 Resp: 150677
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 7.5 5.0 7.6
68 5.9 4.4 6.6



#8
Nitrobenzene-d5
Concen: 1.560 ng
RT: 9.114 min Scan# 631
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion: 82 Resp: 149646
Ion Ratio Lower Upper
82 100
128 46.7 33.7 50.5
54 45.7 36.6 55.0









#9
Naphthalene
Concen: 1.600 ng
RT: 10.812 min Scan# 765
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 612243
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 12.8 10.4 15.6



#10
Hexachlorobutadiene
Concen: 1.699 ng
RT: 11.116 min Scan# 789
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:225 Resp: 120826
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 1.790 ng
RT: 12.345 min Scan# 1008
Delta R.T. -0.005 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 422083
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2



#12
2-Methylnaphthalene
Concen: 1.784 ng
RT: 12.420 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:142 Resp: 418338
Ion Ratio Lower Upper
142 100
141 86.6 70.0 105.0
115 32.2 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.573 min Scan# 1207
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 81510
Ion Ratio Lower Upper
164 100
162 101.4 80.2 120.2
160 46.3 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 1.467 ng
RT: 16.056 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:330 Resp: 47285
Ion Ratio Lower Upper
330 100
332 97.8 0.0 0.0#
141 34.2 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 1.655 ng
RT: 13.215 min Scan# 1100
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 487164
Ion Ratio Lower Upper
172 100
171 33.2 27.4 41.0
170 21.4 18.4 27.6



#16
Acenaphthylene
Concen: 1.661 ng
RT: 14.294 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:152 Resp: 537640
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.6
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 1.603 ng
RT: 14.636 min Scan# 117
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 362324
Ion Ratio Lower Upper
154 100
153 110.4 90.9 136.3
152 52.5 45.0 67.4



#18
Fluorene
Concen: 1.713 ng
RT: 15.613 min Scan# 1289
Delta R.T. -0.013 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:166 Resp: 434923
Ion Ratio Lower Upper
166 100
165 96.8 77.8 116.6
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.309 min Scan# 1427
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 149578
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.2 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 1.230 ng
RT: 15.689 min Scan# 1295
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:198 Resp: 12930
Ion Ratio Lower Upper
198 100
182 10.5 15.5 23.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.652 ng
RT: 16.506 min Scan# 1361
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 135840
Ion Ratio Lower Upper
248 100
250 95.1 74.4 111.6
141 59.7 43.7 65.5



#22
Hexachlorobenzene
Concen: 1.365 ng
RT: 16.628 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:284 Resp: 149374
Ion Ratio Lower Upper
284 100
142 33.3 21.6 32.4#
249 30.3 23.5 35.3









#23
Atrazine
Concen: 1.731 ng
RT: 16.774 min Scan# 1383
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 94705
Ion Ratio Lower Upper
200 100
173 21.5 20.3 30.5
215 50.0 40.4 60.6



#24
Pentachlorophenol
Concen: 1.807 ng
RT: 16.969 min Scan# 1399
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:266 Resp: 48108
Ion Ratio Lower Upper
266 100
264 62.6 49.5 74.3
268 64.0 50.6 76.0









#25
Phenanthrene
Concen: 1.589 ng
RT: 17.358 min Scan# 1431
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 668178
Ion Ratio Lower Upper
178 100
176 18.8 15.3 22.9
179 15.4 12.3 18.5



#26
Anthracene
Concen: 1.679 ng
RT: 17.443 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:178 Resp: 583413
Ion Ratio Lower Upper
178 100
176 18.1 14.8 22.2
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.794 ng
RT: 19.341 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 766495
Ion Ratio Lower Upper
212 100
106 13.1 8.4 12.6#
104 7.1 4.5 6.7#



#28
Fluoranthene
Concen: 1.636 ng
RT: 19.370 min Scan# 1758
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:202 Resp: 730509
Ion Ratio Lower Upper
202 100
101 11.1 7.2 10.8#
203 17.4 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.482 min Scan# 2026
Delta R.T. -0.011 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 112660
Ion Ratio Lower Upper
240 100
120 13.8 6.0 9.0#
236 26.9 21.4 32.2



#30
Pyrene
Concen: 1.895 ng
RT: 19.728 min Scan# 1830
Delta R.T. -0.005 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:202 Resp: 755649
Ion Ratio Lower Upper
202 100
200 20.5 16.6 25.0
203 17.5 14.1 21.1









#31
 Terphenyl-d14
 Concen: 2.236 ng
 RT: 19.936 min Scan# 1872
 Delta R.T. -0.000 min
 Lab File: BN018452.D
 Acq: 24 Jan 2022 15:42




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 545337
 Ion Ratio Lower Upper
 244 100
 212 7.3 6.2 9.2
 122 10.7 7.9 11.9



#32
 Benzo(a)anthracene
 Concen: 1.796 ng
 RT: 21.471 min Scan# 2024
 Delta R.T. -0.000 min
 Lab File: BN018452.D
 Acq: 24 Jan 2022 15:42

Tgt Ion:228 Resp: 596096
 Ion Ratio Lower Upper
 228 100
 226 26.1 21.3 31.9
 229 20.0 15.9 23.9









#33
Chrysene
Concen: 1.578 ng
RT: 21.524 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 594603
Ion Ratio Lower Upper
228 100
226 28.7 24.2 36.4
229 19.9 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 2.179 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.011 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:149 Resp: 433584
Ion Ratio Lower Upper
149 100
167 28.6 23.7 35.5
279 4.0 3.6 5.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.895 min Scan# 21
Delta R.T. -0.007 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264 Resp: 92955
Ion Ratio Lower Upper
264 100
260 24.8 20.2 30.4
265 44.4 18.2 27.2#



#36
Indeno(1,2,3-cd)pyrene
Concen: 1.015 ng
RT: 26.397 min Scan# 3278
Delta R.T. -0.011 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:276 Resp: 299833
Ion Ratio Lower Upper
276 100
138 24.5 14.9 22.3#
277 24.6 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 1.874 ng
RT: 23.159 min Scan# 21
Delta R.T. -0.011 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 564626
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 10.3 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 1.666 ng
RT: 23.210 min Scan# 2347
Delta R.T. -0.007 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:252 Resp: 490890
Ion Ratio Lower Upper
252 100
253 22.6 17.7 26.5
125 11.2 6.2 9.2#









#39
Benzo(a)pyrene
Concen: 1.521 ng
RT: 23.785 min Scan# 2515
Delta R.T. -0.011 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 423547
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.6
125 11.6 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.902 ng
RT: 26.421 min Scan# 3285
Delta R.T. -0.011 min
Lab File: BN018452.D
Acq: 24 Jan 2022 15:42

Tgt Ion:278 Resp: 199380
Ion Ratio Lower Upper
278 100
139 17.9 10.8 16.2#
279 24.8 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 1.100 ng
 RT: 27.164 min Scan# 31
 Delta R.T. -0.007 min
 Lab File: BN018452.D
 Acq: 24 Jan 2022 15:42

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:276 Resp: 297286

Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.0	28.6
138	22.0	12.6	19.0#

