

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012523\
 Data File : BN023795.D
 Acq On : 25 Jan 2023 15:26
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020396

Quant Time: Jan 25 22:42:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 00:53:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.881	152	225531	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	1009412	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	617182	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.280	188	1336652	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	1286814	20.000	ng/u1	0.01
88) Perylene-d12	23.898	264	1183792	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	40900	7.804	ng/uL	0.00
4) Pyridine-d5	3.687	84	302626	19.245	ng/u1	0.00
7) Phenol-d5	7.046	99	377043	18.863	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	236037	19.249	ng/u1	0.00
11) 2-Chlorophenol-d4	7.410	132	304783	19.910	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	294983	18.490	ng/u1	0.00
21) Nitrobenzene-d5	9.052	128	147354	19.359	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	161358	20.025	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.310	165	288095	18.745	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	428657	17.731	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	776091	16.904	ng/u1	0.00
49) Acenaphthylene-d8	14.228	160	969648	19.061	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	160551	17.178	ng/u1	0.00
60) Fluorene-d10	15.528	176	720977	18.378	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	141465	17.544	ng/u1	0.00
73) Anthracene-d10	17.380	188	1119767	18.721	ng/u1	0.00
81) Pyrene-d10	19.674	212	1318572	17.917	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	1094578	18.697	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	44214	7.992	ng/uL	96
5) Pyridine	3.711	79	305207	18.960	ng/u1	96
6) Benzaldehyde	7.016	77	168544	21.773	ng/u1	97
8) Phenol	7.075	94	381981	18.773	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.305	93	312381	19.431	ng/u1	100
12) 2-Chlorophenol	7.440	128	313433	19.813	ng/u1	98
13) 2-Methylphenol	8.328	108	284362	18.571	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.410	45	249053	11.583	ng/u1	99
16) Acetophenone	8.710	105	468657	19.006	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.699	70	242906	19.565	ng/u1	99
18) 4-Methylphenol	8.663	108	313166	18.572	ng/u1	97
19) Hexachloroethane	8.963	117	126672	20.020	ng/u1	98
22) Nitrobenzene	9.093	77	364537	19.136	ng/u1	99
23) Isophorone	9.622	82	664134	19.276	ng/u1	99
25) 2-Nitrophenol	9.810	139	173882	19.414	ng/u1	98
26) 2,4-Dimethylphenol	9.869	107	341294	18.776	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.110	93	418053	18.707	ng/u1	99
29) 2,4-Dichlorophenol	10.340	162	288244	18.743	ng/u1	97
30) Naphthalene	10.746	128	994610	18.329	ng/u1	99
32) 4-Chloroaniline	10.857	127	430061	17.762	ng/u1	99
33) Hexachlorobutadiene	11.028	225	181422	19.073	ng/u1	98
34) Caprolactam	11.640	113	80642	15.613	ng/u1	98
35) 4-Chloro-3-methylphenol	11.987	107	284378	16.959	ng/u1	97
36) 2-Methylnaphthalene	12.357	142	576611	15.940	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012523\
 Data File : BN023795.D
 Acq On : 25 Jan 2023 15:26
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020396

Quant Time: Jan 25 22:42:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jan 25 00:53:39 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.575	142	607113	16.428	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.722	216	322694	17.679	ng/ul	99
40) Hexachlorocyclopentadiene	12.698	237	204898	17.810	ng/ul	97
41) 2,4,6-Trichlorophenol	12.969	196	194031	16.126	ng/ul	95
42) 2,4,5-Trichlorophenol	13.034	196	209911	15.952	ng/ul	91
43) 1,1'-Biphenyl	13.369	154	744185	15.629	ng/ul	99
44) 2-Chloronaphthalene	13.410	162	585195	15.841	ng/ul	99
45) 2-Nitroaniline	13.622	65	179178	17.588	ng/ul	98
47) Dimethylphthalate	13.992	163	807249	17.498	ng/ul	99
48) 2,6-Dinitrotoluene	14.116	165	166902	19.137	ng/ul	98
50) Acenaphthylene	14.257	152	1060125	18.743	ng/ul	99
51) 3-Nitroaniline	14.445	138	161281	18.454	ng/ul	100
52) Acenaphthene	14.598	153	739870	18.553	ng/ul	95
53) 2,4-Dinitrophenol	14.651	184	83029	14.566	ng/ul	99
55) 4-Nitrophenol	14.751	109	131603	17.435	ng/ul	98
56) Dibenzofuran	14.934	168	1030613	18.567	ng/ul	100
57) 2,4-Dinitrotoluene	14.904	165	250462	19.671	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.157	232	209059	18.541	ng/ul	98
59) Diethylphthalate	15.357	149	874829	18.930	ng/ul	97
61) Fluorene	15.581	166	817112	18.179	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.575	204	404254	18.474	ng/ul	98
63) 4-Nitroaniline	15.610	138	149519	18.581	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.663	198	141888	17.831	ng/ul	98
67) N-Nitrosodiphenylamine	15.792	169	703104	18.744	ng/ul	98
68) 4-Bromophenyl-phenylether	16.469	248	249119	18.594	ng/ul	93
69) Hexachlorobenzene	16.586	284	288377	18.367	ng/ul	100
70) Atrazine	16.745	200	259087	18.455	ng/ul	99
71) Pentachlorophenol	16.933	266	177512	17.715	ng/ul	95
72) Phenanthrene	17.322	178	1311057	18.305	ng/ul	100
74) Anthracene	17.416	178	1332293	18.691	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.334	216	283110	15.732	ng/ul	97
76) Pentachlorobenzene	14.851	250	333289	18.489	ng/ul	94
77) Carbazole	17.692	167	1217148	18.806	ng/ul	97
78) Di-n-butylphthalate	18.251	149	1499117	19.509	ng/ul	100
80) Fluoranthene	19.339	202	1499922	17.425	ng/ul	98
82) Pyrene	19.704	202	1627911	17.905	ng/ul	98
83) Butylbenzylphthalate	20.604	149	698917	19.417	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.398	252	535978	19.067	ng/ul	96
85) Benzo(a)anthracene	21.463	228	1591993	18.264	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.386	149	1049471	20.081	ng/ul	99
87) Chrysene	21.516	228	1481296	18.125	ng/ul	97
89) Di-n-octyl phthalate	22.316	149	1700829	20.818	ng/ul	100
90) Benzo(b)fluoranthene	23.163	252	1492484	19.792	ng/ul	99
91) Benzo(k)fluoranthene	23.210	252	1395470	18.710	ng/ul	99
93) Benzo(a)pyrene	23.786	252	1198661	18.329	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.409	276	1425217	17.634	ng/ul	98
95) Dibenzo(a,h)anthracene	26.421	278	1158847	17.341	ng/ul	99
96) Benzo(g,h,i)perylene	27.180	276	1121065	17.417	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN012523\
Data File : BN023795.D
Acq On : 25 Jan 2023 15:26
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD020396

Quant Time: Jan 25 22:42:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN010523.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jan 25 00:53:39 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.992 ng/uL
RT: 3.305 min Scan# 31
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion: 88 Resp: 44214
Ion Ratio Lower Upper
88 100
43 28.0 24.6 36.8
58 67.1 56.2 84.4



#5
Pyridine
Concen: 18.960 ng/uL
RT: 3.711 min Scan# 106
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion: 79 Resp: 305207
Ion Ratio Lower Upper
79 100
52 69.7 59.9 89.9
51 35.2 28.6 42.8









#6
Benzaldehyde
Concen: 21.773 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion: 77 Resp: 168544
Ion Ratio Lower Upper
77 100
105 92.3 75.1 112.7
106 87.2 72.8 109.2



#8
Phenol
Concen: 18.773 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion: 94 Resp: 381981
Ion Ratio Lower Upper
94 100
65 29.4 23.7 35.5
66 36.9 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 19.431 ng/ul
RT: 7.305 min Scan# 718
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion: 93 Resp: 312381
Ion Ratio Lower Upper
93 100
63 76.3 61.2 91.8
95 32.9 26.6 40.0



#12
2-Chlorophenol
Concen: 19.813 ng/ul
RT: 7.440 min Scan# 740
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:128 Resp: 313433
Ion Ratio Lower Upper
128 100
64 45.0 37.4 56.2
130 31.2 25.6 38.4





#13

2-Methylphenol

Concen: 18.571 ng/ul

RT: 8.328 min Scan# 891

Delta R.T. -0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :





SSTD020396

Tgt Ion:108 Resp: 284362

Ion Ratio Lower Upper

108 100

107 90.5 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 11.583 ng/ul

RT: 8.410 min Scan# 905

Delta R.T. -0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion: 45 Resp: 249053


Ion Ratio Lower Upper

45 100

77 14.6 11.4 17.2

79 10.6 8.8 13.2









#16
Acetophenone
Concen: 19.006 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:105 Resp: 468657
Ion Ratio Lower Upper
105 100
77 80.5 64.2 96.2
51 27.3 22.0 33.0



#17
N-Nitroso-di-n-propylamine
Concen: 19.565 ng/ul
RT: 8.699 min Scan# 954
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion: 70 Resp: 242906
Ion Ratio Lower Upper
70 100
42 60.9 47.8 71.6
101 9.5 7.8 11.6
130 20.5 16.3 24.5









#18
4-Methylphenol
Concen: 18.572 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:108 Resp: 313166
Ion Ratio Lower Upper
108 100
107 110.5 90.7 136.1



#19
Hexachloroethane
Concen: 20.020 ng/ul
RT: 8.963 min Scan# 999
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:117 Resp: 126672
Ion Ratio Lower Upper
117 100
201 97.1 76.9 115.3
199 59.3 44.6 67.0





#22

Nitrobenzene

Concen: 19.136 ng/ul

RT: 9.093 min Scan# 1101

Delta R.T. -0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion: 77 Resp: 364537

Ion Ratio Lower Upper

77 100

123 44.8 35.6 53.4

65 12.9 11.0 16.4



#23

Isophorone

Concen: 19.276 ng/ul

RT: 9.622 min Scan# 1111

Delta R.T. -0.000 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion: 82 Resp: 664134


Ion Ratio Lower Upper

82 100

95 6.6 5.5 8.3

138 16.9 13.8 20.8









#25
2-Nitrophenol
Concen: 19.414 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:139 Resp: 173882
Ion Ratio Lower Upper
139 100
65 52.4 40.3 60.5
109 29.9 24.6 37.0



#26
2,4-Dimethylphenol
Concen: 18.776 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:107 Resp: 341294
Ion Ratio Lower Upper
107 100
121 51.1 40.4 60.6
122 86.8 67.0 100.4





#27

Bis(2-Chloroethoxy)methane

Concen: 18.707 ng/ul

RT: 10.110 min Scan# 1194

Delta R.T. -0.000 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion: 93 Resp: 418053

Ion Ratio Lower Upper

93 100

95 32.2 25.5 38.3

123 11.1 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 18.743 ng/ul

RT: 10.340 min Scan# 1233

Delta R.T. -0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion:162 Resp: 288244


Ion Ratio Lower Upper

162 100

164 64.8 53.8 80.6

98 35.0 28.8 43.2









#30
Naphthalene
Concen: 18.329 ng/ul
RT: 10.746 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:128 Resp: 994610
Ion Ratio Lower Upper
128 100
129 11.2 9.1 13.7
127 13.2 10.8 16.2



#32
4-Chloroaniline
Concen: 17.762 ng/ul
RT: 10.857 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:127 Resp: 430061
Ion Ratio Lower Upper
127 100
129 33.4 26.4 39.6









#33
Hexachlorobutadiene
Concen: 19.073 ng/ul
RT: 11.028 min Scan# 111
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:225 Resp: 181422
Ion Ratio Lower Upper
225 100
223 63.0 51.7 77.5
227 61.5 50.8 76.2



#34
Caprolactam
Concen: 15.613 ng/ul
RT: 11.640 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:113 Resp: 80642
Ion Ratio Lower Upper
113 100
55 162.8 129.1 193.7
56 121.3 94.2 141.4









#35
 4-Chloro-3-methylphenol
 Concen: 16.959 ng/ul
 RT: 11.987 min Scan# 1513
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020396


Tgt Ion:107 Resp: 284378
 Ion Ratio Lower Upper
 107 100
 144 26.1 21.9 32.9
 142 82.3 64.0 96.0



#36
 2-Methylnaphthalene
 Concen: 15.940 ng/ul
 RT: 12.357 min Scan# 1576
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26

Tgt Ion:142 Resp: 576611
 Ion Ratio Lower Upper
 142 100
 141 89.4 73.6 110.4









#37
1-Methylnaphthalene
Concen: 16.428 ng/ul
RT: 12.575 min Scan# 1613
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:142 Resp: 607113
Ion Ratio Lower Upper
142 100
141 94.5 73.8 110.6
116 4.0 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.679 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:216 Resp: 322694
Ion Ratio Lower Upper
216 100
214 76.9 62.4 93.6
179 21.4 16.2 24.2
108 18.7 14.6 21.8





#40

Hexachlorocyclopentadiene

Concen: 17.810 ng/ul

RT: 12.698 min Scan# 1634

Delta R.T. -0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:237 Resp: 204898

Ion Ratio Lower Upper

237 100

235 66.4 51.0 76.4

272 13.3 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 16.126 ng/ul

RT: 12.969 min Scan# 1680

Delta R.T. -0.000 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion:196 Resp: 194031


Ion Ratio Lower Upper

196 100

198 94.2 71.4 107.0

200 31.0 22.5 33.7









#42
2,4,5-Trichlorophenol
Concen: 15.952 ng/ul
RT: 13.034 min Scan# 1691
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:196 Resp: 209911
Ion Ratio Lower Upper
196 100
198 95.4 68.1 102.1
200 28.4 24.1 36.1



#43
1,1'-Biphenyl
Concen: 15.629 ng/ul
RT: 13.369 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:154 Resp: 744185
Ion Ratio Lower Upper
154 100
153 41.0 32.5 48.7
76 13.1 11.0 16.6









#44
2-Chloronaphthalene
Concen: 15.841 ng/ul
RT: 13.410 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:162 Resp: 585195
Ion Ratio Lower Upper
162 100
164 31.9 26.2 39.2
127 39.5 31.8 47.8



#45
2-Nitroaniline
Concen: 17.588 ng/ul
RT: 13.622 min Scan# 1791
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion: 65 Resp: 179178
Ion Ratio Lower Upper
65 100
92 68.0 53.5 80.3
138 102.2 84.0 126.0









#47
Dimethylphthalate
Concen: 17.498 ng/ul
RT: 13.992 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:163 Resp: 807249
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 19.137 ng/ul
RT: 14.116 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:165 Resp: 166902
Ion Ratio Lower Upper
165 100
89 64.3 49.9 74.9
121 23.2 19.1 28.7









#50
Acenaphthylene
Concen: 18.743 ng/ul
RT: 14.257 min Scan# 1899
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:152 Resp: 1060125
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 13.2 10.2 15.4



#51
3-Nitroaniline
Concen: 18.454 ng/ul
RT: 14.445 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:138 Resp: 161281
Ion Ratio Lower Upper
138 100
108 10.3 8.1 12.1
92 116.6 93.2 139.8









#52
 Acenaphthene
 Concen: 18.553 ng/ul
 RT: 14.598 min Scan# 1957
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020396

Tgt Ion:153 Resp: 739870




Ion	Ratio	Lower	Upper
153	100		
152	46.0	38.5	57.7
154	84.4	71.5	107.3




#53
 2,4-Dinitrophenol
 Concen: 14.566 ng/ul
 RT: 14.651 min Scan# 1966
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26

Tgt Ion:184 Resp: 83029

Ion	Ratio	Lower	Upper
184	100		
63	73.3	58.2	87.4
154	62.8	49.4	74.0









#55
4-Nitrophenol
Concen: 17.435 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:109 Resp: 131603
Ion Ratio Lower Upper
109 100
139 128.8 104.8 157.2
65 130.6 107.3 160.9



#56
Dibenzofuran
Concen: 18.567 ng/ul
RT: 14.934 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:168 Resp: 1030613
Ion Ratio Lower Upper
168 100
139 36.7 29.6 44.4









#57
2,4-Dinitrotoluene
Concen: 19.671 ng/ul
RT: 14.904 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:165 Resp: 250462
Ion Ratio Lower Upper
165 100
63 45.2 37.2 55.8
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.541 ng/ul
RT: 15.157 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:232 Resp: 209059
Ion Ratio Lower Upper
232 100
131 49.9 39.5 59.3
130 2.8 2.2 3.4
166 33.8 25.1 37.7





#59

Diethylphthalate

Concen: 18.930 ng/ul

RT: 15.357 min Scan# 2086

Delta R.T. -0.000 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:149 Resp: 874829

Ion Ratio Lower Upper

149 100

177 20.6 17.9 26.9

150 12.2 9.9 14.9



#61

Fluorene

Concen: 18.179 ng/ul

RT: 15.581 min Scan# 2124

Delta R.T. -0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion:166 Resp: 817112


Ion Ratio Lower Upper

166 100

165 99.5 77.8 116.6

167 14.0 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 18.474 ng/ul
RT: 15.575 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:204 Resp: 404254
Ion Ratio Lower Upper
204 100
206 34.9 27.0 40.6
141 62.4 51.0 76.4



#63
4-Nitroaniline
Concen: 18.581 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:138 Resp: 149519
Ion Ratio Lower Upper
138 100
92 52.7 41.3 61.9
108 85.5 64.2 96.2









#66
4,6-Dinitro-2-methylphenol
Concen: 17.831 ng/ul
RT: 15.663 min Scan# 2138
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:198 Resp: 141888
Ion Ratio Lower Upper
198 100
182 4.6 3.6 5.4
77 27.7 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 18.744 ng/ul
RT: 15.792 min Scan# 2160
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:169 Resp: 703104
Ion Ratio Lower Upper
169 100
168 68.5 53.4 80.2
167 37.2 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 18.594 ng/ul
RT: 16.469 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:248 Resp: 249119
Ion Ratio Lower Upper
248 100
250 100.1 76.4 114.6
141 80.8 58.0 87.0



#69
Hexachlorobenzene
Concen: 18.367 ng/ul
RT: 16.586 min Scan# 2295
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:284 Resp: 288377
Ion Ratio Lower Upper
284 100
142 36.1 28.9 43.3
249 30.2 24.2 36.2









#70
Atrazine
 Concen: 18.455 ng/ul
 RT: 16.745 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26




Instrument :
 BNA_N
ClientSampleId :
 SSTD020396


Tgt Ion:	200	Resp:	259087
Ion Ratio	Lower	Upper	
200	100		
173	26.7	21.4	32.2
215	48.2	39.0	58.6



#71
Pentachlorophenol
 Concen: 17.715 ng/ul
 RT: 16.933 min Scan# 2354
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26

Tgt Ion:	266	Resp:	177512
Ion Ratio	Lower	Upper	
266	100		
264	60.7	51.4	77.0
268	64.2	54.5	81.7









#72
 Phenanthrene
 Concen: 18.305 ng/ul
 RT: 17.322 min Scan# 2420
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020396

Tgt Ion:178 Resp: 1311057




Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.3	18.5
176	19.1	15.2	22.8




#74
 Anthracene
 Concen: 18.691 ng/ul
 RT: 17.416 min Scan# 2436
 Delta R.T. -0.000 min
 Lab File: BN023795.D
 Acq: 25 Jan 2023 15:26

Tgt Ion:178 Resp: 1332293

Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.6	19.0
176	19.1	15.1	22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 15.732 ng/uL
RT: 13.334 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:216 Resp: 283110
Ion Ratio Lower Upper
216 100
214 80.6 61.0 91.4
218 49.6 39.8 59.8



#76
Pentachlorobenzene
Concen: 18.489 ng/uL
RT: 14.851 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:250 Resp: 333289
Ion Ratio Lower Upper
250 100
248 66.2 50.3 75.5
252 67.4 48.9 73.3





#77

Carbazole

Concen: 18.806 ng/ul

RT: 17.692 min Scan# 2483

Delta R.T. 0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:167 Resp: 1217148

Ion Ratio Lower Upper

167 100

166 20.8 17.4 26.2

139 12.3 11.0 16.6



#78

Di-n-butylphthalate

Concen: 19.509 ng/ul

RT: 18.251 min Scan# 2578

Delta R.T. -0.000 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion:149 Resp: 1499117


Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.2

104 5.3 4.3 6.5









#80
Fluoranthene
Concen: 17.425 ng/ul
RT: 19.339 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:202 Resp: 1499922
Ion Ratio Lower Upper
202 100
101 12.9 9.5 14.3
100 9.8 7.3 10.9



#82
Pyrene
Concen: 17.905 ng/ul
RT: 19.704 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:202 Resp: 1627911
Ion Ratio Lower Upper
202 100
101 14.1 10.9 16.3
100 11.9 8.9 13.3





#83

Butylbenzylphthalate

Concen: 19.417 ng/ul

RT: 20.604 min Scan# 2978

Delta R.T. -0.000 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:149 Resp: 698917

Ion Ratio Lower Upper

149 100

91 77.1 57.9 86.9

206 19.8 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 19.067 ng/ul

RT: 21.398 min Scan# 3113

Delta R.T. 0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion:252 Resp: 535978


Ion Ratio Lower Upper

252 100

254 67.7 51.2 76.8

126 12.1 8.8 13.2









#85
Benzo(a)anthracene
Concen: 18.264 ng/ul
RT: 21.463 min Scan# 3124
Delta R.T. 0.012 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:228 Resp: 1591993
Ion Ratio Lower Upper
228 100
229 18.7 15.0 22.4
226 27.0 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.081 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. 0.006 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:149 Resp: 1049471
Ion Ratio Lower Upper
149 100
167 26.6 22.0 33.0
279 4.2 3.4 5.2





#87

Chrysene

Concen: 18.125 ng/ul

RT: 21.516 min Scan# 3133

Delta R.T. 0.006 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:228 Resp: 1481296

Ion Ratio Lower Upper

228 100

226 29.8 24.6 37.0

229 19.3 16.8 25.2



#89

Di-n-octyl phthalate

Concen: 20.818 ng/ul




RT: 22.316 min Scan# 3269


Delta R.T. 0.012 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Tgt Ion:149 Resp: 1700829









#90
Benzo(b)fluoranthene
Concen: 19.792 ng/ul
RT: 23.163 min Scan# 3410
Delta R.T. 0.023 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26




Instrument :
BNA_N
ClientSampleId :
SSTD020396


Tgt Ion:252 Resp: 1492484
Ion Ratio Lower Upper
252 100
253 21.9 17.2 25.8
125 10.6 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 18.710 ng/ul
RT: 23.210 min Scan# 3421
Delta R.T. 0.018 min
Lab File: BN023795.D
Acq: 25 Jan 2023 15:26

Tgt Ion:252 Resp: 1395470
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.8
125 10.2 8.6 12.8





#93

Benzo(a)pyrene

Concen: 18.329 ng/ul

RT: 23.786 min Scan# 3

Delta R.T. 0.018 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:252 Resp: 1198661

Ion Ratio Lower Upper

252 100

253 21.7 16.6 25.0

125 11.7 9.8 14.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.634 ng/ul

RT: 26.409 min Scan# 3965

Delta R.T. 0.041 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26




Tgt Ion:276 Resp: 1425217


Ion Ratio Lower Upper

276 100

138 24.8 20.6 30.8

277 24.4 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 17.341 ng/ul

RT: 26.421 min Scan# 3962

Delta R.T. 0.035 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Instrument :

BNA_N

ClientSampleId :

SSTD020396





Tgt Ion:278 Resp: 1158847

Ion Ratio Lower Upper

278 100

139 16.2 14.1 21.1

279 22.6 18.0 27.0



#96

Benzo(g,h,i)perylene

Concen: 17.417 ng/ul

RT: 27.180 min Scan# 4096

Delta R.T. 0.047 min

Lab File: BN023795.D

Acq: 25 Jan 2023 15:26

Tgt Ion:276 Resp: 1121065

Ion Ratio Lower Upper

276 100

138 22.2 17.0 25.6

277 21.5 17.6 26.4

