

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020122\
 Data File : BN018527.D
 Acq On : 01 Feb 2022 23:23
 Operator : CG/JU
 Sample : PB142390BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK390

Quant Time: Feb 02 05:28:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN013122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 01 03:59:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.927	152	387	0.400	ng/ul	0.00
4) Naphthalene-d8	10.733	136	1411	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.558	164	821	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.294	188	1564	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.468	240	1225	0.400	ng/ul	# 0.00
23) Perylene-d12	23.862	264	1315	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.295	96	2868	6.434	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.323	152	745	0.392	ng/ul	0.00
18) Fluoranthene-d10	19.314	212	1568	0.458	ng/ul	0.00
Target Compounds						
14) Pentachlorophenol	16.948	266	80	0.576	ng/ul	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020122\
Data File : BN018527.D
Acq On : 01 Feb 2022 23:23
Operator : CG/JU
Sample : PB142390BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK390

Quant Time: Feb 02 05:28:58 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN013122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 01 03:59:58 2022
Response via : Initial Calibration

