

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN020122\  
 Data File : BN018532.D  
 Acq On : 02 Feb 2022 02:20  
 Operator : CG/JU  
 Sample : N1307-14  
 Misc :  
 ALS Vial : 28 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 C0Q05

Quant Time: Feb 02 05:31:14 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN013122.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Feb 01 03:59:58 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|----------|--------|----------|
| -----                       |        |      |          |          |        |          |
| Internal Standards          |        |      |          |          |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.927  | 152  | 431      | 0.400    | ng/ul  | 0.00     |
| 4) Naphthalene-d8           | 10.733 | 136  | 1548     | 0.400    | ng/ul  | # 0.00   |
| 9) Acenaphthene-d10         | 14.558 | 164  | 988      | 0.400    | ng/ul  | # 0.00   |
| 13) Phenanthrene-d10        | 17.298 | 188  | 1827     | 0.400    | ng/ul  | # 0.00   |
| 17) Chrysene-d12            | 21.465 | 240  | 1658     | 0.400    | ng/ul  | # 0.00   |
| 23) Perylene-d12            | 23.859 | 264  | 1600     | 0.400    | ng/ul  | -0.01    |
| System Monitoring Compounds |        |      |          |          |        |          |
| 3) 1,4-Dioxane-d8           | 3.295  | 96   | 1683     | 3.390    | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.317 | 152  | 679      | 0.325    | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 19.319 | 212  | 1860     | 0.402    | ng/ul  | 0.00     |
| Target Compounds            |        |      |          |          |        |          |
|                             |        |      |          |          | Qvalue |          |
| 5) Naphthalene              | 10.788 | 128  | 298      | 0.070    | ng/ul# | 1        |
| 8) 1-Methylnaphthalene      | 12.609 | 142  | 285      | 0.103    | ng/ul  | 94       |
| 11) Acenaphthene            | 14.618 | 153  | 216      | 0.063    | ng/ul# | 88       |
| 12) Fluorene                | 15.604 | 166  | 2864     | 0.783    | ng/ul  | 99       |
| 14) Pentachlorophenol       | 16.952 | 266  | 214902   | 1323.665 | ng/ul  | 99       |
| 15) Phenanthrene            | 17.336 | 178  | 386      | 0.066    | ng/ul# | 50       |
| 16) Anthracene              | 17.429 | 178  | 366      | 0.075    | ng/ul# | 42       |
| 19) Fluoranthene            | 19.346 | 202  | 150      | 0.023    | ng/ul# | 8        |
| -----                       |        |      |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN020122\  
Data File : BN018532.D  
Acq On : 02 Feb 2022 02:20  
Operator : CG/JU  
Sample : N1307-14  
Misc :  
ALS Vial : 28 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
C0Q05

Quant Time: Feb 02 05:31:14 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN013122.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Feb 01 03:59:58 2022  
Response via : Initial Calibration

