

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023888.D
 Acq On : 01 Feb 2023 13:45
 Operator : CG/JU
 Sample : SSTD00559
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005224

Quant Time: Feb 02 01:16:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 01:11:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	151814	20.000	ng/ul	0.00
20) Naphthalene-d8	10.658	136	614938	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.504	164	410879	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.251	188	941437	20.000	ng/ul	0.00
79) Chrysene-d12	21.445	240	910459	20.000	ng/ul	0.00
88) Perylene-d12	23.851	264	879889	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	7686	2.064	ng/uL	0.00
4) Pyridine-d5	3.670	84	45242	4.305	ng/ul	0.00
7) Phenol-d5	7.017	99	57971	4.373	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	40255	4.805	ng/ul	0.00
11) 2-Chlorophenol-d4	7.381	132	47176	4.551	ng/ul	0.00
15) 4-Methylphenol-d8	8.569	113	45441	4.321	ng/ul	0.00
21) Nitrobenzene-d5	9.022	128	22508	4.511	ng/ul	0.00
24) 2-Nitrophenol-d4	9.740	143	25102	4.422	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	48701	4.594	ng/ul	0.00
31) 4-Chloroaniline-d4	10.805	131	70230	4.857	ng/ul	0.00
46) Dimethylphthalate-d6	13.916	166	154858	4.721	ng/ul	0.00
49) Acenaphthylene-d8	14.198	160	158920	4.478	ng/ul	0.00
54) 4-Nitrophenol-d4	14.710	143	24285	4.083	ng/ul	0.00
60) Fluorene-d10	15.498	176	132515	4.819	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	23267	3.493	ng/ul	0.00
73) Anthracene-d10	17.351	188	205826	4.696	ng/ul	0.00
81) Pyrene-d10	19.645	212	256206	4.635	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.686	264	214310	4.693	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.276	88	7431	1.918	ng/uL#	1
5) Pyridine	3.687	79	47323	4.481	ng/ul	95
6) Benzaldehyde	6.987	77	24794	4.301	ng/ul	96
8) Phenol	7.046	94	59850	4.491	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.275	93	49763	4.583	ng/ul	95
12) 2-Chlorophenol	7.411	128	47468	4.525	ng/ul	99
13) 2-Methylphenol	8.305	108	43704	4.322	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.375	45	67156	4.782	ng/ul#	91
16) Acetophenone	8.681	105	75933	4.662	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.663	70	41086	4.595	ng/ul	97
18) 4-Methylphenol	8.628	108	47610	4.347	ng/ul	96
19) Hexachloroethane	8.928	117	19613	4.464	ng/ul	96
22) Nitrobenzene	9.063	77	58209	4.647	ng/ul	99
23) Isophorone	9.587	82	108048	4.571	ng/ul	100
25) 2-Nitrophenol	9.775	139	25535	4.332	ng/ul	99
26) 2,4-Dimethylphenol	9.840	107	53875	4.615	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.075	93	69601	4.819	ng/ul	99
29) 2,4-Dichlorophenol	10.310	162	47372	4.631	ng/ul	98
30) Naphthalene	10.710	128	164484	4.939	ng/ul	99
32) 4-Chloroaniline	10.828	127	69757	4.847	ng/ul	97
33) Hexachlorobutadiene	10.987	225	34162	4.766	ng/ul	94
34) Caprolactam	11.599	113	13777	4.342	ng/ul#	81
35) 4-Chloro-3-methylphenol	11.957	107	50267	4.585	ng/ul	97
36) 2-Methylnaphthalene	12.322	142	107861	4.767	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023888.D
 Acq On : 01 Feb 2023 13:45
 Operator : CG/JU
 Sample : SST00559
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST005224

Quant Time: Feb 02 01:16:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 01:11:49 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.546	142	110926	4.792	ng/ul	93
39) 1,2,4,5-Tetrachloroben...	12.693	216	66675	4.723	ng/ul	95
40) Hexachlorocyclopentadiene	12.663	237	39672	3.903	ng/ul	91
41) 2,4,6-Trichlorophenol	12.934	196	41898	4.498	ng/ul	100
42) 2,4,5-Trichlorophenol	13.004	196	45877	4.547	ng/ul	96
43) 1,1'-Biphenyl	13.334	154	158857	4.906	ng/ul	97
44) 2-Chloronaphthalene	13.375	162	118422	4.701	ng/ul	99
45) 2-Nitroaniline	13.593	65	33178	4.386	ng/ul	94
47) Dimethylphthalate	13.963	163	155169	4.766	ng/ul	99
48) 2,6-Dinitrotoluene	14.087	165	26484	4.103	ng/ul	91
50) Acenaphthylene	14.222	152	178014	4.671	ng/ul	99
51) 3-Nitroaniline	14.416	138	24368	4.194	ng/ul	93
52) Acenaphthene	14.569	153	126247	4.772	ng/ul	96
53) 2,4-Dinitrophenol	14.622	184	15202	3.350	ng/ul	95
55) 4-Nitrophenol	14.728	109	22211	4.393	ng/ul	97
56) Dibenzofuran	14.904	168	183141	4.852	ng/ul	100
57) 2,4-Dinitrotoluene	14.869	165	37998	4.161	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.128	232	41438	4.605	ng/ul#	97
59) Diethylphthalate	15.328	149	156072	4.825	ng/ul	99
61) Fluorene	15.551	166	155413	5.035	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.545	204	80828	4.992	ng/ul	98
63) 4-Nitroaniline	15.575	138	17652	3.531	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.634	198	25185	3.927	ng/ul#	97
67) N-Nitrosodiphenylamine	15.763	169	128977	4.745	ng/ul	99
68) 4-Bromophenyl-phenylether	16.440	248	50958	4.622	ng/ul	96
69) Hexachlorobenzene	16.551	284	60656	4.762	ng/ul	92
70) Atrazine	16.716	200	49655	4.546	ng/ul	97
71) Pentachlorophenol	16.904	266	35036	4.296	ng/ul	98
72) Phenanthrene	17.292	178	244198	4.816	ng/ul	99
74) Anthracene	17.387	178	239203	4.733	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.299	216	65867	4.548	ng/ul	97
76) Pentachlorobenzene	14.822	250	67535	4.676	ng/ul	94
77) Carbazole	17.663	167	215062	4.856	ng/ul	99
78) Di-n-butylphthalate	18.222	149	980369	10.329	ng/ul	100
80) Fluoranthene	19.310	202	293895	4.554	ng/ul	100
82) Pyrene	19.675	202	311697	4.749	ng/ul	98
83) Butylbenzylphthalate	20.581	149	120739	4.411	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.369	252	92314	4.317	ng/ul	96
85) Benzo(a)anthracene	21.433	228	295862	4.689	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.357	149	182308	4.514	ng/ul	95
87) Chrysene	21.486	228	279832	4.807	ng/ul	96
89) Di-n-octyl phthalate	22.280	149	287576	4.344	ng/ul	100
90) Benzo(b)fluoranthene	23.116	252	292087	4.813	ng/ul	98
91) Benzo(k)fluoranthene	23.163	252	275300	4.804	ng/ul	98
93) Benzo(a)pyrene	23.739	252	242283	4.746	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.327	276	296671	4.817	ng/ul	99
95) Dibenzo(a,h)anthracene	26.351	278	243948	4.788	ng/ul	99
96) Benzo(g,h,i)perylene	27.080	276	227173	4.683	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023888.D
 Acq On : 01 Feb 2023 13:45
 Operator : CG/JU
 Sample : SSTD00559
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD005224

Quant Time: Feb 02 01:16:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 01:11:49 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 1.918 ng/uL
 RT: 3.276 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45




Instrument :
 BNA_N
 ClientSampleId :
 SST005224


Tgt Ion: 88 Resp: 7431
 Ion Ratio Lower Upper
 88 100
 43 222.7 42.1 63.1#
 58 62.1 53.8 80.6



#5
 Pyridine
 Concen: 4.481 ng/uL
 RT: 3.687 min Scan# 102
 Delta R.T. 0.006 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45

Tgt Ion: 79 Resp: 47323
 Ion Ratio Lower Upper
 79 100
 52 72.6 56.4 84.6
 51 38.2 26.6 39.8









#6
Benzaldehyde
Concen: 4.301 ng/ul
RT: 6.987 min Scan# 664
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion: 77 Resp: 24794
Ion Ratio Lower Upper
77 100
105 94.8 73.6 110.4
106 83.6 70.2 105.4



#8
Phenol
Concen: 4.491 ng/ul
RT: 7.046 min Scan# 673
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion: 94 Resp: 59850
Ion Ratio Lower Upper
94 100
65 28.3 24.1 36.1
66 42.6 32.4 48.6





#10

Bis(2-Chloroethyl)ether

Concen: 4.583 ng/ul

RT: 7.275 min Scan# 712

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion: 93 Resp: 49763

Ion Ratio Lower Upper

93 100

63 80.5 60.6 91.0

95 34.1 26.2 39.4



#12

2-Chlorophenol

Concen: 4.525 ng/ul

RT: 7.411 min Scan# 735

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:128 Resp: 47468


Ion Ratio Lower Upper

128 100

64 49.3 38.4 57.6

130 31.9 25.6 38.4





#13

2-Methylphenol

Concen: 4.322 ng/ul

RT: 8.305 min Scan# 887

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :





SSTD005224

Tgt Ion:108 Resp: 43704

Ion Ratio Lower Upper

108 100

107 92.6 72.6 109.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.782 ng/ul

RT: 8.375 min Scan# 899

Delta R.T. -0.018 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion: 45 Resp: 67156


Ion Ratio Lower Upper

45 100

77 12.1 12.7 19.1#

79 9.0 9.8 14.8#









#16
Acetophenone
Concen: 4.662 ng/ul
RT: 8.681 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:105 Resp: 75933
Ion Ratio Lower Upper
105 100
77 78.3 64.7 97.1
51 27.3 21.7 32.5



#17
N-Nitroso-di-n-propylamine
Concen: 4.595 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion: 70 Resp: 41086
Ion Ratio Lower Upper
70 100
42 62.0 47.6 71.4
101 9.5 7.4 11.0
130 18.4 15.1 22.7









#18
4-Methylphenol
Concen: 4.347 ng/ul
RT: 8.628 min Scan# 944
Delta R.T. -0.012 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:108 Resp: 47610
Ion Ratio Lower Upper
108 100
107 115.9 88.9 133.3



#19
Hexachloroethane
Concen: 4.464 ng/ul
RT: 8.928 min Scan# 993
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:117 Resp: 19613
Ion Ratio Lower Upper
117 100
201 117.0 88.8 133.2
199 70.7 55.0 82.4









#22
Nitrobenzene
Concen: 4.647 ng/ul
RT: 9.063 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion: 77 Resp: 58209
Ion Ratio Lower Upper
77 100
123 42.0 34.2 51.2
65 13.7 10.9 16.3



#23
Isophorone
Concen: 4.571 ng/ul
RT: 9.587 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion: 82 Resp: 108048
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 16.6 13.4 20.2





#25

2-Nitrophenol

Concen: 4.332 ng/ul

RT: 9.775 min Scan# 1137

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:139 Resp: 25535

Ion Ratio Lower Upper

139 100

65 54.9 44.5 66.7

109 31.2 26.0 39.0



#26

2,4-Dimethylphenol

Concen: 4.615 ng/ul

RT: 9.840 min Scan# 1148

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:107 Resp: 53875


Ion Ratio Lower Upper

107 100

121 53.0 39.4 59.2

122 86.8 65.8 98.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.819 ng/ul
 RT: 10.075 min Scan# 1188
 Delta R.T. -0.006 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45




Instrument :
 BNA_N
 ClientSampleId :
 SST005224


Tgt Ion: 93 Resp: 69601
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.4 38.2
 123 11.6 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 4.631 ng/ul
 RT: 10.310 min Scan# 1228
 Delta R.T. 0.000 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45

Tgt Ion: 162 Resp: 47372
 Ion Ratio Lower Upper
 162 100
 164 65.8 51.4 77.0
 98 35.8 28.0 42.0





#30

Naphthalene

Concen: 4.939 ng/ul

RT: 10.710 min Scan# 1296

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:128 Resp: 164484

Ion Ratio Lower Upper

128 100

129 10.2 8.5 12.7

127 12.9 10.3 15.5



#32

4-Chloroaniline

Concen: 4.847 ng/ul

RT: 10.828 min Scan# 1316

Delta R.T. -0.006 min

Lab File: BN023888.D




Acq: 01 Feb 2023 13:45


Tgt Ion:127 Resp: 69757

Ion Ratio Lower Upper

127 100

129 30.9 25.9 38.9





#33

Hexachlorobutadiene

Concen: 4.766 ng/ul

RT: 10.987 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:225 Resp: 34162

Ion Ratio Lower Upper

225 100

223 59.7 50.2 75.4

227 58.2 52.2 78.2



#34

Caprolactam

Concen: 4.342 ng/ul

RT: 11.599 min Scan# 1447

Delta R.T. -0.018 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:113 Resp: 13777


Ion Ratio Lower Upper

113 100

55 176.2 125.6 188.4

56 143.8 94.1 141.1#









#35
4-Chloro-3-methylphenol
Concen: 4.585 ng/ul
RT: 11.957 min Scan# 1508
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:107 Resp: 50267
Ion Ratio Lower Upper
107 100
144 25.0 21.7 32.5
142 82.2 63.7 95.5



#36
2-Methylnaphthalene
Concen: 4.767 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:142 Resp: 107861
Ion Ratio Lower Upper
142 100
141 90.0 71.3 106.9









#37
1-Methylnaphthalene
Concen: 4.792 ng/ul
RT: 12.546 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:142 Resp: 110926
Ion Ratio Lower Upper
142 100
141 96.8 72.2 108.2
116 3.9 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.723 ng/ul
RT: 12.693 min Scan# 1633
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:216 Resp: 66675
Ion Ratio Lower Upper
216 100
214 73.8 63.0 94.4
179 18.3 16.5 24.7
108 15.1 13.3 19.9









#40
Hexachlorocyclopentadiene
Concen: 3.903 ng/ul
RT: 12.663 min Scan# 1628
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:237 Resp: 39672
Ion Ratio Lower Upper
237 100
235 58.0 53.0 79.4
272 14.1 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 4.498 ng/ul
RT: 12.934 min Scan# 1674
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:196 Resp: 41898
Ion Ratio Lower Upper
196 100
198 96.4 76.8 115.2
200 31.5 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 4.547 ng/ul
RT: 13.004 min Scan# 1
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:196 Resp: 45877
Ion Ratio Lower Upper
196 100
198 100.4 77.6 116.4
200 33.3 24.6 37.0



#43
1,1'-Biphenyl
Concen: 4.906 ng/ul
RT: 13.334 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:154 Resp: 158857
Ion Ratio Lower Upper
154 100
153 38.0 32.6 48.8
76 13.7 10.7 16.1









#44
2-Chloronaphthalene
Concen: 4.701 ng/ul
RT: 13.375 min Scan# 1749
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:162 Resp: 118422
Ion Ratio Lower Upper
162 100
164 31.8 25.9 38.9
127 38.9 30.7 46.1



#45
2-Nitroaniline
Concen: 4.386 ng/ul
RT: 13.593 min Scan# 1786
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion: 65 Resp: 33178
Ion Ratio Lower Upper
65 100
92 71.9 53.2 79.8
138 102.6 78.3 117.5









#47
Dimethylphthalate
Concen: 4.766 ng/ul
RT: 13.963 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:163 Resp: 155169
Ion Ratio Lower Upper
163 100
194 4.6 3.5 5.3
164 10.6 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 4.103 ng/ul
RT: 14.087 min Scan# 1870
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:165 Resp: 26484
Ion Ratio Lower Upper
165 100
89 71.4 50.8 76.2
121 23.0 16.8 25.2









#50
Acenaphthylene
Concen: 4.671 ng/ul
RT: 14.222 min Scan# 1893
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:152 Resp: 178014
Ion Ratio Lower Upper
152 100
151 19.7 16.3 24.5
153 13.3 10.3 15.5



#51
3-Nitroaniline
Concen: 4.194 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:138 Resp: 24368
Ion Ratio Lower Upper
138 100
108 12.2 9.1 13.7
92 130.8 97.6 146.4









#52
Acenaphthene
Concen: 4.772 ng/ul
RT: 14.569 min Scan# 1952
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:153 Resp: 126247
Ion Ratio Lower Upper
153 100
152 45.7 38.9 58.3
154 87.0 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 3.350 ng/ul
RT: 14.622 min Scan# 1961
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:184 Resp: 15202
Ion Ratio Lower Upper
184 100
63 73.6 53.2 79.8
154 64.0 50.6 76.0









#55
4-Nitrophenol
Concen: 4.393 ng/ul
RT: 14.728 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:109 Resp: 22211
Ion Ratio Lower Upper
109 100
139 125.1 98.1 147.1
65 130.9 101.6 152.4



#56
Dibenzofuran
Concen: 4.852 ng/ul
RT: 14.904 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:168 Resp: 183141
Ion Ratio Lower Upper
168 100
139 36.1 29.0 43.6





#57

2,4-Dinitrotoluene

Concen: 4.161 ng/ul

RT: 14.869 min Scan# 2003

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:165 Resp: 37998

Ion Ratio Lower Upper

165 100

63 52.9 40.7 61.1

182 3.5 2.3 3.5#



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.605 ng/ul

RT: 15.128 min Scan# 2047

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Tgt Ion:232 Resp: 41438




Ion Ratio Lower Upper


232 100

131 41.4 32.0 48.0

130 3.1 1.8 2.6#

166 28.1 21.0 31.4





#59

Diethylphthalate

Concen: 4.825 ng/ul

RT: 15.328 min Scan# 2081

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:149 Resp: 156072

Ion Ratio Lower Upper

149 100

177 21.8 17.8 26.6

150 12.3 9.4 14.2



#61

Fluorene

Concen: 5.035 ng/ul

RT: 15.551 min Scan# 2119

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:166 Resp: 155413


Ion Ratio Lower Upper

166 100

165 96.6 74.1 111.1

167 14.1 10.6 16.0





#62

4-Chlorophenyl-phenylether

Concen: 4.992 ng/ul

RT: 15.545 min Scan# 2118

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:204 Resp: 80828

Ion Ratio Lower Upper

204 100

206 32.1 27.2 40.8

141 57.1 45.2 67.8



#63

4-Nitroaniline

Concen: 3.531 ng/ul

RT: 15.575 min Scan# 2123

Delta R.T. -0.012 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:138 Resp: 17652


Ion Ratio Lower Upper

138 100

92 54.1 41.5 62.3

108 97.7 71.0 106.6









#66
4,6-Dinitro-2-methylphenol
Concen: 3.927 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:198 Resp: 25185
Ion Ratio Lower Upper
198 100
182 5.3 3.2 4.8#
77 28.0 21.1 31.7



#67
N-Nitrosodiphenylamine
Concen: 4.745 ng/ul
RT: 15.763 min Scan# 2155
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:169 Resp: 128977
Ion Ratio Lower Upper
169 100
168 65.8 52.2 78.2
167 35.5 27.3 40.9









#68
4-Bromophenyl-phenylether
Concen: 4.622 ng/ul
RT: 16.440 min Scan# 211
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:248 Resp: 50958
Ion Ratio Lower Upper
248 100
250 97.0 79.2 118.8
141 67.9 50.4 75.6



#69
Hexachlorobenzene
Concen: 4.762 ng/ul
RT: 16.551 min Scan# 2289
Delta R.T. -0.006 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:284 Resp: 60656
Ion Ratio Lower Upper
284 100
142 34.4 23.2 34.8
249 31.1 22.6 34.0









#70
 Atrazine
 Concen: 4.546 ng/ul
 RT: 16.716 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45




Instrument :
 BNA_N
 ClientSampleId :
 SSTD005224


Tgt Ion:200 Resp: 49655
 Ion Ratio Lower Upper
 200 100
 173 26.3 20.9 31.3
 215 47.3 40.2 60.2



#71
 Pentachlorophenol
 Concen: 4.296 ng/ul
 RT: 16.904 min Scan# 2349
 Delta R.T. 0.000 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45

Tgt Ion:266 Resp: 35036
 Ion Ratio Lower Upper
 266 100
 264 60.3 49.9 74.9
 268 62.2 50.3 75.5





#72

Phenanthrene

Concen: 4.816 ng/ul

RT: 17.292 min Scan# 2416

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:178 Resp: 244198

Ion Ratio Lower Upper

178 100

179 15.6 12.1 18.1

176 18.3 15.2 22.8



#74

Anthracene

Concen: 4.733 ng/ul

RT: 17.387 min Scan# 2431

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:178 Resp: 239203


Ion Ratio Lower Upper

178 100

179 15.6 11.8 17.8

176 18.6 14.4 21.6









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 4.548 ng/uL
 RT: 13.299 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45




Instrument :
 BNA_N
 ClientSampleId :
 SST005224


Tgt Ion:	216	Resp:	65867
Ion Ratio	Lower	Upper	
216	100		
214	75.5	61.9	92.9
218	47.6	36.0	54.0



#76
 Pentachlorobenzene
 Concen: 4.676 ng/uL
 RT: 14.822 min Scan# 1995
 Delta R.T. 0.000 min
 Lab File: BN023888.D
 Acq: 01 Feb 2023 13:45

Tgt Ion:	250	Resp:	67535
Ion Ratio	Lower	Upper	
250	100		
248	58.0	51.2	76.8
252	66.4	50.8	76.2





#77

Carbazole

Concen: 4.856 ng/ul

RT: 17.663 min Scan# 2478

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:167 Resp: 215062

Ion Ratio Lower Upper

167 100

166 21.3 17.4 26.2

139 12.3 10.2 15.4



#78

Di-n-butylphthalate

Concen: 10.329 ng/ul

RT: 18.222 min Scan# 2573

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:149 Resp: 980369


Ion Ratio Lower Upper

149 100

150 9.4 7.4 11.2

104 5.6 4.6 6.8





#80

Fluoranthene

Concen: 4.554 ng/ul

RT: 19.310 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:202 Resp: 293895

Ion Ratio Lower Upper

202 100

101 10.2 8.2 12.2

100 7.9 6.2 9.4



#82

Pyrene

Concen: 4.749 ng/ul

RT: 19.675 min Scan# 2820

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:202 Resp: 311697


Ion Ratio Lower Upper

202 100

101 12.0 9.1 13.7

100 10.1 7.4 11.2









#83
Butylbenzylphthalate
Concen: 4.411 ng/ul
RT: 20.581 min Scan# 2974
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45




Instrument :
BNA_N
ClientSampleId :
SSTD005224


Tgt Ion:149 Resp: 120739
Ion Ratio Lower Upper
149 100
91 72.5 57.9 86.9
206 23.6 19.0 28.4



#84
3,3'-Dichlorobenzidine
Concen: 4.317 ng/ul
RT: 21.369 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BN023888.D
Acq: 01 Feb 2023 13:45

Tgt Ion:252 Resp: 92314
Ion Ratio Lower Upper
252 100
254 69.4 53.0 79.6
126 10.7 8.1 12.1





#85

Benzo(a)anthracene

Concen: 4.689 ng/ul

RT: 21.433 min Scan# 3119

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:228 Resp: 295862

Ion Ratio Lower Upper

228 100

229 21.3 15.6 23.4

226 29.4 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.514 ng/ul

RT: 21.357 min Scan# 3106

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:149 Resp: 182308


Ion Ratio Lower Upper

149 100

167 29.6 21.2 31.8

279 5.3 3.8 5.6





#87

Chrysene

Concen: 4.807 ng/ul

RT: 21.486 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:228 Resp: 279832

Ion Ratio Lower Upper

228 100

226 29.3 25.9 38.9

229 19.5 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 4.344 ng/ul




RT: 22.280 min Scan# 3263


Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Tgt Ion:149 Resp: 287576





#90

Benzo(b)fluoranthene

Concen: 4.813 ng/ul

RT: 23.116 min Scan# 3405

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:252 Resp: 292087

Ion Ratio Lower Upper

252 100

253 21.3 17.8 26.8

125 9.0 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 4.804 ng/ul

RT: 23.163 min Scan# 3413

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:252 Resp: 275300


Ion Ratio Lower Upper

252 100

253 21.3 16.2 24.2

125 8.9 6.6 9.8





#93

Benzo(a)pyrene

Concen: 4.746 ng/ul

RT: 23.739 min Scan# 3

Delta R.T. 0.000 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:252 Resp: 242283

Ion Ratio Lower Upper

252 100

253 22.8 17.1 25.7

125 10.0 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.817 ng/ul

RT: 26.327 min Scan# 3951

Delta R.T. -0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45




Tgt Ion:276 Resp: 296671


Ion Ratio Lower Upper

276 100

138 20.8 16.6 25.0

277 25.5 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 4.788 ng/ul

RT: 26.351 min Scan# 3955

Delta R.T. 0.006 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Instrument :

BNA_N

ClientSampleId :

SSTD005224





Tgt Ion:278 Resp: 243948

Ion Ratio Lower Upper

278 100

139 14.1 10.8 16.2

279 23.6 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 4.683 ng/ul

RT: 27.080 min Scan# 4079

Delta R.T. -0.023 min

Lab File: BN023888.D

Acq: 01 Feb 2023 13:45

Tgt Ion:276 Resp: 227173

Ion Ratio Lower Upper

276 100

138 19.0 14.6 21.8

277 25.0 18.8 28.2

