

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023892.D
 Acq On : 01 Feb 2023 16:11
 Operator : CG/JU
 Sample : SSTD08063
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228

Quant Time: Feb 02 01:18:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 01:11:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.851	152	157888	20.000	ng/u1	0.00
20) Naphthalene-d8	10.669	136	649350	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	401790	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	876444	20.000	ng/u1	0.00
79) Chrysene-d12	21.457	240	655083	20.000	ng/u1	0.00
88) Perylene-d12	23.857	264	596698	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	115269	29.761	ng/uL	0.00
4) Pyridine-d5	3.663	84	854465	78.174	ng/u1	0.00
7) Phenol-d5	7.034	99	1103136	80.008	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.193	67	672170	77.141	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	880499	81.674	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	865470	79.129	ng/u1	0.02
21) Nitrobenzene-d5	9.034	128	442083	83.911	ng/u1	0.01
24) 2-Nitrophenol-d4	9.757	143	507728	84.698	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.298	165	914974	81.744	ng/u1	0.01
31) 4-Chloroaniline-d4	10.816	131	1175325	76.969	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	2533735	78.987	ng/u1	0.01
49) Acenaphthylene-d8	14.204	160	2819261	81.246	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	465398	80.024	ng/u1	0.02
60) Fluorene-d10	15.504	176	2076044	77.199	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	513391	82.779	ng/u1	0.02
73) Anthracene-d10	17.363	188	3179510	77.923	ng/u1	0.01
81) Pyrene-d10	19.657	212	3364501	84.595	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.698	264	2528957	81.659	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.275	88	120789	29.975	ng/uL#	87
5) Pyridine	3.681	79	855871	77.928	ng/u1	98
6) Benzaldehyde	6.993	77	477510	79.647	ng/u1	99
8) Phenol	7.063	94	1089132	78.577	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.287	93	880115	77.938	ng/u1	100
12) 2-Chlorophenol	7.422	128	887468	81.349	ng/u1	99
13) 2-Methylphenol	8.316	108	835268	79.424	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.399	45	1102585	75.493	ng/u1	99
16) Acetophenone	8.698	105	1295107	76.457	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.698	70	701783	75.468	ng/u1	97
18) 4-Methylphenol	8.657	108	891950	78.300	ng/u1	100
19) Hexachloroethane	8.934	117	371326	81.271	ng/u1	96
22) Nitrobenzene	9.075	77	1054089	79.698	ng/u1	98
23) Isophorone	9.616	82	1993357	79.852	ng/u1	100
25) 2-Nitrophenol	9.787	139	523377	84.076	ng/u1	97
26) 2,4-Dimethylphenol	9.851	107	985013	79.904	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.087	93	1180963	77.436	ng/u1	98
29) 2,4-Dichlorophenol	10.322	162	876467	81.146	ng/u1	99
30) Naphthalene	10.722	128	2738909	77.884	ng/u1	100
32) 4-Chloroaniline	10.840	127	1170589	77.025	ng/u1	99
33) Hexachlorobutadiene	10.998	225	615025	81.258	ng/u1	99
34) Caprolactam	11.657	113	165066	49.264	ng/u1	99
35) 4-Chloro-3-methylphenol	11.975	107	917075	79.211	ng/u1	98
36) 2-Methylnaphthalene	12.334	142	1861733	77.914	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023892.D
 Acq On : 01 Feb 2023 16:11
 Operator : CG/JU
 Sample : SSTD08063
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080228

Quant Time: Feb 02 01:18:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 01:11:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.551	142	1874124	76.664	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.704	216	1132476	82.031	ng/ul	98
40) Hexachlorocyclopentadiene	12.675	237	839611	84.474	ng/ul	97
41) 2,4,6-Trichlorophenol	12.945	196	754760	82.861	ng/ul	94
42) 2,4,5-Trichlorophenol	13.022	196	814721	82.575	ng/ul	95
43) 1,1'-Biphenyl	13.345	154	2480495	78.343	ng/ul	99
44) 2-Chloronaphthalene	13.386	162	1968942	79.921	ng/ul	99
45) 2-Nitroaniline	13.604	65	616810	83.383	ng/ul	95
47) Dimethylphthalate	13.981	163	2498327	78.477	ng/ul	100
48) 2,6-Dinitrotoluene	14.104	165	553423	87.683	ng/ul	97
50) Acenaphthylene	14.233	152	2933028	78.697	ng/ul	99
51) 3-Nitroaniline	14.433	138	461683	81.259	ng/ul	94
52) Acenaphthene	14.575	153	2028114	78.391	ng/ul	94
53) 2,4-Dinitrophenol	14.639	184	389708	87.831	ng/ul	96
55) 4-Nitrophenol	14.757	109	392749	79.439	ng/ul	98
56) Dibenzofuran	14.916	168	2839792	76.942	ng/ul	99
57) 2,4-Dinitrotoluene	14.892	165	742779	83.179	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.139	232	703068	79.903	ng/ul	97
59) Diethylphthalate	15.345	149	2458550	77.721	ng/ul	99
61) Fluorene	15.563	166	2223696	73.672	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.551	204	1174104	74.148	ng/ul	96
63) 4-Nitroaniline	15.604	138	390094	79.798	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.657	198	496713	83.184	ng/ul	94
67) N-Nitrosodiphenylamine	15.775	169	1985791	78.469	ng/ul	98
68) 4-Bromophenyl-phenylether	16.451	248	800067	77.943	ng/ul	95
69) Hexachlorobenzene	16.563	284	916671	77.298	ng/ul	92
70) Atrazine	16.733	200	781714	76.878	ng/ul	98
71) Pentachlorophenol	16.910	266	612049	80.609	ng/ul	100
72) Phenanthrene	17.304	178	3605472	76.382	ng/ul	100
74) Anthracene	17.398	178	3594856	76.410	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.310	216	1128256	83.673	ng/ul	97
76) Pentachlorobenzene	14.828	250	1075226	79.960	ng/ul	96
77) Carbazole	17.669	167	3145996	76.309	ng/ul	99
78) Di-n-butylphthalate	18.227	149	4182045	47.327	ng/ul	99
80) Fluoranthene	19.321	202	3997475	86.096	ng/ul	100
82) Pyrene	19.686	202	3950008	83.640	ng/ul	98
83) Butylbenzylphthalate	20.586	149	1689105	85.758	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.374	252	1145787	74.471	ng/ul	99
85) Benzo(a)anthracene	21.439	228	3511914	77.361	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.363	149	2366468	81.442	ng/ul	95
87) Chrysene	21.492	228	3190524	76.167	ng/ul	99
89) Di-n-octyl phthalate	22.286	149	3887123	86.592	ng/ul	100
90) Benzo(b)fluoranthene	23.127	252	3273573	79.548	ng/ul	97
91) Benzo(k)fluoranthene	23.174	252	3191317	82.121	ng/ul	99
93) Benzo(a)pyrene	23.757	252	2816404	81.346	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.339	276	1980372	47.412	ng/ul	98
95) Dibenzo(a,h)anthracene	26.368	278	2922904	84.592	ng/ul	97
96) Benzo(g,h,i)perylene	27.127	276	2913900	88.577	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 29.975 ng/uL
 RT: 3.275 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11




Instrument :
 BNA_N
 ClientSampleId :
 SST080228


Tgt Ion: 88 Resp: 120789
 Ion Ratio Lower Upper
 88 100
 43 31.8 42.1 63.1#
 58 67.6 53.8 80.6



#5
 Pyridine
 Concen: 77.928 ng/uL
 RT: 3.681 min Scan# 101
 Delta R.T. -0.000 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11

Tgt Ion: 79 Resp: 855871
 Ion Ratio Lower Upper
 79 100
 52 68.4 56.4 84.6
 51 33.3 26.6 39.8









#6
Benzaldehyde
Concen: 79.647 ng/ul
RT: 6.993 min Scan# 664
Delta R.T. -0.000 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion: 77 Resp: 477510
Ion Ratio Lower Upper
77 100
105 92.0 73.6 110.4
106 89.4 70.2 105.4



#8
Phenol
Concen: 78.577 ng/ul
RT: 7.063 min Scan# 676
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion: 94 Resp: 1089132
Ion Ratio Lower Upper
94 100
65 29.6 24.1 36.1
66 39.6 32.4 48.6









#10
Bis(2-Chloroethyl)ether
Concen: 77.938 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion: 93 Resp: 880115
Ion Ratio Lower Upper
93 100
63 75.5 60.6 91.0
95 32.3 26.2 39.4



#12
2-Chlorophenol
Concen: 81.349 ng/ul
RT: 7.422 min Scan# 737
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:128 Resp: 887468
Ion Ratio Lower Upper
128 100
64 46.9 38.4 57.6
130 32.6 25.6 38.4





#13

2-Methylphenol

Concen: 79.424 ng/ul

RT: 8.316 min Scan# 887

Delta R.T. 0.012 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :





SSTD080228

Tgt Ion:108 Resp: 835268

Ion Ratio Lower Upper

108 100

107 88.6 72.6 109.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 75.493 ng/ul

RT: 8.399 min Scan# 903

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion: 45 Resp: 1102585


Ion Ratio Lower Upper

45 100

77 16.6 12.7 19.1

79 12.5 9.8 14.8









#16
Acetophenone
Concen: 76.457 ng/ul
RT: 8.698 min Scan# 91
Delta R.T. 0.012 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:105 Resp: 1295107
Ion Ratio Lower Upper
105 100
77 80.2 64.7 97.1
51 26.4 21.7 32.5



#17
N-Nitroso-di-n-propylamine
Concen: 75.468 ng/ul
RT: 8.698 min Scan# 954
Delta R.T. 0.029 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion: 70 Resp: 701783
Ion Ratio Lower Upper
70 100
42 57.0 47.6 71.4
101 9.4 7.4 11.0
130 20.8 15.1 22.7









#18
4-Methylphenol
Concen: 78.300 ng/ul
RT: 8.657 min Scan# 944
Delta R.T. 0.017 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:108 Resp: 891950
Ion Ratio Lower Upper
108 100
107 111.0 88.9 133.3



#19
Hexachloroethane
Concen: 81.271 ng/ul
RT: 8.934 min Scan# 994
Delta R.T. -0.000 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:117 Resp: 371326
Ion Ratio Lower Upper
117 100
201 105.0 88.8 133.2
199 69.1 55.0 82.4









#22
Nitrobenzene
Concen: 79.698 ng/ul
RT: 9.075 min Scan# 1107
Delta R.T. 0.012 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion: 77 Resp: 1054089
Ion Ratio Lower Upper
77 100
123 43.9 34.2 51.2
65 13.9 10.9 16.3



#23
Isophorone
Concen: 79.852 ng/ul
RT: 9.616 min Scan# 1110
Delta R.T. 0.023 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion: 82 Resp: 1993357
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 16.8 13.4 20.2









#25
2-Nitrophenol
Concen: 84.076 ng/ul
RT: 9.787 min Scan# 1137
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:139 Resp: 523377
Ion Ratio Lower Upper
139 100
65 52.1 44.5 66.7
109 32.9 26.0 39.0



#26
2,4-Dimethylphenol
Concen: 79.904 ng/ul
RT: 9.851 min Scan# 1150
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:107 Resp: 985013
Ion Ratio Lower Upper
107 100
121 49.5 39.4 59.2
122 84.5 65.8 98.6





#27

Bis(2-Chloroethoxy)methane

Concen: 77.436 ng/ul

RT: 10.087 min Scan# 1189

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion: 93 Resp: 1180963

Ion Ratio Lower Upper

93 100

95 33.0 25.4 38.2

123 11.2 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 81.146 ng/ul

RT: 10.322 min Scan# 1230

Delta R.T. 0.011 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:162 Resp: 876467


Ion Ratio Lower Upper

162 100

164 64.9 51.4 77.0

98 35.6 28.0 42.0









#30
Naphthalene
Concen: 77.884 ng/ul
RT: 10.722 min Scan# 1298
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:128 Resp: 2738909
Ion Ratio Lower Upper
128 100
129 10.5 8.5 12.7
127 13.2 10.3 15.5



#32
4-Chloroaniline
Concen: 77.025 ng/ul
RT: 10.840 min Scan# 1318
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:127 Resp: 1170589
Ion Ratio Lower Upper
127 100
129 31.6 25.9 38.9









#33
Hexachlorobutadiene
Concen: 81.258 ng/ul
RT: 10.998 min Scan# 11
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:225 Resp: 615025
Ion Ratio Lower Upper
225 100
223 64.6 50.2 75.4
227 64.9 52.2 78.2



#34
Caprolactam
Concen: 49.264 ng/ul
RT: 11.657 min Scan# 1457
Delta R.T. 0.041 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:113 Resp: 165066
Ion Ratio Lower Upper
113 100
55 154.6 125.6 188.4
56 117.4 94.1 141.1









#35
4-Chloro-3-methylphenol
Concen: 79.211 ng/ul
RT: 11.975 min Scan# 1511
Delta R.T. 0.012 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:107 Resp: 917075
Ion Ratio Lower Upper
107 100
144 25.8 21.7 32.5
142 81.1 63.7 95.5



#36
2-Methylnaphthalene
Concen: 77.914 ng/ul
RT: 12.334 min Scan# 1572
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:142 Resp: 1861733
Ion Ratio Lower Upper
142 100
141 87.8 71.3 106.9









#37
1-Methylnaphthalene
Concen: 76.664 ng/ul
RT: 12.551 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:142 Resp: 1874124
Ion Ratio Lower Upper
142 100
141 91.9 72.2 108.2
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 82.031 ng/ul
RT: 12.704 min Scan# 1635
Delta R.T. 0.012 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:216 Resp: 1132476
Ion Ratio Lower Upper
216 100
214 77.4 63.0 94.4
179 19.3 16.5 24.7
108 14.9 13.3 19.9









#40
Hexachlorocyclopentadiene
Concen: 84.474 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:237 Resp: 839611
Ion Ratio Lower Upper
237 100
235 64.0 53.0 79.4
272 13.3 10.1 15.1



#41
2,4,6-Trichlorophenol
Concen: 82.861 ng/ul
RT: 12.945 min Scan# 1676
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:196 Resp: 754760
Ion Ratio Lower Upper
196 100
198 88.6 76.8 115.2
200 30.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 82.575 ng/ul
RT: 13.022 min Scan# 1689
Delta R.T. 0.012 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:196 Resp: 814721
Ion Ratio Lower Upper
196 100
198 91.6 77.6 116.4
200 30.0 24.6 37.0



#43
1,1'-Biphenyl
Concen: 78.343 ng/ul
RT: 13.345 min Scan# 1744
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:154 Resp: 2480495
Ion Ratio Lower Upper
154 100
153 40.4 32.6 48.8
76 13.8 10.7 16.1









#44
2-Chloronaphthalene
Concen: 79.921 ng/ul
RT: 13.386 min Scan# 1751
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:162 Resp: 1968942
Ion Ratio Lower Upper
162 100
164 32.7 25.9 38.9
127 38.9 30.7 46.1



#45
2-Nitroaniline
Concen: 83.383 ng/ul
RT: 13.604 min Scan# 1788
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion: 65 Resp: 616810
Ion Ratio Lower Upper
65 100
92 68.1 53.2 79.8
138 105.7 78.3 117.5









#47
Dimethylphthalate
Concen: 78.477 ng/ul
RT: 13.981 min Scan# 1852
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:163 Resp: 2498327
Ion Ratio Lower Upper
163 100
194 4.4 3.5 5.3
164 10.2 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 87.683 ng/ul
RT: 14.104 min Scan# 1873
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:165 Resp: 553423
Ion Ratio Lower Upper
165 100
89 60.6 50.8 76.2
121 20.4 16.8 25.2





#50

Acenaphthylene

Concen: 78.697 ng/ul

RT: 14.233 min Scan# 1895

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:152 Resp: 2933028

Ion Ratio Lower Upper

152 100

151 19.8 16.3 24.5

153 13.3 10.3 15.5



#51

3-Nitroaniline

Concen: 81.259 ng/ul

RT: 14.433 min Scan# 1929

Delta R.T. 0.011 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:138 Resp: 461683


Ion Ratio Lower Upper

138 100

108 10.6 9.1 13.7

92 114.3 97.6 146.4









#52
Acenaphthene
Concen: 78.391 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:153 Resp: 2028114
Ion Ratio Lower Upper
153 100
152 46.0 38.9 58.3
154 84.6 72.7 109.1



#53
2,4-Dinitrophenol
Concen: 87.831 ng/ul
RT: 14.639 min Scan# 1964
Delta R.T. 0.011 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:184 Resp: 389708
Ion Ratio Lower Upper
184 100
63 62.9 53.2 79.8
154 59.9 50.6 76.0









#55
4-Nitrophenol
Concen: 79.439 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.023 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:109 Resp: 392749
Ion Ratio Lower Upper
109 100
139 124.0 98.1 147.1
65 123.9 101.6 152.4



#56
Dibenzofuran
Concen: 76.942 ng/ul
RT: 14.916 min Scan# 2011
Delta R.T. 0.012 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:168 Resp: 2839792
Ion Ratio Lower Upper
168 100
139 37.0 29.0 43.6









#57
2,4-Dinitrotoluene
Concen: 83.179 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. 0.017 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:165 Resp: 742779
Ion Ratio Lower Upper
165 100
63 47.0 40.7 61.1
182 3.0 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 79.903 ng/ul
RT: 15.139 min Scan# 2049
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:232 Resp: 703068
Ion Ratio Lower Upper
232 100
131 40.1 32.0 48.0
130 2.4 1.8 2.6
166 29.4 21.0 31.4





#59

Diethylphthalate

Concen: 77.721 ng/ul

RT: 15.345 min Scan# 2084

Delta R.T. 0.011 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:149 Resp: 2458550

Ion Ratio Lower Upper

149 100

177 22.8 17.8 26.6

150 12.5 9.4 14.2



#61

Fluorene

Concen: 73.672 ng/ul

RT: 15.563 min Scan# 2121

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:166 Resp: 2223696


Ion Ratio Lower Upper

166 100

165 98.5 74.1 111.1

167 13.3 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 74.148 ng/ul
RT: 15.551 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:204 Resp: 1174104
Ion Ratio Lower Upper
204 100
206 31.1 27.2 40.8
141 58.7 45.2 67.8



#63
4-Nitroaniline
Concen: 79.798 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.017 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:138 Resp: 390094
Ion Ratio Lower Upper
138 100
92 52.0 41.5 62.3
108 88.4 71.0 106.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 83.184 ng/ul
 RT: 15.657 min Scan# 2137
 Delta R.T. 0.017 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11




Instrument :
 BNA_N
 ClientSampleId :
 SST080228


Tgt Ion:198 Resp: 496713
 Ion Ratio Lower Upper
 198 100
 182 3.3 3.2 4.8
 77 22.9 21.1 31.7



#67
 N-Nitrosodiphenylamine
 Concen: 78.469 ng/ul
 RT: 15.775 min Scan# 2157
 Delta R.T. 0.006 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11

Tgt Ion:169 Resp: 1985791
 Ion Ratio Lower Upper
 169 100
 168 66.3 52.2 78.2
 167 35.6 27.3 40.9









#68
4-Bromophenyl-phenylether
Concen: 77.943 ng/ul
RT: 16.451 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:248 Resp: 800067
Ion Ratio Lower Upper
248 100
250 102.4 79.2 118.8
141 68.8 50.4 75.6



#69
Hexachlorobenzene
Concen: 77.298 ng/ul
RT: 16.563 min Scan# 2291
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:284 Resp: 916671
Ion Ratio Lower Upper
284 100
142 33.1 23.2 34.8
249 32.8 22.6 34.0





#70

Atrazine

Concen: 76.878 ng/ul

RT: 16.733 min Scan# 21

Delta R.T. 0.012 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:200 Resp: 781714

Ion Ratio Lower Upper

200 100

173 26.7 20.9 31.3

215 48.8 40.2 60.2



#71

Pentachlorophenol

Concen: 80.609 ng/ul

RT: 16.910 min Scan# 2350

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:266 Resp: 612049


Ion Ratio Lower Upper

266 100

264 62.4 49.9 74.9

268 62.9 50.3 75.5





#72

Phenanthrene

Concen: 76.382 ng/ul

RT: 17.304 min Scan# 2416

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:178 Resp: 3605472

Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 19.0 15.2 22.8



#74

Anthracene

Concen: 76.410 ng/ul

RT: 17.398 min Scan# 2433

Delta R.T. 0.012 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:178 Resp: 3594856


Ion Ratio Lower Upper

178 100

179 15.6 11.8 17.8

176 18.3 14.4 21.6









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 83.673 ng/uL
 RT: 13.310 min Scan# 1738
 Delta R.T. 0.006 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11

Instrument :
 BNA_N
 ClientSampleId :
 SST080228

Tgt Ion:216 Resp: 1128256




Ion	Ratio	Lower	Upper
216	100		
214	75.0	61.9	92.9
218	46.8	36.0	54.0




#76
 Pentachlorobenzene
 Concen: 79.960 ng/uL
 RT: 14.828 min Scan# 1996
 Delta R.T. 0.006 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11

Tgt Ion:250 Resp: 1075226

Ion	Ratio	Lower	Upper
250	100		
248	59.8	51.2	76.8
252	61.4	50.8	76.2









#77
 Carbazole
 Concen: 76.309 ng/ul
 RT: 17.669 min Scan# 2478
 Delta R.T. 0.006 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11




Instrument :
 BNA_N
 ClientSampleId :
 SST080228


Tgt Ion:167 Resp: 3145996
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.4 26.2
 139 13.2 10.2 15.4



#78
 Di-n-butylphthalate
 Concen: 47.327 ng/ul
 RT: 18.227 min Scan# 2574
 Delta R.T. 0.006 min
 Lab File: BN023892.D
 Acq: 01 Feb 2023 16:11

Tgt Ion:149 Resp: 4182045
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.4 11.2
 104 5.8 4.6 6.8





#80

Fluoranthene

Concen: 86.096 ng/ul

RT: 19.321 min Scan# 21

Delta R.T. 0.012 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:202 Resp: 3997475

Ion Ratio Lower Upper

202 100

101 10.4 8.2 12.2

100 7.8 6.2 9.4



#82

Pyrene

Concen: 83.640 ng/ul

RT: 19.686 min Scan# 2822

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:202 Resp: 3950008


Ion Ratio Lower Upper

202 100

101 12.2 9.1 13.7

100 9.9 7.4 11.2









#83
Butylbenzylphthalate
Concen: 85.758 ng/ul
RT: 20.586 min Scan# 2975
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:149 Resp: 1689105
Ion Ratio Lower Upper
149 100
91 73.3 57.9 86.9
206 23.3 19.0 28.4



#84
3,3'-Dichlorobenzidine
Concen: 74.471 ng/ul
RT: 21.374 min Scan# 3109
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:252 Resp: 1145787
Ion Ratio Lower Upper
252 100
254 66.9 53.0 79.6
126 10.6 8.1 12.1









#85
Benzo(a)anthracene
Concen: 77.361 ng/ul
RT: 21.439 min Scan# 3119
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:228 Resp: 3511914
Ion Ratio Lower Upper
228 100
229 19.0 15.6 23.4
226 27.8 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 81.442 ng/ul
RT: 21.363 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:149 Resp: 2366468
Ion Ratio Lower Upper
149 100
167 29.5 21.2 31.8
279 5.2 3.8 5.6









#87
Chrysene
Concen: 76.167 ng/ul
RT: 21.492 min Scan# 3128
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11




Instrument :
BNA_N
ClientSampleId :
SSTD080228


Tgt Ion:228 Resp: 3190524
Ion Ratio Lower Upper
228 100
226 31.6 25.9 38.9
229 19.9 16.4 24.6



#89
Di-n-octyl phthalate
Concen: 86.592 ng/ul
RT: 22.286 min Scan# 3264
Delta R.T. 0.006 min
Lab File: BN023892.D
Acq: 01 Feb 2023 16:11

Tgt Ion:149 Resp: 3887123





#90

Benzo(b)fluoranthene

Concen: 79.548 ng/ul

RT: 23.127 min Scan# 3405

Delta R.T. 0.012 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:252 Resp: 3273573

Ion Ratio Lower Upper

252 100

253 20.9 17.8 26.8

125 8.6 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 82.121 ng/ul

RT: 23.174 min Scan# 3415

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:252 Resp: 3191317


Ion Ratio Lower Upper

252 100

253 20.4 16.2 24.2

125 8.7 6.6 9.8





#93

Benzo(a)pyrene

Concen: 81.346 ng/ul

RT: 23.757 min Scan# 3

Delta R.T. 0.017 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:252 Resp: 2816404

Ion Ratio Lower Upper

252 100

253 21.9 17.1 25.7

125 9.4 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 47.412 ng/ul

RT: 26.339 min Scan# 3953

Delta R.T. 0.006 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11




Tgt Ion:276 Resp: 1980372


Ion Ratio Lower Upper

276 100

138 19.5 16.6 25.0

277 24.8 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 84.592 ng/ul

RT: 26.368 min Scan# 3958

Delta R.T. 0.023 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Instrument :

BNA_N

ClientSampleId :

SSTD080228





Tgt Ion:278 Resp: 2922904

Ion Ratio Lower Upper

278 100

139 14.5 10.8 16.2

279 24.3 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 88.577 ng/ul

RT: 27.127 min Scan# 4087

Delta R.T. 0.023 min

Lab File: BN023892.D

Acq: 01 Feb 2023 16:11

Tgt Ion:276 Resp: 2913900

Ion Ratio Lower Upper

276 100

138 17.1 14.6 21.8

277 23.1 18.8 28.2

