

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023894.D
 Acq On : 01 Feb 2023 17:26
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV230

Quant Time: Feb 02 02:16:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 02:10:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.852	152	145340	20.000	ng/u1	0.00
20) Naphthalene-d8	10.663	136	623053	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.510	164	412771	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.257	188	968106	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	874242	20.000	ng/u1	0.00
88) Perylene-d12	23.857	264	862864	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	28759	7.986	ng/uL	0.00
4) Pyridine-d5	3.664	84	211364	21.007	ng/u1	0.00
7) Phenol-d5	7.028	99	253072	19.939	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	159004	19.823	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	195681	19.718	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	203235	20.186	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	101339	20.047	ng/u1	0.00
24) 2-Nitrophenol-d4	9.746	143	117229	20.381	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.287	165	216533	20.162	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	298002	20.339	ng/u1	0.00
46) Dimethylphthalate-d6	13.922	166	663254	20.126	ng/u1	0.00
49) Acenaphthylene-d8	14.198	160	749344	21.020	ng/u1	0.00
54) 4-Nitrophenol-d4	14.722	143	124166	20.782	ng/u1	0.00
60) Fluorene-d10	15.498	176	576529	20.868	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	130575	19.061	ng/u1	0.00
73) Anthracene-d10	17.351	188	912147	20.238	ng/u1	0.00
81) Pyrene-d10	19.651	212	1087649	20.492	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.686	264	903738	20.180	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.276	88	31546	8.504	ng/uL#	88
5) Pyridine	3.687	79	205104	20.287	ng/u1	97
6) Benzaldehyde	6.993	77	132068	23.930	ng/u1	97
8) Phenol	7.052	94	264211	20.708	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.281	93	210742	20.273	ng/u1	97
12) 2-Chlorophenol	7.417	128	210807	20.992	ng/u1	97
13) 2-Methylphenol	8.311	108	201970	20.863	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.393	45	276952	20.864	ng/u1	100
16) Acetophenone	8.687	105	333961	21.418	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.675	70	182526	21.323	ng/u1	98
18) 4-Methylphenol	8.640	108	219732	20.954	ng/u1	99
19) Hexachloroethane	8.934	117	88939	21.146	ng/u1	99
22) Nitrobenzene	9.069	77	267119	21.049	ng/u1	100
23) Isophorone	9.593	82	508886	21.246	ng/u1	98
25) 2-Nitrophenol	9.781	139	128688	21.545	ng/u1	99
26) 2,4-Dimethylphenol	9.846	107	253200	21.406	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.081	93	296260	20.246	ng/u1	99
29) 2,4-Dichlorophenol	10.316	162	218974	21.129	ng/u1	96
30) Naphthalene	10.716	128	697914	20.684	ng/u1	99
32) 4-Chloroaniline	10.834	127	303381	20.805	ng/u1	99
33) Hexachlorobutadiene	10.993	225	154041	21.211	ng/u1	96
34) Caprolactam	11.616	113	46486	14.354	ng/u1	95
35) 4-Chloro-3-methylphenol	11.963	107	239997	21.604	ng/u1	98
36) 2-Methylnaphthalene	12.328	142	504455	22.003	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023894.D
 Acq On : 01 Feb 2023 17:26
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV230

Quant Time: Feb 02 02:16:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 02:10:45 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.546	142	497017	21.189	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.699	216	304632	21.479	ng/ul	96
40) Hexachlorocyclopentadiene	12.669	237	186307	18.246	ng/ul	95
41) 2,4,6-Trichlorophenol	12.940	196	197728	21.130	ng/ul	93
42) 2,4,5-Trichlorophenol	13.010	196	216558	21.365	ng/ul	99
43) 1,1'-Biphenyl	13.340	154	671215	20.635	ng/ul	100
44) 2-Chloronaphthalene	13.381	162	539797	21.328	ng/ul	100
45) 2-Nitroaniline	13.599	65	162964	21.444	ng/ul	95
47) Dimethylphthalate	13.969	163	698482	21.357	ng/ul	100
48) 2,6-Dinitrotoluene	14.093	165	142419	21.964	ng/ul	99
50) Acenaphthylene	14.228	152	808222	21.109	ng/ul	100
51) 3-Nitroaniline	14.422	138	134446	23.034	ng/ul	96
52) Acenaphthene	14.575	153	564029	21.221	ng/ul	96
53) 2,4-Dinitrophenol	14.628	184	96390	21.146	ng/ul	98
55) 4-Nitrophenol	14.734	109	113336	22.314	ng/ul	97
56) Dibenzofuran	14.904	168	809296	21.344	ng/ul	97
57) 2,4-Dinitrotoluene	14.881	165	203767	22.211	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.134	232	197846	21.887	ng/ul	97
59) Diethylphthalate	15.334	149	699533	21.526	ng/ul	97
61) Fluorene	15.557	166	655543	21.141	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.551	204	353768	21.747	ng/ul	97
63) 4-Nitroaniline	15.587	138	125000	23.596	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.640	198	139672	21.176	ng/ul	99
67) N-Nitrosodiphenylamine	15.769	169	591569	21.163	ng/ul	98
68) 4-Bromophenyl-phenylether	16.445	248	229738	20.262	ng/ul	98
69) Hexachlorobenzene	16.557	284	269649	20.585	ng/ul	91
70) Atrazine	16.722	200	218061	19.415	ng/ul	97
71) Pentachlorophenol	16.904	266	171766	20.480	ng/ul	100
72) Phenanthrene	17.298	178	1061896	20.366	ng/ul	99
74) Anthracene	17.392	178	1088371	20.943	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.304	216	299865	20.133	ng/uL#	89
76) Pentachlorobenzene	14.822	250	325355	21.904	ng/uL	98
77) Carbazole	17.663	167	975792	21.428	ng/ul	100
78) Di-n-butylphthalate	18.228	149	1296801	13.286	ng/ul	99
80) Fluoranthene	19.316	202	1334140	21.531	ng/ul	98
82) Pyrene	19.681	202	1351875	21.450	ng/ul	99
83) Butylbenzylphthalate	20.581	149	561546	21.363	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.369	252	466435	22.716	ng/ul	95
85) Benzo(a)anthracene	21.433	228	1320768	21.801	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.357	149	833212	21.486	ng/ul	97
87) Chrysene	21.486	228	1233152	22.059	ng/ul	95
89) Di-n-octyl phthalate	22.286	149	1359714	20.950	ng/ul	100
90) Benzo(b)fluoranthene	23.122	252	1213769	20.396	ng/ul	98
91) Benzo(k)fluoranthene	23.169	252	1183121	21.053	ng/ul	99
93) Benzo(a)pyrene	23.745	252	1017942	20.332	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.333	276	1093637	18.102	ng/ul	97
95) Dibenzo(a,h)anthracene	26.345	278	1025679	20.528	ng/ul	100
96) Benzo(g,h,i)perylene	27.104	276	954059	20.055	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
 Data File : BN023894.D
 Acq On : 01 Feb 2023 17:26
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV230

Quant Time: Feb 02 02:16:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 02:10:45 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.504 ng/uL
RT: 3.276 min Scan# 31
Delta R.T. -0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion: 88 Resp: 31546
Ion Ratio Lower Upper
88 100
43 34.0 42.1 63.1#
58 67.4 53.8 80.6



#5
Pyridine
Concen: 20.287 ng/uL
RT: 3.687 min Scan# 102
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion: 79 Resp: 205104
Ion Ratio Lower Upper
79 100
52 67.7 56.4 84.6
51 34.4 26.6 39.8









#6
Benzaldehyde
Concen: 23.930 ng/ul
RT: 6.993 min Scan# 664
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion: 77 Resp: 132068
Ion Ratio Lower Upper
77 100
105 94.1 73.6 110.4
106 91.7 70.2 105.4



#8
Phenol
Concen: 20.708 ng/ul
RT: 7.052 min Scan# 674
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion: 94 Resp: 264211
Ion Ratio Lower Upper
94 100
65 28.6 24.1 36.1
66 40.7 32.4 48.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.273 ng/ul
 RT: 7.281 min Scan# 713
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Instrument :
 BNA_N
 ClientSampleId :
 SICV230

Tgt Ion: 93 Resp: 210742




Ion	Ratio	Lower	Upper
93	100		
63	72.6	60.6	91.0
95	31.6	26.2	39.4




#12
 2-Chlorophenol
 Concen: 20.992 ng/ul
 RT: 7.417 min Scan# 736
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Tgt Ion:128 Resp: 210807

Ion	Ratio	Lower	Upper
128	100		
64	49.0	38.4	57.6
130	34.8	25.6	38.4









#13
2-Methylphenol
Concen: 20.863 ng/ul
RT: 8.311 min Scan# 887
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:108 Resp: 201970
Ion Ratio Lower Upper
108 100
107 88.1 72.6 109.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.864 ng/ul
RT: 8.393 min Scan# 902
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion: 45 Resp: 276952
Ion Ratio Lower Upper
45 100
77 16.1 12.7 19.1
79 12.5 9.8 14.8









#16
Acetophenone
Concen: 21.418 ng/ul
RT: 8.687 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:105 Resp: 333961
Ion Ratio Lower Upper
105 100
77 78.4 64.7 97.1
51 25.3 21.7 32.5



#17
N-Nitroso-di-n-propylamine
Concen: 21.323 ng/ul
RT: 8.675 min Scan# 950
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion: 70 Resp: 182526
Ion Ratio Lower Upper
70 100
42 57.9 47.6 71.4
101 9.8 7.4 11.0
130 20.6 15.1 22.7









#18
4-Methylphenol
Concen: 20.954 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:108 Resp: 219732
Ion Ratio Lower Upper
108 100
107 111.8 88.9 133.3



#19
Hexachloroethane
Concen: 21.146 ng/ul
RT: 8.934 min Scan# 994
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:117 Resp: 88939
Ion Ratio Lower Upper
117 100
201 109.6 88.8 133.2
199 67.7 55.0 82.4









#22
Nitrobenzene
Concen: 21.049 ng/ul
RT: 9.069 min Scan# 1016
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion: 77 Resp: 267119
Ion Ratio Lower Upper
77 100
123 42.5 34.2 51.2
65 13.8 10.9 16.3



#23
Isophorone
Concen: 21.246 ng/ul
RT: 9.593 min Scan# 1106
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion: 82 Resp: 508886
Ion Ratio Lower Upper
82 100
95 7.3 5.6 8.4
138 17.6 13.4 20.2









#25
2-Nitrophenol
Concen: 21.545 ng/ul
RT: 9.781 min Scan# 1138
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:139 Resp: 128688
Ion Ratio Lower Upper
139 100
65 55.5 44.5 66.7
109 34.5 26.0 39.0



#26
2,4-Dimethylphenol
Concen: 21.406 ng/ul
RT: 9.846 min Scan# 1149
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:107 Resp: 253200
Ion Ratio Lower Upper
107 100
121 49.9 39.4 59.2
122 83.0 65.8 98.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.246 ng/ul
 RT: 10.081 min Scan# 1189
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26




Instrument :
 BNA_N
 ClientSampleId :
 SICV230


Tgt Ion: 93 Resp: 296260
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.4 38.2
 123 10.4 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 21.129 ng/ul
 RT: 10.316 min Scan# 1229
 Delta R.T. 0.006 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Tgt Ion: 162 Resp: 218974
 Ion Ratio Lower Upper
 162 100
 164 69.3 51.4 77.0
 98 34.5 28.0 42.0





#30

Naphthalene

Concen: 20.684 ng/ul

RT: 10.716 min Scan# 1296

Delta R.T. 0.006 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:128 Resp: 697914

Ion Ratio Lower Upper

128 100

129 10.7 8.5 12.7

127 13.5 10.3 15.5



#32

4-Chloroaniline

Concen: 20.805 ng/ul

RT: 10.834 min Scan# 1317

Delta R.T. 0.000 min

Lab File: BN023894.D




Acq: 01 Feb 2023 17:26


Tgt Ion:127 Resp: 303381

Ion Ratio Lower Upper

127 100

129 33.1 25.9 38.9









#33
Hexachlorobutadiene
Concen: 21.211 ng/ul
RT: 10.993 min Scan# 1344
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:225 Resp: 154041
Ion Ratio Lower Upper
225 100
223 66.6 50.2 75.4
227 68.0 52.2 78.2



#34
Caprolactam
Concen: 14.354 ng/ul
RT: 11.616 min Scan# 1450
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:113 Resp: 46486
Ion Ratio Lower Upper
113 100
55 154.7 125.6 188.4
56 126.5 94.1 141.1









#35
4-Chloro-3-methylphenol
Concen: 21.604 ng/ul
RT: 11.963 min Scan# 1509
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:107 Resp: 239997
Ion Ratio Lower Upper
107 100
144 26.8 21.7 32.5
142 81.7 63.7 95.5



#36
2-Methylnaphthalene
Concen: 22.003 ng/ul
RT: 12.328 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:142 Resp: 504455
Ion Ratio Lower Upper
142 100
141 90.0 71.3 106.9





#37

1-Methylnaphthalene

Concen: 21.189 ng/ul

RT: 12.546 min Scan# 1608

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:142 Resp: 497017

Ion Ratio Lower Upper

142 100

141 92.7 72.2 108.2

116 4.2 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.479 ng/ul

RT: 12.699 min Scan# 1634

Delta R.T. 0.006 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Tgt Ion:216 Resp: 304632




Ion Ratio Lower Upper


216 100

214 75.5 63.0 94.4

179 18.5 16.5 24.7

108 15.4 13.3 19.9





#40

Hexachlorocyclopentadiene

Concen: 18.246 ng/ul

RT: 12.669 min Scan# 1629

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:237 Resp: 186307

Ion Ratio Lower Upper

237 100

235 62.0 53.0 79.4

272 11.8 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 21.130 ng/ul

RT: 12.940 min Scan# 1675

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:196 Resp: 197728


Ion Ratio Lower Upper

196 100

198 103.9 76.8 115.2

200 32.7 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 21.365 ng/ul

RT: 13.010 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:196 Resp: 216558

Ion Ratio Lower Upper

196 100

198 96.3 77.6 116.4

200 30.7 24.6 37.0



#43

1,1'-Biphenyl

Concen: 20.635 ng/ul

RT: 13.340 min Scan# 1743

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:154 Resp: 671215


Ion Ratio Lower Upper

154 100

153 40.9 32.6 48.8

76 13.4 10.7 16.1









#44
2-Chloronaphthalene
Concen: 21.328 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:162 Resp: 539797
Ion Ratio Lower Upper
162 100
164 32.0 25.9 38.9
127 38.6 30.7 46.1



#45
2-Nitroaniline
Concen: 21.444 ng/ul
RT: 13.599 min Scan# 1787
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion: 65 Resp: 162964
Ion Ratio Lower Upper
65 100
92 67.3 53.2 79.8
138 105.2 78.3 117.5









#47
Dimethylphthalate
Concen: 21.357 ng/ul
RT: 13.969 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:163 Resp: 698482
Ion Ratio Lower Upper
163 100
194 4.3 3.5 5.3
164 10.0 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 21.964 ng/ul
RT: 14.093 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:165 Resp: 142419
Ion Ratio Lower Upper
165 100
89 62.3 50.8 76.2
121 20.8 16.8 25.2





#50

Acenaphthylene

Concen: 21.109 ng/ul

RT: 14.228 min Scan# 1894

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:152 Resp: 808222

Ion Ratio Lower Upper

152 100

151 20.5 16.3 24.5

153 12.8 10.3 15.5



#51

3-Nitroaniline

Concen: 23.034 ng/ul

RT: 14.422 min Scan# 1927

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:138 Resp: 134446


Ion Ratio Lower Upper

138 100

108 9.7 9.1 13.7

92 117.0 97.6 146.4









#52
 Acenaphthene
 Concen: 21.221 ng/ul
 RT: 14.575 min Scan# 1953
 Delta R.T. 0.006 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26




Instrument :
 BNA_N
 ClientSampleId :
 SICV230


Tgt Ion:153 Resp: 564029
 Ion Ratio Lower Upper
 153 100
 152 45.6 38.9 58.3
 154 87.4 72.7 109.1



#53
 2,4-Dinitrophenol
 Concen: 21.146 ng/ul
 RT: 14.628 min Scan# 1962
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Tgt Ion:184 Resp: 96390
 Ion Ratio Lower Upper
 184 100
 63 65.5 53.2 79.8
 154 66.0 50.6 76.0





#55

4-Nitrophenol

Concen: 22.314 ng/ul

RT: 14.734 min Scan# 1980

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:109 Resp: 113336

Ion Ratio Lower Upper

109 100

139 118.6 98.1 147.1

65 128.7 101.6 152.4



#56

Dibenzofuran

Concen: 21.344 ng/ul

RT: 14.904 min Scan# 2009

Delta R.T. 0.000 min

Lab File: BN023894.D




Acq: 01 Feb 2023 17:26


Tgt Ion:168 Resp: 809296

Ion Ratio Lower Upper

168 100

139 38.3 29.0 43.6





#57

2,4-Dinitrotoluene

Concen: 22.211 ng/ul

RT: 14.881 min Scan# 2005

Delta R.T. 0.006 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:165 Resp: 203767

Ion Ratio Lower Upper

165 100

63 45.6 40.7 61.1

182 3.3 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.887 ng/ul

RT: 15.134 min Scan# 2048

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Tgt Ion:232 Resp: 197846




Ion Ratio Lower Upper


232 100

131 42.2 32.0 48.0

130 2.4 1.8 2.6

166 27.9 21.0 31.4





#59

Diethylphthalate

Concen: 21.526 ng/ul

RT: 15.334 min Scan# 2082

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:149 Resp: 699533

Ion Ratio Lower Upper

149 100

177 23.2 17.8 26.6

150 13.1 9.4 14.2



#61

Fluorene

Concen: 21.141 ng/ul

RT: 15.557 min Scan# 2120

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:166 Resp: 655543


Ion Ratio Lower Upper

166 100

165 103.0 74.1 111.1

167 14.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.747 ng/ul
RT: 15.551 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:204 Resp: 353768
Ion Ratio Lower Upper
204 100
206 36.1 27.2 40.8
141 58.7 45.2 67.8



#63
4-Nitroaniline
Concen: 23.596 ng/ul
RT: 15.587 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:138 Resp: 125000
Ion Ratio Lower Upper
138 100
92 53.3 41.5 62.3
108 91.4 71.0 106.6









#66
4,6-Dinitro-2-methylphenol
Concen: 21.176 ng/ul
RT: 15.640 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:198 Resp: 139672
Ion Ratio Lower Upper
198 100
182 3.9 3.2 4.8
77 26.0 21.1 31.7



#67
N-Nitrosodiphenylamine
Concen: 21.163 ng/ul
RT: 15.769 min Scan# 2156
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:169 Resp: 591569
Ion Ratio Lower Upper
169 100
168 64.6 52.2 78.2
167 35.8 27.3 40.9









#68
4-Bromophenyl-phenylether
Concen: 20.262 ng/ul
RT: 16.445 min Scan# 2271
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:248 Resp: 229738
Ion Ratio Lower Upper
248 100
250 100.8 79.2 118.8
141 65.0 50.4 75.6



#69
Hexachlorobenzene
Concen: 20.585 ng/ul
RT: 16.557 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:284 Resp: 269649
Ion Ratio Lower Upper
284 100
142 34.4 23.2 34.8
249 32.5 22.6 34.0





#70

Atrazine

Concen: 19.415 ng/ul

RT: 16.722 min Scan# 218061

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:200 Resp: 218061

Ion Ratio Lower Upper

200 100

173 25.2 20.9 31.3

215 47.2 40.2 60.2



#71

Pentachlorophenol

Concen: 20.480 ng/ul

RT: 16.904 min Scan# 2349

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:266 Resp: 171766


Ion Ratio Lower Upper

266 100

264 62.0 49.9 74.9

268 63.2 50.3 75.5









#72
Phenanthrene
Concen: 20.366 ng/ul
RT: 17.298 min Scan# 2416
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:178 Resp: 1061896
Ion Ratio Lower Upper
178 100
179 15.5 12.1 18.1
176 19.7 15.2 22.8



#74
Anthracene
Concen: 20.943 ng/ul
RT: 17.392 min Scan# 2432
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:178 Resp: 1088371
Ion Ratio Lower Upper
178 100
179 15.0 11.8 17.8
176 17.7 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.133 ng/uL

RT: 13.304 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:216 Resp: 299865

Ion Ratio Lower Upper

216 100

214 85.3 61.9 92.9

218 54.2 36.0 54.0#



#76

Pentachlorobenzene

Concen: 21.904 ng/uL

RT: 14.822 min Scan# 1995

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:250 Resp: 325355


Ion Ratio Lower Upper

250 100

248 65.4 51.2 76.8

252 65.4 50.8 76.2









#77
 Carbazole
 Concen: 21.428 ng/ul
 RT: 17.663 min Scan# 2478
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Instrument :
 BNA_N
 ClientSampleId :
 SICV230

Tgt Ion:167 Resp: 975792




Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.4	26.2
139	12.8	10.2	15.4




#78
 Di-n-butylphthalate
 Concen: 13.286 ng/ul
 RT: 18.228 min Scan# 2574
 Delta R.T. 0.006 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Tgt Ion:149 Resp: 1296801

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.4	11.2
104	5.3	4.6	6.8









#80
Fluoranthene
Concen: 21.531 ng/u1
RT: 19.316 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:202 Resp: 1334140
Ion Ratio Lower Upper
202 100
101 9.2 8.2 12.2
100 7.3 6.2 9.4



#82
Pyrene
Concen: 21.450 ng/u1
RT: 19.681 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:202 Resp: 1351875
Ion Ratio Lower Upper
202 100
101 11.7 9.1 13.7
100 9.9 7.4 11.2





#83

Butylbenzylphthalate

Concen: 21.363 ng/ul

RT: 20.581 min Scan# 2974

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:149 Resp: 561546

Ion Ratio Lower Upper

149 100

91 74.6 57.9 86.9

206 23.3 19.0 28.4



#84

3,3'-Dichlorobenzidine

Concen: 22.716 ng/ul

RT: 21.369 min Scan# 3108

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:252 Resp: 466435


Ion Ratio Lower Upper

252 100

254 61.7 53.0 79.6

126 10.0 8.1 12.1









#85
Benzo(a)anthracene
Concen: 21.801 ng/ul
RT: 21.433 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26




Instrument :
BNA_N
ClientSampleId :
SICV230


Tgt Ion:228 Resp: 1320768
Ion Ratio Lower Upper
228 100
229 20.2 15.6 23.4
226 26.6 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.486 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BN023894.D
Acq: 01 Feb 2023 17:26

Tgt Ion:149 Resp: 833212
Ion Ratio Lower Upper
149 100
167 28.1 21.2 31.8
279 4.2 3.8 5.6





#87

Chrysene

Concen: 22.059 ng/u1

RT: 21.486 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:228 Resp: 1233152

Ion Ratio Lower Upper

228 100

226 29.2 25.9 38.9

229 19.1 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 20.950 ng/u1




RT: 22.286 min Scan# 3264


Delta R.T. 0.006 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Tgt Ion:149 Resp: 1359714





#90

Benzo(b)fluoranthene

Concen: 20.396 ng/ul

RT: 23.122 min Scan# 3405

Delta R.T. 0.006 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:252 Resp: 1213769

Ion Ratio Lower Upper

252 100

253 23.8 17.8 26.8

125 9.5 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 21.053 ng/ul

RT: 23.169 min Scan# 3414

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:252 Resp: 1183121


Ion Ratio Lower Upper

252 100

253 20.9 16.2 24.2

125 8.1 6.6 9.8





#93

Benzo(a)pyrene

Concen: 20.332 ng/ul

RT: 23.745 min Scan# 3511

Delta R.T. 0.006 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26

Instrument :

BNA_N

ClientSampleId :

SICV230





Tgt Ion:252 Resp: 1017942

Ion Ratio Lower Upper

252 100

253 22.4 17.1 25.7

125 9.5 7.5 11.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.102 ng/ul

RT: 26.333 min Scan# 3952

Delta R.T. 0.000 min

Lab File: BN023894.D

Acq: 01 Feb 2023 17:26




Tgt Ion:276 Resp: 1093637


Ion Ratio Lower Upper

276 100

138 21.4 16.6 25.0

277 26.6 19.7 29.5









#95
 Dibenzo(a,h)anthracene
 Concen: 20.528 ng/ul
 RT: 26.345 min Scan# 3954
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Instrument :
 BNA_N
 ClientSampleId :
 SICV230

Tgt Ion:278 Resp: 1025679

Ion	Ratio	Lower	Upper
278	100		
139	13.5	10.8	16.2
279	23.2	18.3	27.5



#96
 Benzo(g,h,i)perylene
 Concen: 20.055 ng/ul
 RT: 27.104 min Scan# 4083
 Delta R.T. 0.000 min
 Lab File: BN023894.D
 Acq: 01 Feb 2023 17:26

Tgt Ion:276 Resp: 954059

Ion	Ratio	Lower	Upper
276	100		
138	17.9	14.6	21.8
277	23.1	18.8	28.2

