



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020123\
Data File : BN023887.D
Acq On : 01 Feb 2023 12:57
Operator : CG/JU
Sample : DFTPP273
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP273

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
Title : SVOA CALIBRATION
Last Update : Thu Feb 02 02:10:45 2023



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.3	132374	PASS
68	69	0.00	2	2.0	2794	PASS
69	198	0.00	100	38.1	142723	PASS
70	69	0.00	2	0.6	864	PASS
127	198	10	80	44.0	164736	PASS
197	198	0.00	2	0.2	801	PASS
198	198	100	100	100.0	374528	PASS
199	198	5	9	7.0	26291	PASS
275	198	10	60	27.7	103563	PASS
365	198	1	100	4.0	15088	PASS
441	198	0.01	100	13.1	48973	PASS
442	198	50	100	84.6	316715	PASS
443	442	15	24	19.5	61667	PASS

m/z	Abundance
37.00	179.0
38.00	173.0
39.10	1088.0
39.90	1039.0
41.00	2734.0
42.00	2867.0
43.00	16152.0
44.00	2444.0
45.00	12862.0
49.90	305.0
50.80	298.0
51.90	333.0
53.00	357.0
55.00	938.0
56.00	1182.0
57.00	658.0
59.10	171.0
60.00	8215.0
60.90	366.0
62.90	158.0
65.00	168.0
65.90	185.0
67.00	171.0
69.00	213.0
71.10	238.0
72.90	1414.0
74.00	216.0
75.00	343.0
77.00	794.0
78.00	668.0
78.90	302.0
81.10	228.0
82.10	223.0
84.00	276.0
85.00	161.0
91.00	362.0
94.90	226.0
96.00	267.0
97.10	183.0
99.00	169.0
103.10	174.0
104.90	401.0
114.90	211.0
117.00	287.0
119.00	197.0
123.10	424.0
128.10	298.0
129.10	307.0
132.90	316.0
135.00	263.0
142.90	171.0
147.10	443.0
148.90	386.0
168.00	237.0
190.90	273.0
193.00	351.0
207.00	1925.0
208.00	446.0
208.90	365.0
220.90	328.0
266.90	243.0
281.00	922.0
282.00	221.0
354.90	195.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

m/z	Abundance
37.00	871.0
38.00	1939.0
39.00	10422.0
40.00	1107.0
41.00	2450.0
42.00	2781.0
43.00	14348.0
44.00	2125.0
45.00	11501.0
46.90	196.0
47.80	183.0
49.10	1136.0
50.00	35296.0
51.00	126728.0
52.00	6521.0
53.00	444.0
54.10	162.0
55.00	1461.0
56.00	4236.0
57.00	8314.0
58.00	574.0
58.90	171.0
60.00	7897.0
61.00	2092.0
62.00	2032.0
63.00	5366.0
64.00	838.0
65.00	2715.0
66.00	411.0
67.00	515.0
67.90	2717.0
69.00	135552.0
70.00	783.0
71.00	345.0
72.00	161.0
73.10	2380.0
74.00	15071.0
75.00	22568.0
76.00	8607.0
77.00	159744.0
78.00	11786.0
79.00	11448.0
80.00	7727.0
81.00	11077.0
82.00	2711.0
83.00	2454.0
84.00	623.0
85.00	2211.0
86.00	3610.0
86.90	1504.0
88.00	739.0
89.00	359.0
91.00	2508.0
92.00	3101.0
93.00	19464.0
94.00	1303.0
95.10	306.0
96.00	1302.0
96.90	581.0
98.00	15596.0
99.00	12162.0
100.00	1016.0
101.00	6003.0
101.90	309.0
103.00	1736.0
103.90	3490.0
105.00	3608.0
106.00	1293.0
107.00	44968.0
108.00	7011.0
109.00	1627.0
110.00	78904.0
111.00	12791.0
112.00	1594.0
112.90	511.0
113.80	189.0
115.10	420.0
116.00	2441.0
117.00	47232.0
118.00	3617.0
119.00	675.0
120.00	600.0
120.80	214.0
122.00	2804.0
122.90	4596.0
123.90	2238.0
124.90	2061.0
127.00	144320.0
128.00	12410.0
129.00	57824.0
130.00	4978.0
131.00	1439.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

132.00	519.0
132.90	601.0
134.00	1650.0
135.00	4754.0
136.00	2006.0
137.00	2973.0
137.90	535.0
139.00	411.0
140.00	699.0
141.00	8040.0
142.00	2927.0
143.00	1860.0
143.90	638.0
145.00	552.0
146.00	1742.0
147.00	4264.0
148.00	10717.0
149.00	2653.0
149.90	802.0
151.00	1154.0
152.00	653.0
152.90	2512.0
154.00	1874.0
155.00	4652.0
156.00	6668.0
157.00	1298.0
158.00	1606.0
158.90	1319.0
159.90	2515.0
161.00	3873.0
162.00	1103.0
163.00	409.0
164.10	435.0
165.00	3040.0
166.00	2550.0
167.00	18504.0
168.00	11212.0
169.00	1330.0
169.90	606.0
170.90	789.0
172.00	1798.0
172.90	2254.0
174.00	3287.0
175.00	6513.0
175.90	2026.0
177.00	2915.0
177.90	1090.0
179.00	12377.0
180.00	8618.0
181.00	3796.0
182.00	711.0
183.00	435.0
183.90	994.0
185.00	5682.0
186.00	45864.0
187.00	13520.0
188.00	1561.0
189.00	3215.0
189.90	473.0
191.00	1242.0
192.00	4116.0
193.00	5124.0
194.10	970.0
195.00	558.0
196.00	9501.0
197.00	2404.0
198.00	315136.0
199.00	22576.0
200.10	1805.0
201.40	1508.0
202.90	2521.0
204.00	12285.0
205.00	20264.0
206.00	82008.0
207.00	13072.0
208.00	3804.0
209.00	1231.0
210.10	1473.0
211.00	3529.0
211.90	666.0
212.90	271.0
214.00	194.0
215.00	1080.0
216.00	1584.0
217.00	23872.0
218.00	2637.0
219.30	305.0
221.00	13387.0
221.80	3386.0
223.00	5317.0
224.00	48264.0
225.00	13183.0
226.00	1844.0
227.00	18864.0
228.00	3028.0
229.00	4024.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

229.90	623.0
231.10	1609.0
232.00	236.0
233.00	323.0
234.00	1580.0
234.90	1610.0
236.00	1106.0
237.00	2411.0
237.90	391.0
239.00	814.0
239.90	673.0
241.00	1383.0
242.00	2639.0
243.00	2982.0
244.00	38672.0
245.00	4802.0
246.00	7724.0
247.00	1612.0
248.10	285.0
248.90	1344.0
250.10	270.0
250.90	271.0
251.90	418.0
253.00	866.0
254.10	1653.0
255.00	188224.0
256.00	27872.0
257.10	2327.0
258.00	11238.0
259.00	1769.0
260.00	338.0
261.00	332.0
263.90	372.0
265.00	4468.0
265.90	597.0
267.00	207.0
270.00	242.0
270.90	695.0
272.00	655.0
273.00	6428.0
274.00	13516.0
275.00	81048.0
276.00	12247.0
277.00	7554.0
277.90	1563.0
279.00	310.0
281.10	664.0
282.20	247.0
282.90	958.0
284.10	674.0
285.00	1427.0
286.00	322.0
288.90	243.0
290.00	235.0
290.80	166.0
291.80	312.0
293.00	1735.0
294.00	473.0
295.00	659.0
296.00	22432.0
297.00	3515.0
298.00	213.0
301.00	296.0
301.90	433.0
302.90	2301.0
304.00	814.0
305.00	162.0
308.00	293.0
308.80	199.0
310.00	291.0
312.90	190.0
314.00	1157.0
314.90	2496.0
316.00	1515.0
317.00	255.0
321.00	783.0
321.80	318.0
322.20	308.0
323.00	7095.0
323.90	1384.0
326.90	1543.0
327.90	703.0
329.10	159.0
332.00	662.0
333.00	573.0
334.00	4817.0
335.00	1261.0
336.00	192.0
340.00	185.0
341.00	997.0
342.00	294.0
346.00	1837.0
346.80	315.0
352.00	2205.0
353.00	1647.0
354.00	2006.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

354.90	412.0
358.90	290.0
365.00	10655.0
365.90	1643.0
367.00	184.0
370.00	161.0
371.00	636.0
372.00	4196.0
373.10	980.0
382.90	859.0
383.90	367.0
390.00	462.0
391.10	318.0
392.00	290.0
401.00	254.0
401.80	1403.0
403.00	2464.0
404.00	921.0
404.90	345.0
420.90	1794.0
422.00	2248.0
423.00	12645.0
424.00	3684.0
424.90	666.0
440.10	254.0
441.00	35320.0
442.00	195712.0
443.00	41696.0
444.00	3637.0
445.00	285.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

m/z	Abundance
36.00	207.0
37.00	840.0
38.00	2522.0
39.00	12307.0
40.00	1143.0
41.00	2768.0
42.00	2550.0
43.00	13639.0
44.00	2035.0
45.00	10823.0
48.00	154.0
49.00	1270.0
50.00	42720.0
51.00	155776.0
52.00	8082.0
53.00	463.0
55.00	1514.0
56.00	5561.0
57.00	10781.0
57.90	380.0
58.90	165.0
60.00	7201.0
61.00	2352.0
62.00	2749.0
63.00	6218.0
64.00	944.0
65.00	3306.0
65.90	400.0
67.00	641.0
68.00	3234.0
69.00	170240.0
70.00	986.0
71.10	183.0
72.00	158.0
73.00	2666.0
74.00	20368.0
75.00	28912.0
76.00	9930.0
77.00	202752.0
78.00	14808.0
79.00	13701.0
80.00	10132.0
81.00	13960.0
82.00	3206.0
83.00	3468.0
84.00	658.0
85.00	2678.0
86.00	4306.0
87.00	1854.0
88.00	795.0
89.00	360.0
91.00	3252.0
92.00	3837.0
93.00	23864.0
94.00	1547.0
95.00	503.0
96.10	1077.0
97.00	530.0
98.00	19616.0
99.00	14858.0
99.90	1212.0
101.00	7577.0
101.90	457.0
103.00	2795.0
104.00	4956.0
105.00	4277.0
106.00	1891.0
107.00	60960.0
108.00	8812.0
109.10	1939.0
110.00	101592.0
111.00	15785.0
112.00	2312.0
113.00	861.0
114.00	236.0
114.80	425.0
115.20	450.0
116.00	3568.0
117.00	62872.0
118.00	4587.0
119.00	518.0
120.00	844.0
121.00	527.0
122.00	4003.0
123.00	6563.0
124.00	2645.0
125.00	2546.0
127.00	196608.0
128.00	15319.0
129.00	78704.0
130.00	6572.0
131.00	1500.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

131.90	688.0
132.90	484.0
134.00	2318.0
135.00	6977.0
135.90	2625.0
137.00	4046.0
137.90	694.0
139.00	551.0
140.00	967.0
141.00	10860.0
142.00	3298.0
143.00	2503.0
144.00	815.0
145.10	518.0
146.00	1894.0
147.00	5931.0
148.00	14667.0
149.00	3398.0
150.00	1005.0
151.00	1090.0
151.70	817.0
151.90	779.0
153.00	3436.0
154.00	2577.0
155.00	6289.0
156.00	8460.0
157.00	1942.0
157.90	1900.0
159.00	1284.0
160.00	3473.0
161.00	5480.0
161.90	1549.0
163.00	626.0
164.00	448.0
165.00	4085.0
166.00	3286.0
167.00	25960.0
168.00	15009.0
169.00	2100.0
170.10	720.0
171.00	1010.0
172.00	1797.0
173.00	2740.0
174.00	5177.0
175.00	8756.0
176.00	2160.0
177.00	4110.0
177.90	1556.0
179.00	17576.0
180.00	12780.0
181.00	5616.0
181.90	935.0
182.70	436.0
183.10	426.0
184.00	1489.0
185.00	8320.0
186.00	62328.0
187.00	16880.0
188.00	2180.0
189.00	3915.0
190.00	835.0
191.00	2428.0
192.00	6197.0
193.00	6692.0
194.00	1619.0
195.00	739.0
196.00	13198.0
197.90	436288.0
199.00	31952.0
200.00	2460.0
201.60	2188.0
203.00	3753.0
204.00	17608.0
205.00	29152.0
206.10	125256.0
207.10	18048.0
208.00	4758.0
209.00	1658.0
210.00	1773.0
211.10	4922.0
213.20	321.0
213.90	206.0
215.00	1408.0
215.90	3031.0
217.00	34112.0
218.00	4432.0
218.90	441.0
220.20	833.0
221.00	19384.0
223.00	8022.0
224.00	67416.0
225.00	17240.0
226.00	2206.0
227.00	29704.0
228.00	4440.0
229.00	5668.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

230.00	912.0
231.00	2737.0
232.00	524.0
232.90	444.0
234.00	1940.0
234.90	1904.0
236.00	1658.0
237.00	3136.0
238.00	386.0
239.00	1175.0
240.00	1257.0
241.00	1958.0
242.00	3461.0
243.00	4024.0
244.00	59544.0
245.00	7821.0
246.00	11341.0
247.00	2313.0
248.00	556.0
249.00	2005.0
249.90	472.0
250.90	620.0
252.00	703.0
253.00	1306.0
254.10	3189.0
255.00	271808.0
256.00	42544.0
257.00	3307.0
258.00	16512.0
259.00	2926.0
260.00	449.0
260.90	488.0
263.00	214.0
264.00	513.0
265.00	7547.0
265.90	1159.0
267.00	387.0
268.00	224.0
269.10	190.0
270.10	303.0
271.00	682.0
271.90	761.0
273.00	8039.0
274.00	23792.0
275.00	128016.0
276.10	17832.0
276.90	11125.0
278.00	2195.0
279.00	253.0
281.10	722.0
282.00	479.0
283.00	1225.0
284.00	873.0
285.00	1776.0
285.90	376.0
289.00	591.0
290.00	452.0
290.80	256.0
292.10	378.0
293.10	2746.0
294.00	801.0
295.00	937.0
296.00	38920.0
297.00	5568.0
297.90	271.0
301.00	409.0
302.00	667.0
303.00	4177.0
304.00	906.0
304.80	192.0
308.00	637.0
308.90	392.0
310.00	409.0
310.90	159.0
313.00	360.0
314.00	1800.0
315.00	4543.0
316.00	2122.0
316.90	444.0
319.80	152.0
321.10	1266.0
322.00	694.0
323.00	11087.0
324.00	2153.0
324.90	267.0
325.90	268.0
327.00	2271.0
327.90	1093.0
328.90	293.0
332.00	744.0
332.90	1243.0
334.00	7428.0
334.90	1962.0
335.90	207.0
340.00	219.0
340.90	1282.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

341.80	372.0
345.90	2679.0
347.00	459.0
351.10	306.0
352.00	3625.0
353.00	2615.0
354.00	4119.0
355.00	916.0
356.10	153.0
358.90	252.0
364.90	18536.0
365.90	2458.0
366.90	234.0
370.00	320.0
370.90	909.0
372.00	6390.0
372.90	1720.0
374.00	199.0
382.90	1933.0
384.00	602.0
384.90	191.0
389.90	870.0
391.00	533.0
392.00	529.0
400.90	555.0
402.00	2988.0
402.90	3341.0
404.00	1634.0
404.90	371.0
415.00	170.0
421.00	3498.0
422.00	2846.0
423.00	22232.0
424.00	5558.0
425.00	1157.0
439.00	158.0
441.00	57704.0
442.00	379264.0
443.10	74184.0
444.00	6407.0
445.00	245.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

m/z	Abundance
35.00	158.0
37.00	794.0
38.10	1558.0
39.00	9438.0
40.00	1051.0
41.00	2675.0
42.00	2781.0
43.00	15017.0
44.00	2030.0
45.00	11953.0
45.90	194.0
49.00	960.0
50.00	30776.0
51.00	115512.0
52.00	6294.0
53.00	282.0
55.00	1494.0
56.00	4373.0
57.00	8028.0
57.90	406.0
58.90	192.0
60.00	8288.0
61.00	1895.0
62.00	1972.0
63.00	4629.0
64.00	754.0
65.00	2464.0
66.00	316.0
67.00	557.0
68.00	2432.0
69.00	123016.0
70.00	824.0
70.90	208.0
72.00	163.0
73.00	2383.0
74.00	14582.0
75.00	21312.0
76.00	8120.0
77.00	153024.0
78.00	10523.0
79.00	10824.0
80.00	7534.0
81.00	10922.0
82.00	2876.0
82.90	2436.0
83.90	411.0
84.90	2187.0
86.00	3152.0
87.00	1448.0
87.90	436.0
88.90	363.0
91.00	2597.0
92.00	2640.0
93.00	19376.0
94.00	1243.0
95.00	507.0
96.00	1003.0
97.10	659.0
98.00	15287.0
99.00	11254.0
100.00	1163.0
101.00	5427.0
101.90	289.0
103.00	1963.0
103.90	3759.0
105.00	4132.0
105.90	1193.0
107.00	46208.0
108.00	7021.0
109.00	1480.0
110.00	78696.0
111.00	11891.0
112.00	1511.0
113.00	531.0
113.90	255.0
115.00	535.0
116.00	2658.0
117.00	50656.0
118.00	3618.0
118.90	598.0
120.00	712.0
120.90	350.0
122.00	2965.0
123.00	5578.0
123.90	2512.0
125.00	2296.0
126.10	1175.0
127.00	153280.0
128.00	11455.0
129.00	61920.0
130.00	5016.0
131.00	1207.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

132.00	581.0
132.90	459.0
134.00	1819.0
135.00	5534.0
136.00	2577.0
137.00	3020.0
137.90	670.0
138.90	517.0
139.90	921.0
141.00	8604.0
141.90	3035.0
142.90	1739.0
143.90	622.0
145.00	528.0
146.00	1744.0
147.00	5016.0
148.00	11976.0
149.00	2634.0
150.00	810.0
151.10	1137.0
151.80	868.0
153.00	2862.0
154.00	2053.0
155.00	4677.0
156.10	6935.0
157.00	1616.0
157.90	1698.0
158.90	1369.0
160.00	2787.0
161.00	4253.0
162.00	1109.0
163.00	540.0
163.90	400.0
165.00	3761.0
166.00	2683.0
167.00	22248.0
168.00	12194.0
169.00	1924.0
170.00	663.0
171.00	1036.0
171.90	1524.0
173.00	2188.0
174.00	3945.0
175.00	7358.0
176.00	2023.0
177.00	3276.0
177.90	1144.0
178.90	13338.0
180.00	10033.0
181.00	4635.0
182.00	748.0
182.90	272.0
184.00	1352.0
185.00	7512.0
186.00	52160.0
187.00	15545.0
188.00	1568.0
189.00	3746.0
189.90	755.0
191.00	1787.0
192.00	4862.0
193.00	5285.0
194.00	999.0
195.00	680.0
196.00	10447.0
197.90	372160.0
199.00	24344.0
200.00	2256.0
201.50	1702.0
203.00	2762.0
204.00	13534.0
205.00	24840.0
206.10	104608.0
207.00	14435.0
208.00	3822.0
209.00	1293.0
210.10	1702.0
211.00	3997.0
211.90	852.0
213.00	328.0
213.90	208.0
215.00	1136.0
216.00	2632.0
216.90	28072.0
218.00	3824.0
219.00	352.0
221.00	15501.0
221.90	4033.0
223.00	6508.0
224.00	61112.0
225.00	15584.0
226.00	1718.0
227.00	27264.0
228.00	3586.0
229.00	5138.0
230.00	729.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

231.00	2144.0
231.90	409.0
233.10	597.0
234.00	1684.0
235.00	1810.0
236.00	1522.0
236.90	2813.0
238.00	375.0
239.00	1269.0
240.00	863.0
240.90	1396.0
242.00	3180.0
243.00	3499.0
244.10	49968.0
245.00	6246.0
246.00	9634.0
247.00	2251.0
247.90	455.0
249.00	1875.0
250.00	356.0
251.00	476.0
252.00	508.0
253.00	1262.0
254.10	2444.0
255.00	243456.0
256.00	36120.0
257.00	2806.0
258.00	14428.0
259.00	2735.0
259.90	599.0
260.90	388.0
263.90	511.0
265.00	6408.0
266.00	928.0
267.00	256.0
270.00	360.0
270.90	710.0
272.00	581.0
273.00	7372.0
274.00	22344.0
275.00	101624.0
276.00	14666.0
277.00	11054.0
278.00	1819.0
279.00	499.0
280.90	596.0
282.00	455.0
283.00	1187.0
284.10	659.0
285.00	1552.0
286.00	477.0
288.90	454.0
289.90	286.0
290.90	389.0
292.00	462.0
293.00	2256.0
293.90	533.0
295.00	922.0
296.00	33976.0
297.00	4910.0
298.00	491.0
301.00	294.0
301.90	542.0
303.00	3909.0
304.00	1106.0
307.80	405.0
308.90	355.0
309.80	436.0
313.00	337.0
314.00	1726.0
314.90	3626.0
316.00	2164.0
317.00	353.0
319.90	216.0
321.00	1134.0
322.00	613.0
323.10	10900.0
324.00	1792.0
325.00	228.0
325.70	240.0
326.90	2088.0
327.90	1274.0
328.90	186.0
332.00	801.0
333.00	886.0
334.00	7545.0
335.00	1611.0
335.90	239.0
338.90	271.0
340.00	281.0
341.00	1173.0
341.90	540.0
345.90	2544.0
346.90	431.0
351.00	209.0
352.00	3644.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273

353.00	2437.0
354.00	3793.0
355.00	987.0
365.00	16074.0
366.00	2571.0
366.90	187.0
369.90	320.0
371.00	770.0
372.00	7039.0
373.00	1358.0
374.00	297.0
377.00	207.0
382.90	1936.0
384.00	526.0
385.00	215.0
390.00	627.0
390.90	468.0
391.90	478.0
401.00	396.0
401.90	2553.0
403.00	3368.0
404.00	1416.0
404.90	238.0
414.80	214.0
421.00	3722.0
421.90	3078.0
423.00	21216.0
424.00	6101.0
425.00	712.0
439.10	229.0
441.00	53896.0
442.00	375168.0
443.00	69120.0
444.00	7656.0
445.10	473.0

Instrument :
BNA_N
ClientSampleId :
DFTPP273