

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020223\
 Data File : BN023936.D
 Acq On : 03 Feb 2023 10:52
 Operator : CG/JU
 Sample : PB150528BL
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK528

Quant Time: Feb 03 21:50:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 02:10:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	156167	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	654316	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.498	164	448387	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.251	188	1008898	20.000	ng/u1	0.00
79) Chrysene-d12	21.445	240	845803	20.000	ng/u1	0.00
88) Perylene-d12	23.845	264	798682	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.240	96	20462	5.288	ng/uL	0.00
4) Pyridine-d5	3.658	84	296285	27.406	ng/u1	0.00
7) Phenol-d5	7.016	99	366609	26.882	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.181	67	242028	28.082	ng/u1	0.00
11) 2-Chlorophenol-d4	7.375	132	298854	28.027	ng/u1	0.00
15) 4-Methylphenol-d8	8.563	113	296917	27.446	ng/u1	-0.01
21) Nitrobenzene-d5	9.016	128	146501	27.596	ng/u1	0.00
24) 2-Nitrophenol-d4	9.740	143	169953	28.136	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	302615	26.831	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.799	131	425813	27.674	ng/u1	-0.01
46) Dimethylphthalate-d6	13.916	166	1059530	29.597	ng/u1	0.00
49) Acenaphthylene-d8	14.193	160	1120449	28.934	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	156278	24.079	ng/u1	0.00
60) Fluorene-d10	15.492	176	875752	29.181	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	156340	21.899	ng/u1	0.00
73) Anthracene-d10	17.345	188	1416790	30.164	ng/u1	0.00
81) Pyrene-d10	19.645	212	1622424	31.595	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.686	264	1309178	31.582	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020223\
Data File : BN023936.D
Acq On : 03 Feb 2023 10:52
Operator : CG/JU
Sample : PB150528BL
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK528

Quant Time: Feb 03 21:50:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020123.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 02 02:10:45 2023
Response via : Initial Calibration

