

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020623\
 Data File : BN023958.D
 Acq On : 06 Feb 2023 18:25
 Operator : CG/JU
 Sample : 01362-12MEDL2 250X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4488MEDL2

Quant Time: Feb 07 03:01:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 07 02:28:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	142133	20.000	ng/u1	0.00
20) Naphthalene-d8	10.645	136	578713	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	385412	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	850154	20.000	ng/u1	0.00
79) Chrysene-d12	21.433	240	715004	20.000	ng/u1	0.00
88) Perylene-d12	23.827	264	701846	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/uL	
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.010	99	1615	0.133	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.175	67	1045	0.136	ng/u1	0.00
11) 2-Chlorophenol-d4	7.369	132	1173	0.124	ng/u1	0.00
15) 4-Methylphenol-d8	8.563	113	1126	0.119	ng/u1	0.00
21) Nitrobenzene-d5	9.010	128	499	0.108	ng/u1	0.00
24) 2-Nitrophenol-d4	9.734	143	473	0.091	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	1224	0.127	ng/u1	0.00
31) 4-Chloroaniline-d4	10.804	131	1384	0.106	ng/u1	0.01
46) Dimethylphthalate-d6	13.904	166	4028	0.135	ng/u1	0.00
49) Acenaphthylene-d8	14.186	160	3798	0.115	ng/u1	0.00
54) 4-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
60) Fluorene-d10	15.486	176	4187	0.167	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	0.000	200	0d	0.000	ng/u1	
73) Anthracene-d10	17.339	188	6802	0.176	ng/u1	0.00
81) Pyrene-d10	19.633	212	6391	0.148	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.656	264	5430	0.151	ng/u1	-0.02
Target Compounds						
86) Bis(2-ethylhexyl)phtha...	21.345	149	830999	27.402	ng/u1	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020623\
Data File : BN023958.D
Acq On : 06 Feb 2023 18:25
Operator : CG/JU
Sample : 01362-12MEDL2 250X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4488MEDL2

Quant Time: Feb 07 03:01:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN020623.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 07 02:28:26 2023
Response via : Initial Calibration





#86
 Bis(2-ethylhexyl)phthalate
 Concen: 27.402 ng/ul
 RT: 21.345 min Scan# 31
 Delta R.T. -0.006 min
 Lab File: BN023958.D
 Acq: 06 Feb 2023 18:25

Instrument :
 BNA_N
 ClientSampleId :
 A4488MEDL2

Tgt Ion:149 Resp: 830999
 Ion Ratio Lower Upper
 149 100
 167 27.0 22.7 34.1
 279 4.9 3.9 5.9

