

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN020821\
 Data File : BN013599.D
 Acq On : 08 Feb 2021 16:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN020821

Quant Time: Feb 08 16:44:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN020821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 08 15:13:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.448	0.425	5.1	100	0.00
3	n-Nitrosodimethylamine	0.326	0.326	0.0	103	0.00
4 S	2-Fluorophenol	0.866	0.956	-10.4	97	0.00
5 S	Phenol-d6	1.082	1.165	-7.7	96	0.00
6	bis(2-Chloroethyl)ether	0.965	0.929	3.7	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.222	0.219	1.4	98	-0.01
9	Naphthalene	0.982	0.964	1.8	98	-0.01
10	Hexachlorobutadiene	0.163	0.160	1.8	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.624	0.606	2.9	98	0.00
12	2-Methylnaphthalene	0.672	0.656	2.4	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.114	0.100	12.3	92	0.00
15 S	2-Fluorobiphenyl	1.501	1.584	-5.5	98	0.00
16	Acenaphthylene	1.534	1.384	9.8	97	0.00
17	Acenaphthene	1.103	1.067	3.3	99	0.00
18	Fluorene	1.383	1.340	3.1	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.039	0.032	17.9	95	0.00
21	4-Bromophenyl-phenylether	0.198	0.188	5.1	95	-0.01
22	Hexachlorobenzene	0.219	0.214	2.3	99	0.00
23	Atrazine	0.127	0.111	12.6	97	0.00
24	Pentachlorophenol	0.045	0.034	24.4	88	0.00
25	Phenanthrene	1.091	1.071	1.8	100	0.00
26	Anthracene	0.897	0.835	6.9	100	0.00
27 SURR	Fluoranthene-d10	1.077	1.028	4.5	100	0.00
28	Fluoranthene	1.195	1.139	4.7	100	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
30	Pyrene	1.197	1.189	0.7	100	0.00
31 S	Terphenyl-d14	0.890	0.949	-6.6	102	0.00
32	Benzo(a)anthracene	1.066	0.971	8.9	98	0.00
33	Chrysene	1.222	1.225	-0.2	104	0.00
34	Bis(2-ethylhexyl)phthalate	0.314	0.295	6.1	104	0.00
35	Indeno(1,2,3-cd)pyrene	1.277	1.249	2.2	103	0.00
36 I	Perylene-d12	1.000	1.000	0.0	103	0.00
37	Benzo(b)fluoranthene	1.391	1.354	2.7	102	0.00
38	Benzo(k)fluoranthene	1.403	1.404	-0.1	107	0.00
39 C	Benzo(a)pyrene	1.177	1.120	4.8	103	0.00
40	Dibenzo(a,h)anthracene	1.264	1.221	3.4	104	0.00
41	Benzo(g,h,i)perylene	1.327	1.306	1.6	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN020821