


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN020821\  
 Data File : BN013591.D  
 Acq On : 08 Feb 2021 10:14  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN020821.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Feb 08 15:13:18 2021






AutoFind: Scans 2433, 2434, 2435; Background Corrected with Scan 2427

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 35.9         | 175120     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.7          | 2917       | PASS                |
| 69             | 198             | 0.00            | 100             | 36.1         | 176384     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 872        | PASS                |
| 127            | 198             | 10              | 80              | 48.3         | 235520     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.2          | 885        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 487979     | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 32960      | PASS                |
| 275            | 198             | 10              | 60              | 26.0         | 126901     | PASS                |
| 365            | 198             | 1               | 100             | 2.9          | 14151      | PASS                |
| 441            | 198             | 0.01            | 100             | 13.0         | 63440      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 420587     | PASS                |
| 443            | 442             | 15              | 24              | 19.3         | 81304      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN020821\  
Data File : BN013591.D  
Acq On : 08 Feb 2021 10:14  
Operator : CG/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: Feb 16 05:39:51 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Feb 16 05:39:20 2021  
Response via : Initial Calibration



TIC: BN013591.D\data.ms

(77) Benzidine

19.204min ( 0.000) 0.00 ng




response 3684862

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.47  |
| 183.00 | 13.20  | 11.50  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN020821\  
 Data File : BN013591.D  
 Acq On : 08 Feb 2021 10:14  
 Operator : CG/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Quant Time: Feb 16 05:39:51 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270E-Tune.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Feb 16 05:39:20 2021  
 Response via : Initial Calibration



TIC: BN013591.D\data.ms

**(70) Pentachlorophenol (C)**

16.587min ( 0.000) 24714.54 ng

response 562853

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 64.78  |
| 264.00 | 61.60  | 60.74  |
| 0.00   | 0.00   | 0.00   |

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 2/8/2021      | BNA_N            | <u>BN013591.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 1889698          | 20.44              |
| DDD           | 5526             | 20                 |
| DDE           | 2105             | 19.49              |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 7631          | 1897329          | 0.40               |
|               |                  |                    |