

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN021723\  
 Data File : BN024061.D  
 Acq On : 17 Feb 2023 11:33  
 Operator : CG/JU  
 Sample : PB150879BL  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK879

Quant Time: Feb 17 21:35:27 2023  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN013023.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Feb 17 14:45:13 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.829  | 152  | 5291     | 0.400 | ng/u1 | 0.00     |
| 4) Naphthalene-d8           | 10.634 | 136  | 14198    | 0.400 | ng/u1 | 0.00     |
| 9) Acenaphthene-d10         | 14.483 | 164  | 8153     | 0.400 | ng/u1 | 0.00     |
| 13) Phenanthrene-d10        | 17.234 | 188  | 17814    | 0.400 | ng/u1 | # 0.00   |
| 17) Chrysene-d12            | 21.430 | 240  | 12981    | 0.400 | ng/u1 | 0.00     |
| 23) Perylene-d12            | 23.819 | 264  | 10639    | 0.400 | ng/u1 | 0.00     |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.226  | 96   | 29230    | 5.154 | ng/u1 | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.228 | 152  | 5508     | 0.248 | ng/u1 | 0.00     |
| 18) Fluoranthene-d10        | 19.262 | 212  | 12014    | 0.345 | ng/u1 | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN021723\  
Data File : BN024061.D  
Acq On : 17 Feb 2023 11:33  
Operator : CG/JU  
Sample : PB150879BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK879

Quant Time: Feb 17 21:35:27 2023  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN013023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Feb 17 14:45:13 2023  
Response via : Initial Calibration

