

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022123\
 Data File : BN024075.D
 Acq On : 21 Feb 2023 10:00
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4383

Quant Time: Feb 22 01:29:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN013023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 17 14:45:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.830	152	4815	0.400	ng/ul	0.00
4) Naphthalene-d8	10.632	136	12225	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.482	164	6325	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.233	188	12080	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.426	240	9789	0.400	ng/ul	# 0.00
23) Perylene-d12	23.805	264	7758	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.231	96	1905	0.369	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.227	152	6158	0.322	ng/ul	0.00
18) Fluoranthene-d10	19.261	212	10744	0.409	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.265	88	1972	0.361	ng/ul#	84
5) Naphthalene	10.682	128	12319	0.368	ng/ul	99
7) 2-Methylnaphthalene	12.299	142	7386	0.340	ng/ul	100
8) 1-Methylnaphthalene	12.519	142	7643	0.343	ng/ul	100
10) Acenaphthylene	14.200	152	9050	0.374	ng/ul	99
11) Acenaphthene	14.542	153	8062	0.379	ng/ul	99
12) Fluorene	15.532	166	8704	0.338	ng/ul	97
14) Pentachlorophenol	16.886	266	1127	0.283	ng/ul	100
15) Phenanthrene	17.271	178	13282	0.369	ng/ul	99
16) Anthracene	17.364	178	10733	0.356	ng/ul	99
19) Fluoranthene	19.293	202	14027	0.431	ng/ul#	95
20) Pyrene	19.656	202	14047	0.417	ng/ul#	94
21) Benzo(a)anthracene	21.409	228	11857	0.357	ng/ul	99
22) Chrysene	21.461	228	13510	0.386	ng/ul	98
24) Benzo(b)fluoranthene	23.080	252	13964	0.407	ng/ul	98
25) Benzo(k)fluoranthene	23.130	252	13423	0.401	ng/ul#	96
26) Benzo(a)pyrene	23.697	252	10515	0.368	ng/ul#	93
27) Indeno(1,2,3-cd)pyrene	26.255	276	15087	0.377	ng/ul#	95
28) Dibenzo(a,h)anthracene	26.275	278	11853	0.370	ng/ul#	96
29) Benzo(g,h,i)perylene	27.013	276	13425	0.407	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022123\
Data File : BN024075.D
Acq On : 21 Feb 2023 10:00
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4383

Quant Time: Feb 22 01:29:37 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN013023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 17 14:45:13 2023
Response via : Initial Calibration

