

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018676.D
 Acq On : 22 Feb 2022 14:25
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Feb 22 15:04:56 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 13:54:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.919	152	6676	0.400	ng	0.00
7) Naphthalene-d8	10.722	136	15453	0.400	ng	# 0.00
13) Acenaphthene-d10	14.549	164	7530	0.400	ng	0.00
19) Phenanthrene-d10	17.287	188	13278	0.400	ng	0.00
29) Chrysene-d12	21.458	240	10295	0.400	ng	# 0.00
35) Perylene-d12	23.858	264	8400	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.471	112	24780	1.461	ng	0.00
5) Phenol-d6	7.067	99	29192	1.570	ng	0.00
8) Nitrobenzene-d5	9.067	82	19697	1.844	ng	0.00
11) 2-Methylnaphthalene-d10	12.310	152	38889	1.559	ng	0.00
14) 2,4,6-Tribromophenol	16.025	330	5320	1.955	ng	0.00
15) 2-Fluorobiphenyl	13.180	172	57908	1.939	ng	0.00
27) Fluoranthene-d10	19.312	212	58660	1.501	ng	0.00
31) Terphenyl-d14	19.906	244	36324	1.389	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.326	88	12758	2.685	ng	# 58
3) n-Nitrosodimethylamine	3.637	42	13444	1.945	ng	89
6) bis(2-Chloroethyl)ether	7.327	93	30393	1.543	ng	100
9) Naphthalene	10.776	128	73357	1.813	ng	100
10) Hexachlorobutadiene	11.075	225	17401	2.107	ng	# 100
12) 2-Methylnaphthalene	12.382	142	48820	1.795	ng	98
16) Acenaphthylene	14.271	152	60253	1.901	ng	100
17) Acenaphthene	14.613	154	40731	1.923	ng	98
18) Fluorene	15.586	166	49209	1.894	ng	98
20) 4,6-Dinitro-2-methylph...	15.660	198	3896	4.055	ng	# 80
21) 4-Bromophenyl-phenylether	16.477	248	15976	1.984	ng	93
22) Hexachlorobenzene	16.600	284	19393	2.153	ng	# 88
23) Atrazine	16.733	200	8650	1.769	ng	91
24) Pentachlorophenol	16.938	266	4810	1.832	ng	99
25) Phenanthrene	17.328	178	73554	1.894	ng	100
26) Anthracene	17.420	178	62799	1.904	ng	100
28) Fluoranthene	19.341	202	76896	1.797	ng	# 97
30) Pyrene	19.704	202	74876	2.123	ng	100
32) Benzo(a)anthracene	21.449	228	63872	1.867	ng	99
33) Chrysene	21.494	228	70502	2.033	ng	99
34) Bis(2-ethylhexyl)phtha...	21.377	149	24043	1.735	ng	# 96
36) Indeno(1,2,3-cd)pyrene	26.349	276	73586	2.436	ng	# 90
37) Benzo(b)fluoranthene	23.130	252	63105	2.182	ng	# 97
38) Benzo(k)fluoranthene	23.176	252	63109	2.184	ng	# 97
39) Benzo(a)pyrene	23.752	252	50620	2.188	ng	# 96
40) Dibenzo(a,h)anthracene	26.363	278	57934	2.460	ng	# 94
41) Benzo(g,h,i)perylene	27.106	276	62507	2.378	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018676.D
 Acq On : 22 Feb 2022 14:25
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Quant Time: Feb 22 15:04:56 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 13:54:19 2022
 Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.919 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:152 Resp: 6676
Ion Ratio Lower Upper
152 100
150 154.9 121.7 182.5
115 59.8 43.8 65.8



#2
1,4-Dioxane
Concen: 2.685 ng
RT: 3.326 min Scan# 34
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion: 88 Resp: 12758
Ion Ratio Lower Upper
88 100
43 30.9 69.7 104.5#
58 77.8 83.8 125.8#









#3
n-Nitrosodimethylamine
Concen: 1.945 ng
RT: 3.637 min Scan# 71
Delta R.T. -0.007 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 13444
Ion Ratio Lower Upper
42 100
74 121.6 108.6 163.0
44 6.6 5.8 8.8



#4
2-Fluorophenol
Concen: 1.461 ng
RT: 5.471 min Scan# 331
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion: 112 Resp: 24780
Ion Ratio Lower Upper
112 100
64 59.2 48.0 72.0
63 31.7 25.3 37.9









#5
Phenol-d6
Concen: 1.570 ng
RT: 7.067 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion: 99 Resp: 29192
Ion Ratio Lower Upper
99 100
42 20.1 16.3 24.5
71 34.6 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 1.543 ng
RT: 7.327 min Scan# 588
Delta R.T. -0.007 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion: 93 Resp: 30393
Ion Ratio Lower Upper
93 100
63 77.7 61.7 92.5
95 32.8 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.722 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	15453
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.8	13.2
54	8.9	5.0	7.6#
68	6.7	4.4	6.6#



#8
Nitrobenzene-d5
Concen: 1.844 ng
RT: 9.067 min Scan# 795
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:	82	Resp:	19697
Ion Ratio	Lower	Upper	
82	100		
128	37.1	33.7	50.5
54	53.1	36.6	55.0









#9
Naphthalene
Concen: 1.813 ng
RT: 10.776 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 73357
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 12.9 10.4 15.6



#10
Hexachlorobutadiene
Concen: 2.107 ng
RT: 11.075 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:225 Resp: 17401
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 51.0 76.6





#11

2-Methylnaphthalene-d10

Concen: 1.559 ng

RT: 12.310 min Scan# 1235

Delta R.T. -0.000 min

Lab File: BN018676.D

Acq: 22 Feb 2022 14:25

Instrument :

BNA_N

ClientSampleId :





SSTDICC1.6

Tgt Ion:152 Resp: 38889

Ion Ratio Lower Upper

152 100

151 20.6 16.2 24.2



#12

2-Methylnaphthalene

Concen: 1.795 ng

RT: 12.382 min Scan# 1254

Delta R.T. -0.000 min

Lab File: BN018676.D

Acq: 22 Feb 2022 14:25




Tgt Ion:142 Resp: 48820


Ion Ratio Lower Upper

142 100

141 87.7 70.0 105.0

115 37.6 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.549 min Scan# 1478
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 7530
Ion Ratio Lower Upper
164 100
162 103.7 80.2 120.2
160 48.3 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 1.955 ng
RT: 16.025 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:330 Resp: 5320
Ion Ratio Lower Upper
330 100
332 95.9 0.0 0.0#
141 38.3 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 1.939 ng
RT: 13.180 min Scan# 1350
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 57908
Ion Ratio Lower Upper
172 100
171 34.0 27.4 41.0
170 22.4 18.4 27.6



#16
Acenaphthylene
Concen: 1.901 ng
RT: 14.271 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:152 Resp: 60253
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.6
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 1.923 ng
RT: 14.613 min Scan# 1484
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 40731
Ion Ratio Lower Upper
154 100
153 111.2 90.9 136.3
152 55.0 45.0 67.4



#18
Fluorene
Concen: 1.894 ng
RT: 15.586 min Scan# 1575
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:166 Resp: 49209
Ion Ratio Lower Upper
166 100
165 99.2 77.8 116.6
167 13.6 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.287 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN018676.D

Acq: 22 Feb 2022 14:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 13278

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.2 7.0 10.4



#20

4,6-Dinitro-2-methylphenol

Concen: 4.055 ng

RT: 15.660 min Scan# 1582

Delta R.T. -0.000 min

Lab File: BN018676.D

Acq: 22 Feb 2022 14:25




Tgt Ion:198 Resp: 3896


Ion Ratio Lower Upper

198 100

182 10.3 15.5 23.3#

77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.984 ng
RT: 16.477 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 15976
Ion Ratio Lower Upper
248 100
250 97.0 74.4 111.6
141 63.3 43.7 65.5



#22
Hexachlorobenzene
Concen: 2.153 ng
RT: 16.600 min Scan# 1673
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:284 Resp: 19393
Ion Ratio Lower Upper
284 100
142 36.6 21.6 32.4#
249 32.4 23.5 35.3









#23
 Atrazine
 Concen: 1.769 ng
 RT: 16.733 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: BN018676.D
 Acq: 22 Feb 2022 14:25




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:200 Resp: 8650
 Ion Ratio Lower Upper
 200 100
 173 29.8 20.3 30.5
 215 44.5 40.4 60.6



#24
 Pentachlorophenol
 Concen: 1.832 ng
 RT: 16.938 min Scan# 1706
 Delta R.T. -0.000 min
 Lab File: BN018676.D
 Acq: 22 Feb 2022 14:25

Tgt Ion:266 Resp: 4810
 Ion Ratio Lower Upper
 266 100
 264 63.1 49.5 74.3
 268 64.4 50.6 76.0









#25
Phenanthrene
Concen: 1.894 ng
RT: 17.328 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 73554
Ion Ratio Lower Upper
178 100
176 19.2 15.3 22.9
179 15.2 12.3 18.5



#26
Anthracene
Concen: 1.904 ng
RT: 17.420 min Scan# 1753
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:178 Resp: 62799
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.501 ng
RT: 19.312 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 58660
Ion Ratio Lower Upper
212 100
106 15.3 8.4 12.6#
104 8.5 4.5 6.7#



#28
Fluoranthene
Concen: 1.797 ng
RT: 19.341 min Scan# 2127
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:202 Resp: 76896
Ion Ratio Lower Upper
202 100
101 11.7 7.2 10.8#
203 17.1 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.458 min Scan# 2448
Delta R.T. -0.009 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 10295
Ion Ratio Lower Upper
240 100
120 18.7 6.0 9.0#
236 30.5 21.4 32.2



#30
Pyrene
Concen: 2.123 ng
RT: 19.704 min Scan# 2213
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:202 Resp: 74876
Ion Ratio Lower Upper
202 100
200 21.0 16.6 25.0
203 17.6 14.1 21.1









#31
 Terphenyl-d14
 Concen: 1.389 ng
 RT: 19.906 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN018676.D
 Acq: 22 Feb 2022 14:25




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:244 Resp: 36324
 Ion Ratio Lower Upper
 244 100
 212 10.1 6.2 9.2#
 122 0.0 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 1.867 ng
 RT: 21.449 min Scan# 2446
 Delta R.T. -0.000 min
 Lab File: BN018676.D
 Acq: 22 Feb 2022 14:25

Tgt Ion:228 Resp: 63872
 Ion Ratio Lower Upper
 228 100
 226 27.0 21.3 31.9
 229 19.4 15.9 23.9





#33

Chrysene

Concen: 2.033 ng

RT: 21.494 min Scan# 2451

Delta R.T. -0.009 min

Lab File: BN018676.D

Acq: 22 Feb 2022 14:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 70502

Ion Ratio Lower Upper

228 100

226 31.1 24.2 36.4

229 18.9 15.5 23.3



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.735 ng

RT: 21.377 min Scan# 2438

Delta R.T. -0.000 min

Lab File: BN018676.D

Acq: 22 Feb 2022 14:25




Tgt Ion:149 Resp: 24043


Ion Ratio Lower Upper

149 100

167 27.5 23.7 35.5

279 2.7 3.6 5.4#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.858 min Scan# 3049
Delta R.T. -0.003 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264 Resp: 8400
Ion Ratio Lower Upper
264 100
260 28.1 20.2 30.4
265 35.6 18.2 27.2#



#36
Indeno(1,2,3-cd)pyrene
Concen: 2.436 ng
RT: 26.349 min Scan# 3901
Delta R.T. -0.003 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:276 Resp: 73586
Ion Ratio Lower Upper
276 100
138 28.9 14.9 22.3#
277 24.8 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 2.182 ng
RT: 23.130 min Scan# 2800
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 63105
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 11.5 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 2.184 ng
RT: 23.176 min Scan# 2816
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Tgt Ion:252 Resp: 63109
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 11.6 6.2 9.2#









#39
Benzo(a)pyrene
Concen: 2.188 ng
RT: 23.752 min Scan# 3013
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 50620
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 13.0 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 2.460 ng
RT: 26.363 min Scan# 3906
Delta R.T. -0.000 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25


Tgt Ion:278 Resp: 57934
Ion Ratio Lower Upper
278 100
139 20.1 10.8 16.2#
279 23.8 19.0 28.4





#41
Benzo(g,h,i)perylene
Concen: 2.378 ng
RT: 27.106 min Scan# 41
Delta R.T. -0.003 min
Lab File: BN018676.D
Acq: 22 Feb 2022 14:25

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	276	Resp:	62507
Ion	Ratio	Lower	Upper
276	100		
277	23.5	19.0	28.6
138	25.2	12.6	19.0#

