

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018677.D
 Acq On : 22 Feb 2022 15:01
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Feb 22 15:52:47 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 13:54:19 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.919	152	6859	0.400	ng	0.00
7) Naphthalene-d8	10.722	136	14777	0.400	ng	# 0.00
13) Acenaphthene-d10	14.549	164	7208	0.400	ng	0.00
19) Phenanthrene-d10	17.287	188	13343	0.400	ng	# 0.00
29) Chrysene-d12	21.467	240	12034	0.400	ng	# 0.00
35) Perylene-d12	23.861	264	9620	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.471	112	50439	2.895	ng	0.00
5) Phenol-d6	7.067	99	59269	3.103	ng	0.00
8) Nitrobenzene-d5	9.067	82	39309	3.849	ng	0.00
11) 2-Methylnaphthalene-d10	12.310	152	71840	3.012	ng	0.00
14) 2,4,6-Tribromophenol	16.025	330	11434	4.389	ng	0.00
15) 2-Fluorobiphenyl	13.180	172	105595	3.694	ng	0.00
27) Fluoranthene-d10	19.316	212	128513	3.273	ng	0.00
31) Terphenyl-d14	19.906	244	80333	2.627	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.326	88	24754	5.071	ng	# 57
3) n-Nitrosodimethylamine	3.637	42	26633	3.750	ng	89
6) bis(2-Chloroethyl)ether	7.327	93	57734	2.853	ng	99
9) Naphthalene	10.776	128	135914	3.513	ng	100
10) Hexachlorobutadiene	11.075	225	32177	4.074	ng	# 100
12) 2-Methylnaphthalene	12.382	142	89298	3.434	ng	98
16) Acenaphthylene	14.271	152	118671	3.912	ng	100
17) Acenaphthene	14.613	154	77789	3.836	ng	97
18) Fluorene	15.586	166	95098	3.823	ng	99
20) 4,6-Dinitro-2-methylph...	15.660	198	9299	9.631	ng	# 76
21) 4-Bromophenyl-phenylether	16.477	248	31313	3.869	ng	93
22) Hexachlorobenzene	16.600	284	37609	4.156	ng	# 88
23) Atrazine	16.733	200	18857	3.837	ng	# 91
24) Pentachlorophenol	16.938	266	11817	4.478	ng	99
25) Phenanthrene	17.328	178	148176	3.797	ng	100
26) Anthracene	17.420	178	132322	3.993	ng	100
28) Fluoranthene	19.341	202	166689	3.876	ng	# 97
30) Pyrene	19.704	202	165719	4.020	ng	100
32) Benzo(a)anthracene	21.449	228	152086	3.803	ng	99
33) Chrysene	21.503	228	159887	3.945	ng	99
34) Bis(2-ethylhexyl)phtha...	21.378	149	60931	3.762	ng	# 95
36) Indeno(1,2,3-cd)pyrene	26.346	276	168785	4.879	ng	# 90
37) Benzo(b)fluoranthene	23.127	252	143788	4.341	ng	# 96
38) Benzo(k)fluoranthene	23.176	252	145862	4.407	ng	# 97
39) Benzo(a)pyrene	23.752	252	117373	4.429	ng	# 96
40) Dibenzo(a,h)anthracene	26.360	278	132860	4.926	ng	# 94
41) Benzo(g,h,i)perylene	27.109	276	142165	4.722	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018677.D
 Acq On : 22 Feb 2022 15:01
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Quant Time: Feb 22 15:52:47 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 13:54:19 2022
 Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.919 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN018677.D
 Acq: 22 Feb 2022 15:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2






Tgt Ion:152 Resp: 6859
 Ion Ratio Lower Upper
 152 100
 150 154.5 121.7 182.5
 115 60.0 43.8 65.8



#2
 1,4-Dioxane
 Concen: 5.071 ng
 RT: 3.326 min Scan# 34
 Delta R.T. -0.000 min
 Lab File: BN018677.D
 Acq: 22 Feb 2022 15:01

Tgt Ion: 88 Resp: 24754
 Ion Ratio Lower Upper
 88 100
 43 30.4 69.7 104.5#
 58 77.1 83.8 125.8#









#3
n-Nitrosodimethylamine
Concen: 3.750 ng
RT: 3.637 min Scan# 71
Delta R.T. -0.007 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 26633
Ion Ratio Lower Upper
42 100
74 122.4 108.6 163.0
44 6.8 5.8 8.8



#4
2-Fluorophenol
Concen: 2.895 ng
RT: 5.471 min Scan# 331
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion: 112 Resp: 50439
Ion Ratio Lower Upper
112 100
64 59.2 48.0 72.0
63 31.3 25.3 37.9









#5
Phenol-d6
Concen: 3.103 ng
RT: 7.067 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion: 99 Resp: 59269
Ion Ratio Lower Upper
99 100
42 20.3 16.3 24.5
71 34.5 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 2.853 ng
RT: 7.327 min Scan# 588
Delta R.T. -0.007 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion: 93 Resp: 57734
Ion Ratio Lower Upper
93 100
63 77.8 61.7 92.5
95 32.6 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.722 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:136 Resp: 14777
Ion Ratio Lower Upper
136 100
137 11.5 8.8 13.2
54 9.2 5.0 7.6#
68 6.9 4.4 6.6#



#8
Nitrobenzene-d5
Concen: 3.849 ng
RT: 9.067 min Scan# 795
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion: 82 Resp: 39309
Ion Ratio Lower Upper
82 100
128 37.4 33.7 50.5
54 52.2 36.6 55.0









#9
Naphthalene
Concen: 3.513 ng
RT: 10.776 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 135914
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.9 10.4 15.6



#10
Hexachlorobutadiene
Concen: 4.074 ng
RT: 11.075 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:225 Resp: 32177
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 3.012 ng
RT: 12.310 min Scan# 1235
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 71840
Ion Ratio Lower Upper
152 100
151 20.7 16.2 24.2



#12
2-Methylnaphthalene
Concen: 3.434 ng
RT: 12.382 min Scan# 1254
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:142 Resp: 89298
Ion Ratio Lower Upper
142 100
141 88.7 70.0 105.0
115 37.7 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.549 min Scan# 1478
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 7208
Ion Ratio Lower Upper
164 100
162 102.2 80.2 120.2
160 48.0 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 4.389 ng
RT: 16.025 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:330 Resp: 11434
Ion Ratio Lower Upper
330 100
332 97.8 0.0 0.0#
141 39.5 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 3.694 ng
RT: 13.180 min Scan# 1350
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 105595
Ion Ratio Lower Upper
172 100
171 34.1 27.4 41.0
170 22.3 18.4 27.6



#16
Acenaphthylene
Concen: 3.912 ng
RT: 14.271 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:152 Resp: 118671
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.2 10.5 15.7









#17
Acenaphthene
Concen: 3.836 ng
RT: 14.613 min Scan# 1484
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 77789
Ion Ratio Lower Upper
154 100
153 109.6 90.9 136.3
152 54.1 45.0 67.4



#18
Fluorene
Concen: 3.823 ng
RT: 15.586 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:166 Resp: 95098
Ion Ratio Lower Upper
166 100
165 98.5 77.8 116.6
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.287 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 13343
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.5 7.0 10.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 9.631 ng
RT: 15.660 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:198 Resp: 9299
Ion Ratio Lower Upper
198 100
182 8.4 15.5 23.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.869 ng
RT: 16.477 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 31313
Ion Ratio Lower Upper
248 100
250 97.5 74.4 111.6
141 63.3 43.7 65.5



#22
Hexachlorobenzene
Concen: 4.156 ng
RT: 16.600 min Scan# 1673
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:284 Resp: 37609
Ion Ratio Lower Upper
284 100
142 36.7 21.6 32.4#
249 32.5 23.5 35.3









#23
Atrazine
Concen: 3.837 ng
RT: 16.733 min Scan# 1
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 18857
Ion Ratio Lower Upper
200 100
173 30.6 20.3 30.5#
215 45.0 40.4 60.6



#24
Pentachlorophenol
Concen: 4.478 ng
RT: 16.938 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:266 Resp: 11817
Ion Ratio Lower Upper
266 100
264 62.9 49.5 74.3
268 64.1 50.6 76.0









#25
Phenanthrene
Concen: 3.797 ng
RT: 17.328 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 148176
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.2 12.3 18.5



#26
Anthracene
Concen: 3.993 ng
RT: 17.420 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:178 Resp: 132322
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.273 ng
RT: 19.316 min Scan# 2120
Delta R.T. 0.004 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 128513
Ion Ratio Lower Upper
212 100
106 15.3 8.4 12.6#
104 8.6 4.5 6.7#



#28
Fluoranthene
Concen: 3.876 ng
RT: 19.341 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:202 Resp: 166689
Ion Ratio Lower Upper
202 100
101 11.7 7.2 10.8#
203 17.3 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.467 min Scan# 2448
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 12034
Ion Ratio Lower Upper
240 100
120 9.9 6.0 9.0#
236 29.0 21.4 32.2



#30
Pyrene
Concen: 4.020 ng
RT: 19.704 min Scan# 2213
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:202 Resp: 165719
Ion Ratio Lower Upper
202 100
200 21.0 16.6 25.0
203 17.7 14.1 21.1









#31
 Terphenyl-d14
 Concen: 2.627 ng
 RT: 19.906 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN018677.D
 Acq: 22 Feb 2022 15:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 80333
 Ion Ratio Lower Upper
 244 100
 212 10.0 6.2 9.2#
 122 0.0 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 3.803 ng
 RT: 21.449 min Scan# 2446
 Delta R.T. -0.000 min
 Lab File: BN018677.D
 Acq: 22 Feb 2022 15:01

Tgt Ion:228 Resp: 152086
 Ion Ratio Lower Upper
 228 100
 226 27.1 21.3 31.9
 229 19.8 15.9 23.9





#33

Chrysene

Concen: 3.945 ng

RT: 21.503 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BN018677.D

Acq: 22 Feb 2022 15:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 159887

Ion Ratio Lower Upper

228 100

226 29.6 24.2 36.4

229 19.7 15.5 23.3



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.762 ng

RT: 21.378 min Scan# 2438

Delta R.T. -0.000 min

Lab File: BN018677.D

Acq: 22 Feb 2022 15:01




Tgt Ion:149 Resp: 60931


Ion Ratio Lower Upper

149 100

167 27.3 23.7 35.5

279 2.7 3.6 5.4#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.861 min Scan# 3050
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:264 Resp: 9620
Ion Ratio Lower Upper
264 100
260 28.3 20.2 30.4
265 33.6 18.2 27.2#



#36
Indeno(1,2,3-cd)pyrene
Concen: 4.879 ng
RT: 26.346 min Scan# 3900
Delta R.T. -0.006 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Tgt Ion:276 Resp: 168785
Ion Ratio Lower Upper
276 100
138 28.8 14.9 22.3#
277 24.8 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 4.341 ng
RT: 23.127 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 143788
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 11.3 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 4.407 ng
RT: 23.176 min Scan# 2816
Delta R.T. -0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01


Tgt Ion:252 Resp: 145862
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 11.0 6.2 9.2#








#39
Benzo(a)pyrene
Concen: 4.429 ng
RT: 23.752 min Scan# 3013
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2





Tgt Ion:252 Resp: 117373
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 12.4 7.0 10.4#




#40
Dibenzo(a,h)anthracene
Concen: 4.926 ng
RT: 26.360 min Scan# 3905
Delta R.T. -0.003 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01




Tgt Ion:278 Resp: 132860
Ion Ratio Lower Upper
278 100
139 20.0 10.8 16.2#
279 23.7 19.0 28.4





#41
Benzo(g,h,i)perylene
Concen: 4.722 ng
RT: 27.109 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN018677.D
Acq: 22 Feb 2022 15:01

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2



Tgt Ion:276 Resp: 142165
Ion Ratio Lower Upper
276 100
277 23.6 19.0 28.6
138 24.9 12.6 19.0#

