

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018687.D
 Acq On : 22 Feb 2022 21:22
 Operator : CG/JU
 Sample : N1481-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW7B-GW-20220209


Quant Time: Feb 23 02:17:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	4	0.400	ng	# 0.01
7) Naphthalene-d8	10.786	136	3	0.400	ng	# 0.06
13) Acenaphthene-d10	14.634	164	5	0.400	ng	# 0.09
19) Phenanthrene-d10	17.359	188	3	0.400	ng	# 0.07
29) Chrysene-d12	21.485	240	59	0.400	ng	# 0.02
35) Perylene-d12	23.878	264	2	0.400	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.485	112	1	0.109	ng	0.01
5) Phenol-d6	7.103	99	1	0.095	ng	0.04
8) Nitrobenzene-d5	9.121	82	6	2.586	ng	0.05
11) 2-Methylnaphthalene-d10	12.374	152	2	0.442	ng	0.06
14) 2,4,6-Tribromophenol	16.148	330	1	0.512	ng	0.12
15) 2-Fluorobiphenyl	13.234	172	3	0.127	ng	0.05
27) Fluoranthene-d10	19.375	212	1	0.123	ng	0.06
31) Terphenyl-d14	19.919	244	21	0.164	ng	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.340	88	6	1.292	ng	# 39
3) n-Nitrosodimethylamine	3.629	42	6	1.302	ng	# 1
6) bis(2-Chloroethyl)ether	7.349	93	10	0.924	ng	# 20
9) Naphthalene	10.840	128	3	0.347	ng	# 1
10) Hexachlorobutadiene	11.075	225	4	1.921	ng	# 86
16) Acenaphthylene	14.335	152	1	0.042	ng	# 1
17) Acenaphthene	14.655	154	2	0.122	ng	# 21
18) Fluorene	15.714	166	4	0.203	ng	# 10
21) 4-Bromophenyl-phenylether	16.487	248	7	3.283	ng	# 65
22) Hexachlorobenzene	16.672	284	9	3.372	ng	# 1
23) Atrazine	16.784	200	1	0.850	ng	# 7
24) Pentachlorophenol	17.000	266	6	7.241	ng	# 15
25) Phenanthrene	17.431	178	13	1.295	ng	# 60
26) Anthracene	17.482	178	20	2.352	ng	# 1
28) Fluoranthene	19.421	202	3	0.286	ng	# 1
30) Pyrene	19.712	202	74	0.281	ng	# 89
32) Benzo(a)anthracene	21.467	228	52	0.232	ng	# 1
33) Chrysene	21.512	228	109	0.437	ng	# 12
34) Bis(2-ethylhexyl)phtha...	21.377	149	83	0.913	ng	# 75
36) Indeno(1,2,3-cd)pyrene	26.357	276	2	0.189	ng	# 1
37) Benzo(b)fluoranthene	23.138	252	43	4.732	ng	# 1
38) Benzo(k)fluoranthene	23.171	252	4	0.444	ng	# 1
39) Benzo(a)pyrene	23.787	252	4	0.552	ng	# 1
40) Dibenzo(a,h)anthracene	26.369	278	4	0.486	ng	# 1
41) Benzo(g,h,i)perylene	27.115	276	10	1.107	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Feb 23 02:17:52 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 22 16:23:51 2022
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.934 min Scan# 61
 Delta R.T. 0.015 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22

Instrument :
 BNA_N
 ClientSampleId :
 RW7B-GW-20220209






Tgt Ion:152 Resp: 4
 Ion Ratio Lower Upper
 152 100
 150 98.0 121.7 182.5#
 115 154.9 43.8 65.8#



#2
 1,4-Dioxane
 Concen: 1.292 ng
 RT: 3.340 min Scan# 36
 Delta R.T. 0.014 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22

Tgt Ion: 88 Resp: 6
 Ion Ratio Lower Upper
 88 100
 43 133.3 69.7 104.5#
 58 33.3 83.8 125.8#









#3
 n-Nitrosodimethylamine
 Concen: 1.302 ng
 RT: 3.629 min Scan# 70
 Delta R.T. -0.014 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22




Instrument :
 BNA_N
 ClientSampleId :
 RW7B-GW-20220209


Tgt Ion: 42 Resp: 6
 Ion Ratio Lower Upper
 42 100
 74 166.7 108.6 163.0#
 44 716.7 5.8 8.8#



#4
 2-Fluorophenol
 Concen: 0.109 ng
 RT: 5.485 min Scan# 333
 Delta R.T. 0.014 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22

Tgt Ion: 112 Resp: 1
 Ion Ratio Lower Upper
 112 100
 64 0.0 48.0 72.0#
 63 600.0 25.3 37.9#









#5
Phenol-d6
Concen: 0.095 ng
RT: 7.103 min Scan# 51
Delta R.T. 0.036 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209

Tgt Ion: 99 Resp: 1
Ion Ratio Lower Upper
99 100
42 0.0 16.3 24.5#
71 100.0 26.8 40.2#



#6
bis(2-Chloroethyl)ether
Concen: 0.924 ng
RT: 7.349 min Scan# 591
Delta R.T. 0.014 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion: 93 Resp: 10
Ion Ratio Lower Upper
93 100
63 0.0 61.7 92.5#
95 0.0 26.2 39.4#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.786 min Scan# 91
Delta R.T. 0.064 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:136 Resp: 3
Ion Ratio Lower Upper
136 100
137 98.0 8.8 13.2#
54 112.2 5.0 7.6#
68 114.3 4.4 6.6#



#8
Nitrobenzene-d5
Concen: 2.586 ng
RT: 9.121 min Scan# 800
Delta R.T. 0.053 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion: 82 Resp: 6
Ion Ratio Lower Upper
82 100
128 90.2 33.7 50.5#
54 93.4 36.6 55.0#









#9
Naphthalene
Concen: 0.347 ng
RT: 10.840 min Scan# 90
Delta R.T. 0.064 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:128 Resp: 3
Ion Ratio Lower Upper
128 100
129 87.7 8.7 13.1#
127 84.2 10.4 15.6#



#10
Hexachlorobutadiene
Concen: 1.921 ng
RT: 11.075 min Scan# 983
Delta R.T. -0.000 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:225 Resp: 4
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 75.0 51.0 76.6





#11

2-Methylnaphthalene-d10

Concen: 0.442 ng

RT: 12.374 min Scan# 11

Delta R.T. 0.064 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22

Instrument :

BNA_N

ClientSampleId :





RW7B-GW-20220209

Tgt Ion:152 Resp: 2

Ion Ratio Lower Upper

152 100

151 0.0 16.2 24.2#



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.634 min Scan# 1486

Delta R.T. 0.085 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22




Tgt Ion:164 Resp: 5


Ion Ratio Lower Upper

164 100

162 98.0 80.2 120.2


160 90.2 36.9 55.3#








#14
2,4,6-Tribromophenol
Concen: 0.512 ng
RT: 16.148 min Scan# 1617
Delta R.T. 0.123 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209







Tgt Ion:330 Resp: 1
Ion Ratio Lower Upper
330 100
332 200.0 0.0 0.0#
141 400.0 21.3 31.9#



#15
2-Fluorobiphenyl
Concen: 0.127 ng
RT: 13.234 min Scan# 1355
Delta R.T. 0.053 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:172 Resp: 3
Ion Ratio Lower Upper
172 100
171 97.9 27.4 41.0#
170 95.7 18.4 27.6#









#16
Acenaphthylene
Concen: 0.042 ng
RT: 14.335 min Scan# 1452
Delta R.T. 0.064 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:152 Resp: 1
Ion Ratio Lower Upper
152 100
151 0.0 15.8 23.6#
153 400.0 10.5 15.7#



#17
Acenaphthene
Concen: 0.122 ng
RT: 14.655 min Scan# 1488
Delta R.T. 0.043 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:154 Resp: 2
Ion Ratio Lower Upper
154 100
153 200.0 90.9 136.3#
152 0.0 45.0 67.4#









#18
Fluorene
Concen: 0.203 ng
RT: 15.714 min Scan# 1587
Delta R.T. 0.128 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:166 Resp: 4
Ion Ratio Lower Upper
166 100
165 0.0 77.8 116.6#
167 25.0 10.9 16.3#



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.359 min Scan# 1747
Delta R.T. 0.072 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:188 Resp: 3
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 108.3 7.0 10.4#









#21
4-Bromophenyl-phenylether
Concen: 3.283 ng
RT: 16.487 min Scan# 1662
Delta R.T. 0.010 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:248 Resp: 7
Ion Ratio Lower Upper
248 100
250 104.3 74.4 111.6
141 108.5 43.7 65.5#



#22
Hexachlorobenzene
Concen: 3.372 ng
RT: 16.672 min Scan# 1680
Delta R.T. 0.072 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:284 Resp: 9
Ion Ratio Lower Upper
284 100
142 44.4 21.6 32.4#
249 122.2 23.5 35.3#









#23
 Atrazine
 Concen: 0.850 ng
 RT: 16.784 min Scan# 1
 Delta R.T. 0.051 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22




Instrument :
 BNA_N
 ClientSampleId :
 RW7B-GW-20220209


Tgt Ion:200 Resp: 1
 Ion Ratio Lower Upper
 200 100
 173 97.8 20.3 30.5#
 215 97.8 40.4 60.6#



#24
 Pentachlorophenol
 Concen: 7.241 ng
 RT: 17.000 min Scan# 1712
 Delta R.T. 0.062 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22

Tgt Ion:266 Resp: 6
 Ion Ratio Lower Upper
 266 100
 264 16.7 49.5 74.3#
 268 150.0 50.6 76.0#









#25
Phenanthrene
Concen: 1.295 ng
RT: 17.431 min Scan# 1754
Delta R.T. 0.103 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:178 Resp: 13
Ion Ratio Lower Upper
178 100
176 0.0 15.3 22.9#
179 0.0 12.3 18.5#



#26
Anthracene
Concen: 2.352 ng
RT: 17.482 min Scan# 1759
Delta R.T. 0.062 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:178 Resp: 20
Ion Ratio Lower Upper
178 100
176 90.0 14.8 22.2#
179 0.0 12.2 18.4#





#27

Fluoranthene-d10

Concen: 0.123 ng

RT: 19.375 min Scan# 21

Delta R.T. 0.063 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22

Instrument :

BNA_N

ClientSampleId :

RW7B-GW-20220209





Tgt Ion:212 Resp: 1

Ion Ratio Lower Upper

212 100

106 300.0 8.4 12.6#

104 0.0 4.5 6.7#



#28

Fluoranthene

Concen: 0.286 ng

RT: 19.421 min Scan# 2146

Delta R.T. 0.080 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22




Tgt Ion:202 Resp: 3


Ion Ratio Lower Upper

202 100

101 66.7 7.2 10.8#

203 200.0 13.7 20.5#









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.485 min Scan# 2448
Delta R.T. 0.018 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22




Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209


Tgt Ion:240 Resp: 59
Ion Ratio Lower Upper
240 100
120 203.1 6.0 9.0#
236 101.5 21.4 32.2#



#30
Pyrene
Concen: 0.281 ng
RT: 19.712 min Scan# 2215
Delta R.T. 0.008 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:202 Resp: 74
Ion Ratio Lower Upper
202 100
200 21.6 16.6 25.0
203 27.0 14.1 21.1#









#31
 Terphenyl-d14
 Concen: 0.164 ng
 RT: 19.919 min Scan# 21
 Delta R.T. 0.013 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22




Instrument :
 BNA_N
 ClientSampleId :
 RW7B-GW-20220209


Tgt Ion:244 Resp: 21
 Ion Ratio Lower Upper
 244 100
 212 94.6 6.2 9.2#
 122 0.0 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 0.232 ng
 RT: 21.467 min Scan# 2448
 Delta R.T. 0.018 min
 Lab File: BN018687.D
 Acq: 22 Feb 2022 21:22

Tgt Ion:228 Resp: 52
 Ion Ratio Lower Upper
 228 100
 226 76.1 21.3 31.9#
 229 77.5 15.9 23.9#





#33

Chrysene

Concen: 0.437 ng

RT: 21.512 min Scan# 2453

Delta R.T. 0.009 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22

Instrument :

BNA_N

ClientSampleId :

RW7B-GW-20220209





Tgt Ion:228 Resp: 109

Ion Ratio Lower Upper

228 100

226 72.9 24.2 36.4#

229 65.9 15.5 23.3#



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.913 ng

RT: 21.377 min Scan# 2438

Delta R.T. -0.000 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22




Tgt Ion:149 Resp: 83


Ion Ratio Lower Upper

149 100

167 15.7 23.7 35.5#

279 9.6 3.6 5.4#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.878 min Scan# 3050

Delta R.T. 0.018 min

Lab File: BN018687.D


Acq: 22 Feb 2022 21:22

Instrument :

BNA_N

ClientSampleId :

RW7B-GW-20220209






Tgt Ion:264 Resp: 2

Ion Ratio Lower Upper

264 100

260 83.1 20.2 30.4#

265 950.8 18.2 27.2#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.189 ng

RT: 26.357 min Scan# 3904

Delta R.T. 0.006 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22




Tgt Ion:276 Resp: 2


Ion Ratio Lower Upper

276 100

138 250.0 14.9 22.3#


277 100.0 19.9 29.9#







#37
Benzo(b)fluoranthene
Concen: 4.732 ng
RT: 23.138 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Instrument :
BNA_N
ClientSampleId :
RW7B-GW-20220209







Tgt Ion:252 Resp: 43
Ion Ratio Lower Upper
252 100
253 143.6 17.8 26.6#
125 97.6 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 0.444 ng
RT: 23.171 min Scan# 2814
Delta R.T. -0.006 min
Lab File: BN018687.D
Acq: 22 Feb 2022 21:22

Tgt Ion:252 Resp: 4
Ion Ratio Lower Upper
252 100
253 143.3 17.7 26.5#
125 103.8 6.2 9.2#





#39

Benzo(a)pyrene

Concen: 0.552 ng

RT: 23.787 min Scan# 3013

Delta R.T. 0.035 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22

Instrument :

BNA_N

ClientSampleId :

RW7B-GW-20220209





Tgt Ion:252 Resp: 4

Ion Ratio Lower Upper

252 100

253 167.2 17.8 26.6#

125 118.2 7.0 10.4#



#40

Dibenzo(a,h)anthracene

Concen: 0.486 ng

RT: 26.369 min Scan# 3908

Delta R.T. 0.006 min

Lab File: BN018687.D

Acq: 22 Feb 2022 21:22




Tgt Ion:278 Resp: 4


Ion Ratio Lower Upper

278 100

139 111.9 10.8 16.2#

279 176.3 19.0 28.4#





#41

Benzo(g,h,i)perylene

Concen: 1.107 ng

RT: 27.115 min Scan# 41

Delta R.T. 0.006 min

Lab File: BN018687.D


Acq: 22 Feb 2022 21:22

Instrument :

BNA_N

ClientSampleId :

RW7B-GW-20220209



Tgt Ion:276 Resp: 10

Ion Ratio Lower Upper

276 100

277 85.1 19.0 28.6#

138 100.0 12.6 19.0#

