

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018679.D
 Acq On : 22 Feb 2022 16:14
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN022222

Quant Time: Feb 22 16:56:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.464	0.462	0.4	94	0.00
3	n-Nitrosodimethylamine	0.461	0.435	5.6	92	0.00
4 S	2-Fluorophenol	0.916	0.883	3.6	96	0.00
5 S	Phenol-d6	1.053	0.981	6.8	95	0.00
6	bis(2-Chloroethyl)ether	1.082	1.028	5.0	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.309	0.278	10.0	97	0.00
9	Naphthalene	1.154	1.104	4.3	96	0.00
10	Hexachlorobutadiene	0.278	0.267	4.0	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.603	0.570	5.5	96	0.00
12	2-Methylnaphthalene	0.773	0.737	4.7	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.156	0.134	14.1	91	0.01
15 S	2-Fluorobiphenyl	1.896	1.827	3.6	96	0.00
16	Acenaphthylene	1.898	1.686	11.2	96	0.00
17	Acenaphthene	1.310	1.201	8.3	94	0.00
18	Fluorene	1.577	1.448	8.2	93	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.059	0.049	16.9	89	0.00
21	4-Bromophenyl-phenylether	0.284	0.267	6.0	96	0.00
22	Hexachlorobenzene	0.356	0.346	2.8	96	0.00
23	Atrazine	0.157	0.139	11.5	96	0.00
24	Pentachlorophenol	0.075	0.058	22.7	81	0.00
25	Phenanthrene	1.339	1.267	5.4	96	0.00
26	Anthracene	1.134	1.027	9.4	94	0.00
27 SURR	Fluoranthene-d10	1.086	1.021	6.0	95	0.00
28	Fluoranthene	1.399	1.294	7.5	95	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
30	Pyrene	1.786	1.704	4.6	96	0.00
31 S	Terphenyl-d14	0.868	0.846	2.5	97	0.00
32	Benzo(a)anthracene	1.517	1.407	7.3	96	0.00
33	Chrysene	1.693	1.630	3.7	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.616	0.573	7.0	97	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	2.112	1.947	7.8	97	0.00
37	Benzo(b)fluoranthene	1.817	1.655	8.9	94	0.00
38	Benzo(k)fluoranthene	1.802	1.675	7.0	97	0.00
39 C	Benzo(a)pyrene	1.449	1.326	8.5	96	0.00
40	Dibenzo(a,h)anthracene	1.647	1.510	8.3	97	0.00
41	Benzo(g,h,i)perylene	1.807	1.670	7.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN022222