

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018679.D
 Acq On : 22 Feb 2022 16:14
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN022222

Quant Time: Feb 22 16:56:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.398	0.5	94	0.00
3	n-Nitrosodimethylamine	0.400	0.378	5.5	92	0.00
4 S	2-Fluorophenol	0.400	0.386	3.5	96	0.00
5 S	Phenol-d6	0.400	0.372	7.0	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.380	5.0	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.360	10.0	97	0.00
9	Naphthalene	0.400	0.382	4.5	96	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	96	0.00
12	2-Methylnaphthalene	0.400	0.382	4.5	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.344	14.0	91	0.01
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	96	0.00
16	Acenaphthylene	0.400	0.355	11.3	96	0.00
17	Acenaphthene	0.400	0.367	8.3	94	0.00
18	Fluorene	0.400	0.367	8.3	93	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.371	7.3	89	0.00
21	4-Bromophenyl-phenylether	0.400	0.376	6.0	96	0.00
22	Hexachlorobenzene	0.400	0.389	2.8	96	0.00
23	Atrazine	0.400	0.353	11.8	96	0.00
24	Pentachlorophenol	0.400	0.361	9.8	81	0.00
25	Phenanthrene	0.400	0.378	5.5	96	0.00
26	Anthracene	0.400	0.362	9.5	94	0.00
27 SURR	Fluoranthene-d10	0.400	0.376	6.0	95	0.00
28	Fluoranthene	0.400	0.370	7.5	95	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.382	4.5	96	0.00
31 S	Terphenyl-d14	0.400	0.390	2.5	97	0.00
32	Benzo(a)anthracene	0.400	0.371	7.3	96	0.00
33	Chrysene	0.400	0.385	3.8	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.372	7.0	97	0.00
35 I	Perylene-d12	0.400	0.400	0.0	96	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	97	0.00
37	Benzo(b)fluoranthene	0.400	0.364	9.0	94	0.00
38	Benzo(k)fluoranthene	0.400	0.372	7.0	97	0.00
39 C	Benzo(a)pyrene	0.400	0.366	8.5	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.367	8.3	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.370	7.5	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN022222