

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018688.D
 Acq On : 22 Feb 2022 21:59
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 23 02:17:58 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	0	0.00
2	1,4-Dioxane	0.400	0.574	-43.5#	0	0.00
3	n-Nitrosodimethylamine	0.400	2.025	-406.3#	0	0.00
4 S	2-Fluorophenol	0.400	0.437	-9.2	0	0.00
5 S	Phenol-d6	0.400	0.380	5.0	0	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.493	-23.2	0	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	0	-0.01
8 S	Nitrobenzene-d5	0.400	5.173	-1193.3#	0	0.00
9	Naphthalene	0.400	1.386	-246.5#	0	0.03
10	Hexachlorobutadiene	0.400	8.644	-2061.0#	0	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.663	-65.8#	0	0.00
12	2-Methylnaphthalene	0.400	0.000	100.0#	0	-12.38#
13 I	Acenaphthene-d10	0.400	0.400	0.0	0	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.285	28.8#	0	-0.05
15 S	2-Fluorobiphenyl	0.400	0.141	64.8#	0	0.00
16	Acenaphthylene	0.400	0.281	29.8#	0	0.01
17	Acenaphthene	0.400	0.102	74.5#	0	0.01
18	Fluorene	0.400	0.338	15.5	0	0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	0	-0.06
20	4,6-Dinitro-2-methylphenol	0.400	22.585	-5546.3#	1	-0.07
21	4-Bromophenyl-phenylether	0.400	1.407	-251.8#	0	-0.09
22	Hexachlorobenzene	0.400	2.248	-462.0#	0	-0.04
23	Atrazine	0.400	5.098	-1174.5#	0	-0.10
24	Pentachlorophenol	0.400	42.671	-10567.8#	1	-0.06
25	Phenanthrene	0.400	1.494	-273.5#	0	-0.06
26	Anthracene	0.400	13.407	-3251.8#	0	-0.08
27 SURR	Fluoranthene-d10	0.400	0.737	-84.2#	0	-0.07
28	Fluoranthene	0.400	0.858	-114.5#	0	-0.11
29 I	Chrysene-d12	0.400	0.400	0.0	0	-0.10
30	Pyrene	0.400	0.224	44.0#	0	-0.09
31 S	Terphenyl-d14	0.400	0.154	61.5#	0	-0.08
32	Benzo(a)anthracene	0.400	0.527	-31.8#	0	-0.09
33	Chrysene	0.400	0.630	-57.5#	0	-0.08
34	Bis(2-ethylhexyl)phthalate	0.400	16.656	-4064.0#	1	-0.11
35 I	Perylene-d12	0.400	0.400	0.0	0	-0.11
36	Indeno(1,2,3-cd)pyrene	0.400	0.095	76.3#	0	-0.12
37	Benzo(b)fluoranthene	0.400	0.550	-37.5#	0	-0.11
38	Benzo(k)fluoranthene	0.400	0.555	-38.8#	0	-0.11
39 C	Benzo(a)pyrene	0.400	0.690	-72.5#	0	-0.11
40	Dibenzo(a,h)anthracene	0.400	0.243	39.3#	0	-0.13
41	Benzo(g,h,i)perylene	0.400	0.221	44.8#	0	-0.14

(#) = Out of Range

SPCC's out = 0 CCC's out = 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4