

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
 Data File : BN018674.D
 Acq On : 22 Feb 2022 13:12
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Feb 22 13:55:42 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 13:54:19 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.919	152	6418	0.400	ng	0.00
7) Naphthalene-d8	10.722	136	15226	0.400	ng	# 0.00
13) Acenaphthene-d10	14.549	164	7252	0.400	ng	0.00
19) Phenanthrene-d10	17.287	188	13009	0.400	ng	0.00
29) Chrysene-d12	21.467	240	9790	0.400	ng	# 0.00
35) Perylene-d12	23.861	264	8770	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.471	112	5706	0.350	ng	0.00
5) Phenol-d6	7.067	99	6355	0.356	ng	0.00
8) Nitrobenzene-d5	9.067	82	4196	0.399	ng	0.00
11) 2-Methylnaphthalene-d10	12.310	152	8647	0.352	ng	0.00
14) 2,4,6-Tribromophenol	16.025	330	1019	0.389	ng	0.00
15) 2-Fluorobiphenyl	13.180	172	13232	0.460	ng	0.00
27) Fluoranthene-d10	19.312	212	13165	0.344	ng	0.00
31) Terphenyl-d14	19.906	244	8213	0.330	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.326	88	3029	0.663	ng	# 58
3) n-Nitrosodimethylamine	3.644	42	2920	0.439	ng	89
6) bis(2-Chloroethyl)ether	7.334	93	6713	0.354	ng	98
9) Naphthalene	10.776	128	16771	0.421	ng	99
10) Hexachlorobutadiene	11.075	225	4074	0.501	ng	# 100
12) 2-Methylnaphthalene	12.382	142	10265	0.383	ng	98
16) Acenaphthylene	14.271	152	12131	0.397	ng	99
17) Acenaphthene	14.613	154	8867	0.435	ng	99
18) Fluorene	15.586	166	10712	0.428	ng	99
20) 4,6-Dinitro-2-methylph...	15.660	198	674	0.716	ng	98
21) 4-Bromophenyl-phenylether	16.477	248	3407	0.432	ng	92
22) Hexachlorobenzene	16.600	284	4431	0.502	ng	# 88
23) Atrazine	16.733	200	1780	0.371	ng	# 89
24) Pentachlorophenol	16.938	266	884	0.344	ng	99
25) Phenanthrene	17.328	178	16284	0.428	ng	99
26) Anthracene	17.420	178	13363	0.414	ng	99
28) Fluoranthene	19.341	202	16742	0.399	ng	# 97
30) Pyrene	19.704	202	16677	0.497	ng	100
32) Benzo(a)anthracene	21.449	228	13814	0.425	ng	99
33) Chrysene	21.503	228	15864	0.481	ng	99
34) Bis(2-ethylhexyl)phtha...	21.378	149	5571	0.423	ng	# 95
36) Indeno(1,2,3-cd)pyrene	26.352	276	16959	0.538	ng	# 90
37) Benzo(b)fluoranthene	23.130	252	14790	0.490	ng	# 92
38) Benzo(k)fluoranthene	23.176	252	14524	0.481	ng	# 92
39) Benzo(a)pyrene	23.752	252	11620	0.481	ng	# 88
40) Dibenzo(a,h)anthracene	26.363	278	13150	0.535	ng	# 90
41) Benzo(g,h,i)perylene	27.109	276	14547	0.530	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022222\
Data File : BN018674.D
Acq On : 22 Feb 2022 13:12
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Feb 22 13:55:42 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 22 13:54:19 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.919 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 6418
Ion Ratio Lower Upper
152 100
150 155.0 121.7 182.5
115 60.0 43.8 65.8



#2
1,4-Dioxane
Concen: 0.663 ng
RT: 3.326 min Scan# 34
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion: 88 Resp: 3029
Ion Ratio Lower Upper
88 100
43 30.8 69.7 104.5#
58 76.9 83.8 125.8#









#3
 n-Nitrosodimethylamine
 Concen: 0.439 ng
 RT: 3.644 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BN018674.D
 Acq: 22 Feb 2022 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion: 42 Resp: 2920
 Ion Ratio Lower Upper
 42 100
 74 121.8 108.6 163.0
 44 8.6 5.8 8.8



#4
 2-Fluorophenol
 Concen: 0.350 ng
 RT: 5.471 min Scan# 331
 Delta R.T. 0.000 min
 Lab File: BN018674.D
 Acq: 22 Feb 2022 13:12

Tgt Ion: 112 Resp: 5706
 Ion Ratio Lower Upper
 112 100
 64 59.7 48.0 72.0
 63 31.7 25.3 37.9









#5
Phenol-d6
Concen: 0.356 ng
RT: 7.067 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 6355
Ion Ratio Lower Upper
99 100
42 20.9 16.3 24.5
71 34.3 26.8 40.2



#6
bis(2-Chloroethyl)ether
Concen: 0.354 ng
RT: 7.334 min Scan# 589
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion: 93 Resp: 6713
Ion Ratio Lower Upper
93 100
63 78.5 61.7 92.5
95 33.6 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.722 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	15226
Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.8	13.2
54	8.9	5.0	7.6#
68	6.9	4.4	6.6#



#8
Nitrobenzene-d5
Concen: 0.399 ng
RT: 9.067 min Scan# 795
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:	82	Resp:	4196
Ion	Ratio	Lower	Upper
82	100		
128	36.9	33.7	50.5
54	55.2	36.6	55.0#









#9
Naphthalene
Concen: 0.421 ng
RT: 10.776 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 16771
Ion Ratio Lower Upper
128 100
129 11.5 8.7 13.1
127 13.4 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.501 ng
RT: 11.075 min Scan# 983
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:225 Resp: 4074
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.352 ng
RT: 12.310 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 8647
Ion Ratio Lower Upper
152 100
151 21.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.383 ng
RT: 12.382 min Scan# 1254
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:142 Resp: 10265
Ion Ratio Lower Upper
142 100
141 88.2 70.0 105.0
115 38.0 27.8 41.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.549 min Scan# 1478
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 7252
Ion Ratio Lower Upper
164 100
162 103.3 80.2 120.2
160 48.1 36.9 55.3



#14
2,4,6-Tribromophenol
Concen: 0.389 ng
RT: 16.025 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:330 Resp: 1019
Ion Ratio Lower Upper
330 100
332 98.4 0.0 0.0#
141 38.5 21.3 31.9#









#15
2-Fluorobiphenyl
Concen: 0.460 ng
RT: 13.180 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	13232
Ion	Ratio	Lower	Upper
172	100		
171	34.6	27.4	41.0
170	22.9	18.4	27.6



#16
Acenaphthylene
Concen: 0.397 ng
RT: 14.271 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:	152	Resp:	12131
Ion	Ratio	Lower	Upper
152	100		
151	20.3	15.8	23.6
153	13.2	10.5	15.7









#17
Acenaphthene
Concen: 0.435 ng
RT: 14.613 min Scan# 1484
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 8867
Ion Ratio Lower Upper
154 100
153 112.4 90.9 136.3
152 55.9 45.0 67.4



#18
Fluorene
Concen: 0.428 ng
RT: 15.586 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:166 Resp: 10712
Ion Ratio Lower Upper
166 100
165 98.4 77.8 116.6
167 13.3 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.287 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:188 Resp: 13009
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.2 7.0 10.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.716 ng
RT: 15.660 min Scan# 1582
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:198 Resp: 674
Ion Ratio Lower Upper
198 100
182 18.4 15.5 23.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.432 ng
RT: 16.477 min Scan# 1661
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	3407
Ion	Ratio	Lower	Upper
248	100		
250	98.4	74.4	111.6
141	64.1	43.7	65.5



#22
Hexachlorobenzene
Concen: 0.502 ng
RT: 16.600 min Scan# 1673
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:	284	Resp:	4431
Ion	Ratio	Lower	Upper
284	100		
142	36.9	21.6	32.4#
249	32.2	23.5	35.3









#23
Atrazine
Concen: 0.371 ng
RT: 16.733 min Scan# 1686
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 1780
Ion Ratio Lower Upper
200 100
173 33.5 20.3 30.5#
215 44.2 40.4 60.6



#24
Pentachlorophenol
Concen: 0.344 ng
RT: 16.938 min Scan# 1706
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:266 Resp: 884
Ion Ratio Lower Upper
266 100
264 62.1 49.5 74.3
268 64.8 50.6 76.0









#25
Phenanthrene
Concen: 0.428 ng
RT: 17.328 min Scan# 1744
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 16284
Ion Ratio Lower Upper
178 100
176 19.4 15.3 22.9
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.414 ng
RT: 17.420 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:178 Resp: 13363
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.1 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.344 ng
RT: 19.312 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 13165
Ion Ratio Lower Upper
212 100
106 15.3 8.4 12.6#
104 8.7 4.5 6.7#



#28
Fluoranthene
Concen: 0.399 ng
RT: 19.341 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:202 Resp: 16742
Ion Ratio Lower Upper
202 100
101 11.4 7.2 10.8#
203 16.8 13.7 20.5









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.467 min Scan# 2448
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 9790
Ion Ratio Lower Upper
240 100
120 11.0 6.0 9.0#
236 28.4 21.4 32.2



#30
Pyrene
Concen: 0.497 ng
RT: 19.704 min Scan# 2213
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:202 Resp: 16677
Ion Ratio Lower Upper
202 100
200 21.2 16.6 25.0
203 17.6 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.330 ng
 RT: 19.906 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN018674.D
 Acq: 22 Feb 2022 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4


Tgt Ion:244 Resp: 8213
 Ion Ratio Lower Upper
 244 100
 212 11.3 6.2 9.2#
 122 0.0 7.9 11.9#



#32
 Benzo(a)anthracene
 Concen: 0.425 ng
 RT: 21.449 min Scan# 2446
 Delta R.T. 0.000 min
 Lab File: BN018674.D
 Acq: 22 Feb 2022 13:12

Tgt Ion:228 Resp: 13814
 Ion Ratio Lower Upper
 228 100
 226 27.1 21.3 31.9
 229 20.2 15.9 23.9









#33
Chrysene
Concen: 0.481 ng
RT: 21.503 min Scan# 2452
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:228 Resp: 15864
Ion Ratio Lower Upper
228 100
226 30.9 24.2 36.4
229 20.2 15.5 23.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.423 ng
RT: 21.378 min Scan# 2438
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:149 Resp: 5571
Ion Ratio Lower Upper
149 100
167 26.8 23.7 35.5
279 3.4 3.6 5.4#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.861 min Scan# 3050
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:264 Resp: 8770
Ion Ratio Lower Upper
264 100
260 28.4 20.2 30.4
265 34.9 18.2 27.2#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.538 ng
RT: 26.352 min Scan# 3902
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:276 Resp: 16959
Ion Ratio Lower Upper
276 100
138 29.4 14.9 22.3#
277 24.9 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.490 ng
RT: 23.130 min Scan# 2800
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 14790
Ion Ratio Lower Upper
252 100
253 24.4 17.8 26.6
125 14.0 6.0 9.0#



#38
Benzo(k)fluoranthene
Concen: 0.481 ng
RT: 23.176 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:252 Resp: 14524
Ion Ratio Lower Upper
252 100
253 24.8 17.7 26.5
125 13.2 6.2 9.2#









#39
Benzo(a)pyrene
Concen: 0.481 ng
RT: 23.752 min Scan# 3013
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 11620
Ion Ratio Lower Upper
252 100
253 25.9 17.8 26.6
125 16.7 7.0 10.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.535 ng
RT: 26.363 min Scan# 3906
Delta R.T. 0.000 min
Lab File: BN018674.D
Acq: 22 Feb 2022 13:12

Tgt Ion:278 Resp: 13150
Ion Ratio Lower Upper
278 100
139 21.3 10.8 16.2#
279 25.7 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.530 ng
 RT: 27.109 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BN018674.D
 Acq: 22 Feb 2022 13:12

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Tgt Ion:	276	Resp:	14547
Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.0	28.6
138	26.5	12.6	19.0#

