

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022422\
 Data File : BN018720.D
 Acq On : 24 Feb 2022 22:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 25 00:49:37 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 22 16:23:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.402	-0.5	100	0.00
3	n-Nitrosodimethylamine	0.400	0.412	-3.0	106	0.00
4 S	2-Fluorophenol	0.400	0.392	2.0	102	0.00
5 S	Phenol-d6	0.400	0.383	4.3	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.349	12.8	99	0.00
9	Naphthalene	0.400	0.385	3.8	102	-0.01
10	Hexachlorobutadiene	0.400	0.373	6.8	98	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	101	0.00
12	2-Methylnaphthalene	0.400	0.387	3.3	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	102	-0.01
16	Acenaphthylene	0.400	0.347	13.3	102	-0.01
17	Acenaphthene	0.400	0.368	8.0	102	-0.01
18	Fluorene	0.400	0.372	7.0	103	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.355	11.3	94	0.00
21	4-Bromophenyl-phenylether	0.400	0.358	10.5	105	0.00
22	Hexachlorobenzene	0.400	0.363	9.3	102	0.00
23	Atrazine	0.400	0.322	19.5	100	0.00
24	Pentachlorophenol	0.400	0.393	1.8	107	0.00
25	Phenanthrene	0.400	0.379	5.3	109	0.00
26	Anthracene	0.400	0.353	11.8	105	-0.01
27 SURR	Fluoranthene-d10	0.400	0.378	5.5	109	0.00
28	Fluoranthene	0.400	0.377	5.8	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
30	Pyrene	0.400	0.384	4.0	112	0.00
31 S	Terphenyl-d14	0.400	0.358	10.5	103	0.00
32	Benzo(a)anthracene	0.400	0.359	10.3	107	0.00
33	Chrysene	0.400	0.387	3.3	112	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.336	16.0	101	0.00
35 I	Perylene-d12	0.400	0.400	0.0	117	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.387	3.3	123	0.00
37	Benzo(b)fluoranthene	0.400	0.367	8.3	115	0.00
38	Benzo(k)fluoranthene	0.400	0.372	7.0	118	0.00
39 C	Benzo(a)pyrene	0.400	0.371	7.3	118	0.00
40	Dibenzo(a,h)anthracene	0.400	0.380	5.0	122	0.00
41	Benzo(g,h,i)perylene	0.400	0.371	7.3	118	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4