

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022521\
 Data File : BN013760.D
 Acq On : 25 Feb 2021 10:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 25 11:31:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 25 11:31:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.484	0.541	-11.8	119	0.00
3	n-Nitrosodimethylamine	0.373	0.378	-1.3	120	0.00
4 S	2-Fluorophenol	1.074	1.114	-3.7	92	0.00
5 S	Phenol-d6	1.428	1.490	-4.3	93	0.00
6	bis(2-Chloroethyl)ether	1.012	1.143	-12.9	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.257	0.288	-12.1	112	0.00
9	Naphthalene	0.983	1.050	-6.8	111	0.00
10	Hexachlorobutadiene	0.189	0.197	-4.2	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.647	0.666	-2.9	107	0.00
12	2-Methylnaphthalene	0.691	0.716	-3.6	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.164	0.159	3.0	91	0.00
15 S	2-Fluorobiphenyl	1.414	1.499	-6.0	101	0.00
16	Acenaphthylene	1.788	1.716	4.0	98	0.00
17	Acenaphthene	1.112	1.147	-3.1	104	0.00
18	Fluorene	1.434	1.438	-0.3	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.028	0.028	0.0	106	0.00
21	4-Bromophenyl-phenylether	0.217	0.221	-1.8	101	0.00
22	Hexachlorobenzene	0.227	0.249	-9.7	108	0.00
23	Atrazine	0.204	0.173	15.2	84	0.00
24	Pentachlorophenol	0.074	0.076	-2.7	110	0.00
25	Phenanthrene	1.042	1.105	-6.0	104	0.00
26	Anthracene	1.018	0.983	3.4	94	0.00
27 SURR	Fluoranthene-d10	1.180	1.163	1.4	97	0.00
28	Fluoranthene	1.248	1.264	-1.3	99	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
30	Pyrene	1.229	1.293	-5.2	97	0.00
31 S	Terphenyl-d14	0.883	0.906	-2.6	90	0.00
32	Benzo(a)anthracene	1.231	1.179	4.2	90	0.00
33	Chrysene	1.202	1.259	-4.7	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.585	0.407	30.4#	66	0.00
35	Indeno(1,2,3-cd)pyrene	1.458	1.375	5.7	88	0.00
36 I	Perylene-d12	1.000	1.000	0.0	82	0.00
37	Benzo(b)fluoranthene	1.279	1.333	-4.2	88	0.00
38	Benzo(k)fluoranthene	1.235	1.342	-8.7	92	0.00
39 C	Benzo(a)pyrene	1.150	1.164	-1.2	87	0.00
40	Dibenzo(a,h)anthracene	1.219	1.262	-3.5	88	0.00
41	Benzo(g,h,i)perylene	1.232	1.275	-3.5	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4