

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022521\
 Data File : BN013760.D
 Acq On : 25 Feb 2021 10:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Feb 25 11:31:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 25 11:31:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.741	152	4912	0.40	ng	0.00
7) Naphthalene-d8	10.530	136	18630	0.40	ng	0.00
13) Acenaphthene-d10	14.372	164	10920	0.40	ng	0.00
19) Phenanthrene-d10	17.128	188	23645	0.40	ng	# 0.00
29) Chrysene-d12	21.300	240	23500	0.40	ng	0.00
36) Perylene-d12	23.552	264	21913	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.325	112	5473	0.41	ng	0.00
5) Phenol-d6	6.903	99	7317	0.42	ng	0.00
8) Nitrobenzene-d5	8.883	82	5362	0.45	ng	0.00
11) 2-Methylnaphthalene-d10	12.120	152	12406	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.869	330	1736	0.39	ng	0.00
15) 2-Fluorobiphenyl	13.001	172	16374	0.42	ng	0.00
27) Fluoranthene-d10	19.156	212	27488	0.39	ng	0.00
31) Terphenyl-d14	19.761	244	21289	0.41	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.271	88	2656	0.45	ng	97
3) n-Nitrosodimethylamine	3.585	42	1855	0.41	ng	# 93
6) bis(2-Chloroethyl)ether	7.169	93	5615	0.45	ng	96
9) Naphthalene	10.568	128	19562	0.43	ng	99
10) Hexachlorobutadiene	10.873	225	3674	0.42	ng	# 99
12) 2-Methylnaphthalene	12.195	142	13339	0.41	ng	99
16) Acenaphthylene	14.092	152	18735	0.38	ng	100
17) Acenaphthene	14.435	154	12523	0.41	ng	99
18) Fluorene	15.425	166	15699	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.501	198	673	0.41	ng	# 66
21) 4-Bromophenyl-phenylether	16.325	248	5227	0.41	ng	# 87
22) Hexachlorobenzene	16.434	284	5876	0.44	ng	100
23) Atrazine	16.592	200	4086	0.34	ng	98
24) Pentachlorophenol	16.775	266	1797	0.41	ng	95
25) Phenanthrene	17.164	178	26120	0.42	ng	100
26) Anthracene	17.250	178	23254	0.39	ng	99
28) Fluoranthene	19.186	202	29889	0.41	ng	99
30) Pyrene	19.548	202	30387	0.42	ng	100
32) Benzo(a)anthracene	21.290	228	27713	0.38	ng	99
33) Chrysene	21.343	228	29590	0.42	ng	97
34) Bis(2-ethylhexyl)phtha...	21.237	149	9559	0.28	ng	100
35) Indeno(1,2,3-cd)pyrene	25.817	276	32311	0.38	ng	100
37) Benzo(b)fluoranthene	22.877	252	29209	0.42	ng	# 93
38) Benzo(k)fluoranthene	22.922	252	29404	0.43	ng	# 93
39) Benzo(a)pyrene	23.456	252	25501	0.40	ng	# 90
40) Dibenzo(a,h)anthracene	25.838	278	27660	0.41	ng	98
41) Benzo(g,h,i)perylene	26.509	276	27945	0.41	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022521\
Data File : BN013760.D
Acq On : 25 Feb 2021 10:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Feb 25 11:31:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 25 11:31:11 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.741 min Scan# 616
Delta R.T. -0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4


Tgt Ion:152 Resp: 4912
Ion Ratio Lower Upper
152 100
150 156.0 121.8 182.8
115 62.3 46.6 69.8



#2
1,4-Dioxane
Concen: 0.45 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion: 88 Resp: 2656
Ion Ratio Lower Upper
88 100
58 78.6 64.6 96.8
43 35.5 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.41 ng
RT: 3.585 min Scan# 100
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 1855
Ion Ratio Lower Upper
42 100
74 112.0 95.0 142.6
44 0.0 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.41 ng
RT: 5.325 min Scan# 316
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion: 112 Resp: 5473
Ion Ratio Lower Upper
112 100
64 60.5 44.3 66.5
63 33.0 23.0 34.6









#5
Phenol-d6
Concen: 0.42 ng
RT: 6.903 min Scan# 512
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 7317
Ion Ratio Lower Upper
99 100
42 23.4 15.2 22.8#
71 33.4 26.1 39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.45 ng
RT: 7.169 min Scan# 545
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion: 93 Resp: 5615
Ion Ratio Lower Upper
93 100
63 80.9 60.8 91.2
95 33.0 26.0 39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.530 min Scan# 862
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	18630
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.0	13.6
54	6.9	6.4	9.6
68	5.6	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.45 ng
RT: 8.883 min Scan# 732
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:	82	Resp:	5362
Ion	Ratio	Lower	Upper
82	100		
128	38.0	39.2	58.8#
54	58.4	40.1	60.1









#9
Naphthalene
Concen: 0.43 ng
RT: 10.568 min Scan# 865
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 19562
Ion Ratio Lower Upper
128 100
129 11.6 9.0 13.4
127 13.5 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.873 min Scan# 889
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:225 Resp: 3674
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.5 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 12.120 min Scan# 1077
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 12406
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.195 min Scan# 1094
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:142 Resp: 13339
Ion Ratio Lower Upper
142 100
141 88.4 70.5 105.7
115 35.0 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.372 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 10920
Ion Ratio Lower Upper
164 100
162 103.3 79.6 119.4
160 48.6 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.39 ng
RT: 15.869 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:330 Resp: 1736
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.0 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 13.001 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 16374
Ion Ratio Lower Upper
172 100
171 34.5 28.6 42.8
170 23.5 18.9 28.3



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.092 min Scan# 1289
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:152 Resp: 18735
Ion Ratio Lower Upper
152 100
151 19.5 15.7 23.5
153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 0.41 ng
RT: 14.435 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 12523
Ion Ratio Lower Upper
154 100
153 111.5 88.4 132.6
152 54.3 43.0 64.4



#18
Fluorene
Concen: 0.40 ng
RT: 15.425 min Scan# 1394
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:166 Resp: 15699
Ion Ratio Lower Upper
166 100
165 98.3 77.8 116.8
167 13.7 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 23645
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 9.0 13.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.41 ng
RT: 15.501 min Scan# 1400
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:198 Resp: 673
Ion Ratio Lower Upper
198 100
51 106.0 61.7 92.5#
105 63.3 33.0 49.6#









#21
4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.325 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 5227
Ion Ratio Lower Upper
248 100
250 98.5 75.5 113.3
141 46.6 53.8 80.8#



#22
Hexachlorobenzene
Concen: 0.44 ng
RT: 16.434 min Scan# 1475
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:284 Resp: 5876
Ion Ratio Lower Upper
284 100
142 34.8 27.9 41.9
249 30.1 24.2 36.2









#23
Atrazine
Concen: 0.34 ng
RT: 16.592 min Scan# 1488
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	4086
Ion Ratio	Lower	Upper	
200	100		
173	24.1	18.6	28.0
215	49.6	41.0	61.6



#24
Pentachlorophenol
Concen: 0.41 ng
RT: 16.775 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:	266	Resp:	1797
Ion Ratio	Lower	Upper	
266	100		
264	62.0	51.6	77.4
268	63.3	54.6	82.0









#25
Phenanthrene
Concen: 0.42 ng
RT: 17.164 min Scan# 1535
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 26120
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.39 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:178 Resp: 23254
Ion Ratio Lower Upper
178 100
176 18.6 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.39 ng
RT: 19.156 min Scan# 1836
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 27488
Ion Ratio Lower Upper
212 100
106 12.7 9.7 14.5
104 7.1 5.5 8.3



#28
Fluoranthene
Concen: 0.41 ng
RT: 19.186 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:202 Resp: 29889
Ion Ratio Lower Upper
202 100
101 10.8 8.1 12.1
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.300 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 23500
Ion Ratio Lower Upper
240 100
120 12.4 10.2 15.4
236 26.6 21.3 31.9



#30
Pyrene
Concen: 0.42 ng
RT: 19.548 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:202 Resp: 30387
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.3 14.0 21.0









#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.761 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 21289
Ion Ratio Lower Upper
244 100
212 7.5 6.1 9.1
122 10.7 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.290 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:228 Resp: 27713
Ion Ratio Lower Upper
228 100
226 26.6 20.5 30.7
229 20.1 16.2 24.4









#33
Chrysene
Concen: 0.42 ng
RT: 21.343 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 228 Resp: 29590




Ion	Ratio	Lower	Upper
228	100		
226	28.6	24.3	36.5
229	20.1	15.1	22.7




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.28 ng
RT: 21.237 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion: 149 Resp: 9559

Ion	Ratio	Lower	Upper
149	100		
167	27.8	22.1	33.1
279	4.2	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.38 ng
RT: 25.817 min Scan# 3230
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 32311
Ion Ratio Lower Upper
276 100
138 24.4 19.8 29.6
277 24.9 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.552 min Scan# 2568
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:264 Resp: 21913
Ion Ratio Lower Upper
264 100
260 24.5 19.3 28.9
265 30.0 66.2 99.4#









#37
Benzo(b)fluoranthene
Concen: 0.42 ng
RT: 22.877 min Scan# 2371
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 29209
Ion Ratio Lower Upper
252 100
253 22.9 20.4 30.6
125 10.6 11.6 17.4#



#38
Benzo(k)fluoranthene
Concen: 0.43 ng
RT: 22.922 min Scan# 2384
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:252 Resp: 29404
Ion Ratio Lower Upper
252 100
253 22.9 20.5 30.7
125 11.5 12.3 18.5#









#39
Benzo(a)pyrene
Concen: 0.40 ng
RT: 23.456 min Scan# 2540
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 25501
Ion Ratio Lower Upper
252 100
253 23.1 21.8 32.6
125 12.1 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.41 ng
RT: 25.838 min Scan# 3236
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Tgt Ion:278 Resp: 27660
Ion Ratio Lower Upper
278 100
139 16.9 13.1 19.7
279 25.2 21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.41 ng
RT: 26.509 min Scan# 3432
Delta R.T. 0.000 min
Lab File: BN013760.D
Acq: 25 Feb 2021 10:09

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:276 Resp: 27945
Ion Ratio Lower Upper
276 100
277 24.2 20.0 30.0
138 20.6 17.8 26.8

