



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022521\
 Data File : BN013759.D
 Acq On : 25 Feb 2021 09:30
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Feb 25 11:31:11 2021






AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2457

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.5	81981	PASS
68	69	0.00	2	1.6	1234	PASS
69	198	0.00	100	39.7	76523	PASS
70	69	0.00	2	0.6	436	PASS
127	198	10	80	48.3	93197	PASS
197	198	0.00	2	0.4	787	PASS
198	198	100	100	100.0	192832	PASS
199	198	5	9	6.5	12613	PASS
275	198	10	60	24.5	47293	PASS
365	198	1	100	3.4	6482	PASS
441	198	0.01	100	12.4	23819	PASS
442	442	50	100	100.0	148245	PASS
443	442	15	24	18.9	28045	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022521\
Data File : BN013759.D
Acq On : 25 Feb 2021 09:30
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 26 07:31:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 26 07:31:37 2021
Response via : Initial Calibration



TIC: BN013759.D\data.ms

(70) Pentachlorophenol (C)

16.775min (0.000) 25018.71 ng




response 200660

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.09
264.00	61.60	64.08
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022521\
 Data File : BN013759.D
 Acq On : 25 Feb 2021 09:30
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Feb 26 07:31:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 26 07:31:37 2021
 Response via : Initial Calibration



TIC: BN013759.D\data.ms

(77) Benzidine

19.375min (0.000) 0.00 ng

response 1342744

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.12
183.00	13.20	11.67
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
02/25/2021	BNA_N	<u>BN013759.D</u>
Compound Name	Response	Retention Time
DDT	540831	20.616
DDD	7762	20.175
DDE	373	19.669
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8135	548966	1.48

Instrument :
BNA_N
ClientSampleId :
DFTPP