

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030421\
 Data File : BN013834.D
 Acq On : 04 Mar 2021 15:56
 Operator : CG/JU
 Sample : PB134915BL
 Misc : SOM-MDL--W-BL-03
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK15

Quant Time: Mar 04 16:27:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030421MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 04 14:09:31 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	152513	20.00	ng/ul	0.00
18) Naphthalene-d8	10.50	136	624658	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.36	164	363081	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.10	188	753237	20.00	ng/ul	0.00
78) Chrysene-d12	21.28	240	736463	20.00	ng/ul	0.00
86) Perylene-d12	23.52	264	775651	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	24761	6.45	ng/uL	0.00
5) Phenol-d5	6.89	99	413682	33.54	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.06	67	261536	34.35	ng/ul	0.00
9) 2-Chlorophenol-d4	7.26	132	335638	35.58	ng/ul	0.00
13) 4-Methylphenol-d8	8.42	113	314835	33.70	ng/ul	0.00
19) Nitrobenzene-d5	8.87	128	148889	35.62	ng/ul	0.00
22) 2-Nitrophenol-d4	9.59	143	147699	36.99	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	292345	33.51	ng/ul	0.00
29) 4-Chloroaniline-d4	10.64	131	441816	40.90	ng/ul	0.00
44) Dimethylphthalate-d6	13.77	166	958252	39.30	ng/ul	0.00
47) Acenaphthylene-d8	14.05	160	1190301	36.92	ng/ul	0.00
52) 4-Nitrophenol-d4	14.55	143	145761	31.87	ng/ul	0.00
58) Fluorene-d10	15.35	176	813100	37.77	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.47	200	102894	29.39	ng/ul	0.00
71) Anthracene-d10	17.20	188	1299643	39.48	ng/ul	0.00
79) Pyrene-d10	19.49	212	1446904	40.39	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.37	264	1550864	40.16	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030421\
Data File : BN013834.D
Acq On : 04 Mar 2021 15:56
Operator : CG/JU
Sample : PB134915BL
Misc : SOM-MDL--W-BL-03
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK15

Quant Time: Mar 04 16:27:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030421MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 04 14:09:31 2021
Response via : Initial Calibration

