

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
 Data File : BN024179.D
 Acq On : 08 Mar 2023 22:22
 Operator : CG/JU
 Sample : 01746-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DCAC1

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/09/2023
 Supervised By : Jagrut Upadhyay 03/09/2023

Quant Time: Mar 09 03:52:01 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 03:48:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.059	152	8673	0.400	ng/u1	0.00
4) Naphthalene-d8	10.894	136	26543	0.400	ng/u1	# 0.02
9) Acenaphthene-d10	14.694	164	22619m	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.434	188	32798m	0.400	ng/u1	0.00
17) Chrysene-d12	21.624	240	23736m	0.400	ng/u1	0.00
23) Perylene-d12	24.132	264	20504m	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	43411	4.548	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.456	152	13500	0.428	ng/u1	0.00
18) Fluoranthene-d10	19.459	212	28988m	0.459	ng/u1	0.00
Target Compounds						
5) Naphthalene	10.955	128	35555184	536.657	ng/u1	Qvalue 95
7) 2-Methylnaphthalene	12.539	142	5197514	129.311	ng/u1	99
8) 1-Methylnaphthalene	12.753	142	4008952	94.906	ng/u1	99
10) Acenaphthylene	14.416	152	722864	7.659	ng/u1	98
11) Acenaphthene	14.763	153	5180778	71.255	ng/u1	97
12) Fluorene	15.744	166	3070366m	37.815	ng/u1	
14) Pentachlorophenol	17.084	266	466942m	51.257	ng/u1	
15) Phenanthrene	17.481	178	3699169m	39.379	ng/u1	
16) Anthracene	17.569	178	388160m	4.709	ng/u1	
19) Fluoranthene	19.487	202	264294m	2.798	ng/u1	
20) Pyrene	19.850	202	123568	1.289	ng/u1	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
Data File : BN024179.D
Acq On : 08 Mar 2023 22:22
Operator : CG/JU
Sample : 01746-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DCAC1

Quant Time: Mar 09 03:52:01 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 03:48:21 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/09/2023
Supervised By :Jagrut Upadhyay 03/09/2023

