

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
 Data File : BN024203.D
 Acq On : 09 Mar 2023 17:12
 Operator : CG/JU
 Sample : 01746-20
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DCAC5

Quant Time: Mar 10 02:24:40 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 09 08:02:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.055	152	8927	0.400	ng/ul	0.00
4) Naphthalene-d8	10.872	136	27486	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.689	164	15919	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.430	188	32264	0.400	ng/ul	0.00
17) Chrysene-d12	21.621	240	20967	0.400	ng/ul	0.00
23) Perylene-d12	24.123	264	17458	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.410	96	41579	4.232	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.456	152	10800	0.331	ng/ul	0.00
18) Fluoranthene-d10	19.455	212	27268	0.489	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.922	128	9078	0.132	ng/ul	97
7) 2-Methylnaphthalene	12.528	142	5059	0.122	ng/ul	98
8) 1-Methylnaphthalene	12.742	142	2897	0.066	ng/ul	93
11) Acenaphthene	14.749	153	1672	0.033	ng/ul	96
12) Fluorene	15.735	166	1895	0.033	ng/ul#	87
14) Pentachlorophenol	17.076	266	4499	0.502	ng/ul#	73
15) Phenanthrene	17.472	178	2821	0.031	ng/ul#	84
16) Anthracene	17.565	178	2824	0.035	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
Data File : BN024203.D
Acq On : 09 Mar 2023 17:12
Operator : CG/JU
Sample : 01746-20
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DCAC5

Quant Time: Mar 10 02:24:40 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 09 08:02:24 2023
Response via : Initial Calibration

