

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
 Data File : BN024221.D
 Acq On : 10 Mar 2023 04:51
 Operator : CG/JU
 Sample : 01784-12
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DCAC0

Quant Time: Mar 10 05:43:22 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 10 03:43:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.050	152	8900	0.400	ng/ul	0.00
4) Naphthalene-d8	10.866	136	27433	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.684	164	16371	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.430	188	33518	0.400	ng/ul	0.00
17) Chrysene-d12	21.620	240	23561	0.400	ng/ul	0.01
23) Perylene-d12	24.137	264	21332	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	40101	4.094	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.450	152	10641	0.326	ng/ul	0.00
18) Fluoranthene-d10	19.450	212	28235	0.450	ng/ul	0.00
Target Compounds						
					Qvalue	
7) 2-Methylnaphthalene	12.522	142	1129	0.027	ng/ul	96
10) Acenaphthylene	14.406	152	11647	0.171	ng/ul	97
11) Acenaphthene	14.749	153	144639	2.749	ng/ul	98
12) Fluorene	15.716	166	7369	0.125	ng/ul#	1
14) Pentachlorophenol	17.075	266	8086	0.869	ng/ul#	73
15) Phenanthrene	17.472	178	4034	0.042	ng/ul#	89
16) Anthracene	17.565	178	7048	0.084	ng/ul#	94
19) Fluoranthene	19.482	202	4028	0.043	ng/ul#	88
20) Pyrene	19.845	202	5154	0.054	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
Data File : BN024221.D
Acq On : 10 Mar 2023 04:51
Operator : CG/JU
Sample : 01784-12
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DCAC0

Quant Time: Mar 10 05:43:22 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 10 03:43:59 2023
Response via : Initial Calibration

