

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
 Data File : BN024227.D
 Acq On : 10 Mar 2023 08:29
 Operator : CG/JU
 Sample : PB151261BS
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS261

Quant Time: Mar 11 00:06:32 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 10 03:43:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.051	152	6397	0.400	ng/ul	0.00
4) Naphthalene-d8	10.867	136	18643	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.684	164	9800	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.430	188	20162	0.400	ng/ul	0.00
17) Chrysene-d12	21.618	240	15752	0.400	ng/ul	0.00
23) Perylene-d12	24.132	264	13075	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.401	96	5192	0.738	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.451	152	9326	0.421	ng/ul	0.00
18) Fluoranthene-d10	19.455	212	17331	0.413	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.439	88	13679	1.786	ng/ul	89
5) Naphthalene	10.916	128	19840	0.426	ng/ul	99
7) 2-Methylnaphthalene	12.522	142	12112	0.429	ng/ul	100
8) 1-Methylnaphthalene	12.742	142	12757	0.430	ng/ul	100
10) Acenaphthylene	14.407	152	16652	0.407	ng/ul	100
11) Acenaphthene	14.749	153	13166	0.418	ng/ul	99
12) Fluorene	15.730	166	14274	0.406	ng/ul	99
14) Pentachlorophenol	17.071	266	4271	0.763	ng/ul#	73
15) Phenanthrene	17.468	178	23980	0.415	ng/ul	100
16) Anthracene	17.561	178	20123	0.397	ng/ul	100
19) Fluoranthene	19.483	202	26014	0.415	ng/ul	100
20) Pyrene	19.845	202	26569	0.418	ng/ul	100
21) Benzo(a)anthracene	21.600	228	21171	0.412	ng/ul	100
22) Chrysene	21.656	228	23453	0.432	ng/ul	100
24) Benzo(b)fluoranthene	23.357	252	22106	0.443	ng/ul	99
25) Benzo(k)fluoranthene	23.410	252	21969	0.424	ng/ul	98
26) Benzo(a)pyrene	24.018	252	17915	0.417	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	26.767	276	19426	0.401	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.790	278	13923	0.378	ng/ul	97
29) Benzo(g,h,i)perylene	27.582	276	17939	0.427	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
Data File : BN024227.D
Acq On : 10 Mar 2023 08:29
Operator : CG/JU
Sample : PB151261BS
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS261

Quant Time: Mar 11 00:06:32 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 10 03:43:59 2023
Response via : Initial Calibration

