

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
 Data File : BN024229.D
 Acq On : 10 Mar 2023 10:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4391

Quant Time: Mar 11 00:07:16 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 10 03:43:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.051	152	6723	0.400	ng/ul	0.00
4) Naphthalene-d8	10.867	136	19352	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.684	164	9666	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.430	188	19569	0.400	ng/ul	0.00
17) Chrysene-d12	21.612	240	14904	0.400	ng/ul	0.00
23) Perylene-d12	24.126	264	12777	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.406	96	2860	0.387	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.451	152	9201	0.400	ng/ul	0.00
18) Fluoranthene-d10	19.450	212	16083	0.406	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.439	88	3065	0.381	ng/ul	96
5) Naphthalene	10.916	128	19638	0.407	ng/ul	99
7) 2-Methylnaphthalene	12.522	142	11721	0.400	ng/ul	100
8) 1-Methylnaphthalene	12.742	142	12253	0.398	ng/ul	100
10) Acenaphthylene	14.407	152	15601	0.387	ng/ul	100
11) Acenaphthene	14.749	153	12609	0.406	ng/ul	100
12) Fluorene	15.730	166	13729	0.396	ng/ul	98
14) Pentachlorophenol	17.071	266	1913	0.352	ng/ul#	73
15) Phenanthrene	17.472	178	22764	0.406	ng/ul	99
16) Anthracene	17.561	178	18471	0.376	ng/ul	100
19) Fluoranthene	19.483	202	24011	0.405	ng/ul	99
20) Pyrene	19.845	202	24401	0.405	ng/ul	99
21) Benzo(a)anthracene	21.598	228	17808	0.366	ng/ul	100
22) Chrysene	21.653	228	20837	0.406	ng/ul	99
24) Benzo(b)fluoranthene	23.354	252	19730	0.404	ng/ul	98
25) Benzo(k)fluoranthene	23.401	252	19435	0.384	ng/ul	97
26) Benzo(a)pyrene	24.012	252	15962	0.380	ng/ul	95
27) Indeno(1,2,3-cd)pyrene	26.760	276	18060	0.382	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.777	278	12781	0.356	ng/ul	98
29) Benzo(g,h,i)perylene	27.568	276	16951	0.413	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
Data File : BN024229.D
Acq On : 10 Mar 2023 10:19
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4391

Quant Time: Mar 11 00:07:16 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 10 03:43:59 2023
Response via : Initial Calibration

