

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
 Data File : BN024241.D
 Acq On : 10 Mar 2023 18:22
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 82 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4392

Quant Time: Mar 11 00:09:17 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 10 03:43:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.051	152	8845	0.400	ng/ul	0.00
4) Naphthalene-d8	10.867	136	25647	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.684	164	13313	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.430	188	27022	0.400	ng/ul	0.00
17) Chrysene-d12	21.618	240	19209	0.400	ng/ul	0.00
23) Perylene-d12	24.135	264	15024	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	3728	0.383	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.450	152	12362	0.406	ng/ul	0.00
18) Fluoranthene-d10	19.455	212	21856	0.428	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.443	88	4046	0.382	ng/ul	94
5) Naphthalene	10.916	128	25992	0.406	ng/ul	100
7) 2-Methylnaphthalene	12.522	142	15736	0.405	ng/ul	99
8) 1-Methylnaphthalene	12.742	142	16604	0.407	ng/ul	100
10) Acenaphthylene	14.407	152	21126	0.380	ng/ul	99
11) Acenaphthene	14.749	153	17171	0.401	ng/ul	99
12) Fluorene	15.730	166	18889	0.395	ng/ul	97
14) Pentachlorophenol	17.075	266	2548	0.339	ng/ul#	73
15) Phenanthrene	17.472	178	31270	0.404	ng/ul	99
16) Anthracene	17.561	178	25714	0.379	ng/ul	100
19) Fluoranthene	19.483	202	32517	0.425	ng/ul	100
20) Pyrene	19.845	202	32868	0.424	ng/ul	99
21) Benzo(a)anthracene	21.597	228	23170	0.370	ng/ul	100
22) Chrysene	21.656	228	26538	0.401	ng/ul	99
24) Benzo(b)fluoranthene	23.360	252	23902	0.417	ng/ul	98
25) Benzo(k)fluoranthene	23.415	252	23236	0.390	ng/ul	98
26) Benzo(a)pyrene	24.012	252	18677	0.378	ng/ul	95
27) Indeno(1,2,3-cd)pyrene	26.763	276	20013	0.360	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.784	278	14274	0.338	ng/ul	98
29) Benzo(g,h,i)perylene	27.582	276	18608	0.386	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030923\
Data File : BN024241.D
Acq On : 10 Mar 2023 18:22
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 82 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4392

Quant Time: Mar 11 00:09:17 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 10 03:43:59 2023
Response via : Initial Calibration

