

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032023\
 Data File : BN024439.D
 Acq On : 21 Mar 2023 09:02
 Operator : CG/JU
 Sample : 01996-06
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 H0A27

Quant Time: Mar 21 17:37:40 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 21 02:42:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.030	152	10685	0.400	ng/ul	0.00
4) Naphthalene-d8	10.845	136	33034	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.666	164	18899	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.413	188	38833	0.400	ng/ul	0.00
17) Chrysene-d12	21.600	240	30611	0.400	ng/ul	0.00
23) Perylene-d12	24.094	264	28899	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	65322	5.555	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.429	152	13148	0.335	ng/ul	0.00
18) Fluoranthene-d10	19.436	212	33869	0.416	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.427	88	1201741	93.955	ng/ul	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032023\
Data File : BN024439.D
Acq On : 21 Mar 2023 09:02
Operator : CG/JU
Sample : 01996-06
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
H0A27

Quant Time: Mar 21 17:37:40 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN030923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 21 02:42:36 2023
Response via : Initial Calibration

