

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032123\
 Data File : BN024457.D
 Acq On : 21 Mar 2023 21:56
 Operator : CG/JU
 Sample : 01880-14DL 2X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20230310DL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/22/2023
 Supervised By : Jagrut Upadhyay 03/22/2023

Quant Time: Mar 22 04:16:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 04:14:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.032	152	10049	0.400	ng	0.00
7) Naphthalene-d8	10.845	136	28974	0.400	ng	0.00
13) Acenaphthene-d10	14.664	164	14725	0.400	ng	0.00
19) Phenanthrene-d10	17.413	188	30645	0.400	ng	0.00
29) Chrysene-d12	21.601	240	22364	0.400	ng	0.00
35) Perylene-d12	24.085	264	19068	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.577	112	1543	0.069	ng	0.00
5) Phenol-d6	7.187	99	1138	0.040	ng	0.00
8) Nitrobenzene-d5	9.200	82	3233m	0.167	ng	0.01
11) 2-Methylnaphthalene-d10	12.429	152	5452	0.153	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	974	0.129	ng	0.01
15) 2-Fluorobiphenyl	13.296	172	8987	0.162	ng	0.00
27) Fluoranthene-d10	19.437	212	11054	0.170	ng	0.00
31) Terphenyl-d14	20.024	244	6583	0.175	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.425	88	28733	2.628	ng	98
34) Bis(2-ethylhexyl)phtha...	21.493	149	1848	0.053	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032123\
Data File : BN024457.D
Acq On : 21 Mar 2023 21:56
Operator : CG/JU
Sample : 01880-14DL 2X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20230310DL

Quant Time: Mar 22 04:16:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 22 04:14:21 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/22/2023
Supervised By :Jagrut Upadhyay 03/22/2023

