

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032123\
 Data File : BN024468.D
 Acq On : 22 Mar 2023 04:38
 Operator : CG/JU
 Sample : 01904-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB02-20230313

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/22/2023
 Supervised By :Jagrut Upadhyay 03/22/2023

Quant Time: Mar 22 05:10:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 04:14:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							
1) 1,4-Dichlorobenzene-d4	8.033	152	9857	0.400	ng	0.00	
7) Naphthalene-d8	10.845	136	28777	0.400	ng	0.00	
13) Acenaphthene-d10	14.665	164	14878	0.400	ng	0.00	
19) Phenanthrene-d10	17.400	188	31661	0.400	ng	#-0.01	
29) Chrysene-d12	21.592	240	20138	0.400	ng	0.00	
35) Perylene-d12	24.073	264	15446	0.400	ng	# 0.00	
System Monitoring Compounds							
4) 2-Fluorophenol	5.577	112	3894	0.178	ng	0.00	
5) Phenol-d6	7.188	99	2802	0.100	ng	0.00	
8) Nitrobenzene-d5	9.190	82	8314m	0.432	ng	0.00	
11) 2-Methylnaphthalene-d10	12.425	152	14057	0.397	ng	0.00	
14) 2,4,6-Tribromophenol	16.147	330	2770	0.364	ng	0.00	
15) 2-Fluorobiphenyl	13.285	172	23608	0.420	ng	-0.01	
27) Fluoranthene-d10	19.433	212	29072	0.433	ng	0.00	
31) Terphenyl-d14	20.025	244	20226	0.596	ng	0.00	
Target Compounds							
25) Phenanthrene	17.450	178	1961	0.021	ng	#	87
34) Bis(2-ethylhexyl)phtha...	21.484	149	3808	0.120	ng	#	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032123\
 Data File : BN024468.D
 Acq On : 22 Mar 2023 04:38
 Operator : CG/JU
 Sample : 01904-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
ClientSampleId :
 EB02-20230313

Quant Time: Mar 22 05:10:01 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 04:14:21 2023
 Response via : Initial Calibration

Manual Integrations
APPROVED
 Reviewed By :Christian Giraldo 03/22/2023
 Supervised By :Jagrut Upadhyay 03/22/2023

