

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032123\  
 Data File : BN024455.D  
 Acq On : 21 Mar 2023 20:05  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Mar 22 04:15:58 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN031323.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 22 04:14:21 2023  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev  | Area% | Dev(min) |
|---------|----------------------------|--------|-------|-------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0   | 100   | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.403 | -0.8  | 96    | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.389 | 2.8   | 96    | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.431 | -7.7  | 108   | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.407 | -1.7  | 104   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.395 | 1.3   | 96    | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0   | 97    | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.468 | -17.0 | 121   | 0.00     |
| 9       | Naphthalene                | 0.400  | 0.404 | -1.0  | 100   | 0.00     |
| 10      | Hexachlorobutadiene        | 0.400  | 0.414 | -3.5  | 100   | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.391 | 2.3   | 98    | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.387 | 3.3   | 97    | 0.00     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0   | 93    | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.338 | 15.5  | 86    | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.406 | -1.5  | 95    | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.366 | 8.5   | 89    | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.376 | 6.0   | 89    | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.349 | 12.8  | 83    | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0   | 89    | 0.00     |
| 20      | 4,6-Dinitro-2-methylphenol | 0.400  | 0.310 | 22.5  | 52    | 0.00     |
| 21      | 4-Bromophenyl-phenylether  | 0.400  | 0.381 | 4.8   | 88    | 0.00     |
| 22      | Hexachlorobenzene          | 0.400  | 0.387 | 3.3   | 89    | 0.00     |
| 23      | Atrazine                   | 0.400  | 0.364 | 9.0   | 92    | 0.00     |
| 24      | Pentachlorophenol          | 0.400  | 0.340 | 15.0  | 72    | 0.00     |
| 25      | Phenanthrene               | 0.400  | 0.378 | 5.5   | 87    | 0.00     |
| 26      | Anthracene                 | 0.400  | 0.360 | 10.0  | 86    | 0.00     |
| 27 SURR | Fluoranthene-d10           | 0.400  | 0.397 | 0.8   | 93    | 0.00     |
| 28      | Fluoranthene               | 0.400  | 0.383 | 4.3   | 89    | 0.00     |
| 29 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0   | 92    | 0.00     |
| 30      | Pyrene                     | 0.400  | 0.403 | -0.8  | 92    | 0.00     |
| 31 S    | Terphenyl-d14              | 0.400  | 0.353 | 11.8  | 80    | 0.00     |
| 32      | Benzo(a)anthracene         | 0.400  | 0.398 | 0.5   | 95    | 0.00     |
| 33      | Chrysene                   | 0.400  | 0.390 | 2.5   | 89    | 0.00     |
| 34      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.462 | -15.5 | 115   | 0.00     |
| 35 I    | Perylene-d12               | 0.400  | 0.400 | 0.0   | 101   | 0.00     |
| 36      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.410 | -2.5  | 109   | 0.00     |
| 37      | Benzo(b)fluoranthene       | 0.400  | 0.378 | 5.5   | 101   | 0.00     |
| 38      | Benzo(k)fluoranthene       | 0.400  | 0.363 | 9.3   | 97    | 0.00     |
| 39 C    | Benzo(a)pyrene             | 0.400  | 0.376 | 6.0   | 101   | 0.00     |
| 40      | Dibenzo(a,h)anthracene     | 0.400  | 0.411 | -2.7  | 110   | 0.00     |
| 41      | Benzo(g,h,i)perylene       | 0.400  | 0.415 | -3.7  | 110   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

**Instrument :**  
BNA\_N  
**LabSampleId :**  
SSTDCCC0.4