

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032123\
 Data File : BN024470.D
 Acq On : 22 Mar 2023 05:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 22 06:20:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 22 04:14:21 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	132	0.00
2	1,4-Dioxane	0.400	0.389	2.8	123	0.00
3	n-Nitrosodimethylamine	0.400	0.403	-0.8	131	0.00
4 S	2-Fluorophenol	0.400	0.410	-2.5	136	0.00
5 S	Phenol-d6	0.400	0.397	0.8	135	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	130	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.468	-17.0	164	0.00
9	Naphthalene	0.400	0.403	-0.8	135	-0.01
10	Hexachlorobutadiene	0.400	0.417	-4.2	137	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	133	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	131	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	0.313	21.8	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	128	-0.01
16	Acenaphthylene	0.400	0.362	9.5	118	0.00
17	Acenaphthene	0.400	0.377	5.8	120	-0.01
18	Fluorene	0.400	0.358	10.5	115	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	118	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.313	21.8	70	-0.01
21	4-Bromophenyl-phenylether	0.400	0.380	5.0	116	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	118	0.00
23	Atrazine	0.400	0.358	10.5	119	-0.01
24	Pentachlorophenol	0.400	0.316	21.0	84	-0.01
25	Phenanthrene	0.400	0.381	4.8	115	0.00
26	Anthracene	0.400	0.361	9.8	113	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	116	0.00
28	Fluoranthene	0.400	0.371	7.3	114	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.442	-10.5	115	0.00
31 S	Terphenyl-d14	0.400	0.382	4.5	100	0.00
32	Benzo(a)anthracene	0.400	0.384	4.0	105	0.00
33	Chrysene	0.400	0.395	1.3	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.453	-13.2	130	0.00
35 I	Perylene-d12	0.400	0.400	0.0	102	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	101	0.00
37	Benzo(b)fluoranthene	0.400	0.398	0.5	108	0.00
38	Benzo(k)fluoranthene	0.400	0.370	7.5	100	0.00
39 C	Benzo(a)pyrene	0.400	0.374	6.5	102	0.00
40	Dibenzo(a,h)anthracene	0.400	0.363	9.3	99	-0.01
41	Benzo(g,h,i)perylene	0.400	0.395	1.3	106	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4