

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032421\
 Data File : BN014074.D
 Acq On : 24 Mar 2021 17:23
 Operator : CG/JU
 Sample : PB135330BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB135330BS

Manual Integrations
 APPROVED

mohammad
 3/25/2021 11:50:47 AM

Quant Time: Mar 24 17:57:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 23 11:06:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	4808	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	18427	0.40	ng	#-0.01
13) Acenaphthene-d10	14.321	164	10845	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	22988	0.40	ng	-0.01
29) Chrysene-d12	21.239	240	21917	0.40	ng	# 0.00
36) Perylene-d12	23.441	264	20605	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	5147	0.47	ng	0.00
5) Phenol-d6	6.863	99	6561	0.47	ng	0.00
8) Nitrobenzene-d5	8.832	82	4956	0.41	ng	-0.01
11) 2-Methylnaphthalene-d10	12.062	152	12892	0.45	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1332	0.38	ng	-0.01
15) 2-Fluorobiphenyl	12.938	172	17103	0.42	ng	-0.01
27) Fluoranthene-d10	19.089	212	27947	0.44	ng	0.00
31) Terphenyl-d14	19.690	244	22532	0.47	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	2611	0.41	ng	99
3) n-Nitrosodimethylamine	3.585	42	1569	0.37	ng	# 90
6) bis(2-Chloroethyl)ether	7.128	93	5616	0.45	ng	95
9) Naphthalene	10.518	128	20417	0.45	ng	99
10) Hexachlorobutadiene	10.809	225	3740	0.44	ng	# 100
12) 2-Methylnaphthalene	12.133	142	13794	0.45	ng	99
16) Acenaphthylene	14.042	152	16626	0.40	ng	100
17) Acenaphthene	14.384	154	12701	0.42	ng	98
18) Fluorene	15.374	166	16366	0.43	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	950	0.37	ng	93
21) 4-Bromophenyl-phenylether	16.264	248	5084	0.43	ng	# 77
22) Hexachlorobenzene	16.374	284	5892	0.43	ng	99
23) Atrazine	16.532	200	3034	0.38	ng	97
24) Pentachlorophenol	16.715	266	1051	0.33	ng	94
25) Phenanthrene	17.104	178	27450	0.44	ng	100
26) Anthracene	17.189	178	21776	0.42	ng	100
28) Fluoranthene	19.119	202	30194	0.43	ng	99
30) Pyrene	19.481	202	30562	0.45	ng	100
32) Benzo(a)anthracene	21.218	228	26700	0.43	ng	98
33) Chrysene	21.271	228	30273	0.45	ng	98
34) Bis(2-ethylhexyl)phtha...	21.154	149	7775	0.33	ng	98
35) Indeno(1,2,3-cd)pyrene	25.646	276	33943	0.46	ng	99
37) Benzo(b)fluoranthene	22.781	252	32002	0.43	ng	97
38) Benzo(k)fluoranthene	22.822	252	31344m	0.43	ng	
39) Benzo(a)pyrene	23.342	252	27935	0.44	ng	98
40) Dibenzo(a,h)anthracene	25.659	278	28746	0.42	ng	99
41) Benzo(g,h,i)perylene	26.327	276	30340	0.42	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032421\
Data File : BN014074.D
Acq On : 24 Mar 2021 17:23
Operator : CG/JU
Sample : PB135330BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM

Quant Time: Mar 24 17:57:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 23 11:06:27 2021
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.700 min Scan# 611
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Tgt Ion: 152 Resp: 4808

Ion	Ratio	Lower	Upper
152	100		
150	151.8	121.8	182.8
115	59.2	46.6	69.8





#2
1,4-Dioxane
Concen: 0.41 ng
RT: 3.239 min Scan# 57
Delta R.T. -0.008 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23



Tgt Ion: 88 Resp: 2611


Ion	Ratio	Lower	Upper
88	100		
58	81.7	64.6	96.8
43	33.7	26.6	40.0



Instrument :
BNA_N
ClientSampled :
PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM







#3
n-Nitrosodimethylamine
Concen: 0.37 ng
RT: 3.585 min Scan# 100
Delta R.T. -0.008 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 42 Resp: 1569
Ion Ratio Lower Upper
42 100
74 108.0 95.0 142.6
44 0.0 5.8 8.8#

Instrument :
BNA_N
ClientSampled :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#4
2-Fluorophenol
Concen: 0.47 ng
RT: 5.292 min Scan# 312
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 112 Resp: 5147
Ion Ratio Lower Upper
112 100
64 60.7 44.3 66.5
63 31.7 23.0 34.6









#5
Phenol-d6
Concen: 0.47 ng
RT: 6.863 min Scan# 507
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 99 Resp: 6561
Ion Ratio Lower Upper
99 100
42 22.1 15.2 22.8
71 34.1 26.1 39.1

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.45 ng
RT: 7.128 min Scan# 540
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23


Tgt Ion: 93 Resp: 5616
Ion Ratio Lower Upper
93 100
63 81.5 60.8 91.2
95 33.0 26.0 39.0





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.012 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Instrument :
BNA_N
ClientSampleId :
PB135330BS



Tgt Ion:	136	Resp:	18427
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	9.2	6.4	9.6
68	6.8	4.5	6.7#




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#8
Nitrobenzene-d5
Concen: 0.41 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.012 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	82	Resp:	4956
Ion	Ratio	Lower	Upper
82	100		
128	31.9	39.2	58.8#
54	63.4	40.1	60.1#









#9
Naphthalene
Concen: 0.45 ng
RT: 10.518 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	128	Resp:	20417
Ion Ratio	Lower	Upper	
128	100		
129	11.4	9.0	13.4
127	13.2	10.4	15.6

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#10
Hexachlorobutadiene
Concen: 0.44 ng
RT: 10.809 min Scan# 884
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	225	Resp:	3740
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.0	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.45 ng
RT: 12.062 min Scan# 1064
Delta R.T. -0.004 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:152 Resp: 12892
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#12
2-Methylnaphthalene
Concen: 0.45 ng
RT: 12.133 min Scan# 1080
Delta R.T. -0.004 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:142 Resp: 13794
Ion Ratio Lower Upper
142 100
141 88.7 70.5 105.7
115 36.2 27.4 41.0





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23





Tgt Ion:164 Resp: 10845

Ion	Ratio	Lower	Upper
164	100		
162	100.0	79.6	119.4
160	44.8	36.0	54.0

Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED




mohammad
3/25/2021 11:50:47 AM




#14
2,4,6-Tribromophenol
Concen: 0.38 ng
RT: 15.805 min Scan# 1424
Delta R.T. -0.012 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:330 Resp: 1332

Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	36.8	26.8	40.2









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.012 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	172	Resp:	17103
Ion	Ratio	Lower	Upper
172	100		
171	35.0	28.6	42.8
170	24.1	18.9	28.3

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM




#16
Acenaphthylene
Concen: 0.40 ng
RT: 14.042 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	152	Resp:	16626
Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.7	23.5
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.42 ng
RT: 14.384 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23





Tgt Ion:154 Resp: 12701
Ion Ratio Lower Upper
154 100
153 112.7 88.4 132.6
152 55.0 43.0 64.4



#18
Fluorene
Concen: 0.43 ng
RT: 15.374 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Tgt Ion:166 Resp: 16366
Ion Ratio Lower Upper
166 100
165 98.2 77.8 116.8
167 13.4 10.6 16.0



Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM







#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.055 min Scan# 1526
Delta R.T. -0.012 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:188 Resp: 22988
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.4 9.0 13.6

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:198 Resp: 950
Ion Ratio Lower Upper
198 100
51 84.0 61.7 92.5
105 38.8 33.0 49.6








#21
4-Bromophenyl-phenylether
Concen: 0.43 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Tgt Ion: 248 Resp: 5084

Ion	Ratio	Lower	Upper
248	100		
250	103.7	75.5	113.3
141	34.3	53.8	80.8#





#22
Hexachlorobenzene
Concen: 0.43 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23



Tgt Ion: 284 Resp: 5892


Ion	Ratio	Lower	Upper
284	100		
142	35.8	27.9	41.9
249	30.7	24.2	36.2



Instrument :
BNA_N
ClientSampleID :
PB135330BS


Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM



#23
Atrazine
Concen: 0.38 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Instrument :
BNA_N
ClientSampleId :
PB135330BS




Tgt Ion:200 Resp: 3034
Ion Ratio Lower Upper
200 100
173 25.3 18.6 28.0
215 49.8 41.0 61.6

Manual Integrations
APPROVED



mohammad
3/25/2021 11:50:47 AM




#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.715 min Scan# 1498
Delta R.T. -0.012 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Tgt Ion:266 Resp: 1051
Ion Ratio Lower Upper
266 100
264 62.6 51.6 77.4
268 60.8 54.6 82.0






#25
Phenanthrene
Concen: 0.44 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23



Tgt Ion:178 Resp: 27450


Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.0	22.4
179	15.3	12.3	18.5




Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM






#26
Anthracene
Concen: 0.42 ng
RT: 17.189 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23



Tgt Ion:178 Resp: 21776

Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.6	21.8
179	15.5	12.2	18.2









#27
Fluoranthene-d10
Concen: 0.44 ng
RT: 19.089 min Scan# 1821
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	212	Resp:	27947
Ion Ratio	Lower	Upper	
212	100		
106	12.7	9.7	14.5
104	7.0	5.5	8.3

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#28
Fluoranthene
Concen: 0.43 ng
RT: 19.119 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	202	Resp:	30194
Ion Ratio	Lower	Upper	
202	100		
101	11.1	8.1	12.1
203	17.3	13.9	20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.239 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 240 Resp: 21917
Ion Ratio Lower Upper
240 100
120 6.6 10.2 15.4#
236 25.4 21.3 31.9

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM




#30
Pyrene
Concen: 0.45 ng
RT: 19.481 min Scan# 1900
Delta R.T. -0.005 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 202 Resp: 30562
Ion Ratio Lower Upper
202 100
200 20.3 16.3 24.5
203 17.5 14.0 21.0









#31
Terphenyl-d14
Concen: 0.47 ng
RT: 19.690 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23





Tgt Ion:244 Resp: 22532
Ion Ratio Lower Upper
244 100
212 7.5 6.1 9.1
122 11.0 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.43 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Tgt Ion:228 Resp: 26700
Ion Ratio Lower Upper
228 100
226 26.9 20.5 30.7
229 19.8 16.2 24.4



Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM







#33
Chrysene
Concen: 0.45 ng
RT: 21.271 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	228	Resp:	30273
Ion Ratio	Lower	Upper	
228	100		
226	28.9	24.3	36.5
229	19.6	15.1	22.7

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.33 ng
RT: 21.154 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	149	Resp:	7775
Ion Ratio	Lower	Upper	
149	100		
167	26.6	22.1	33.1
279	4.9	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.46 ng
RT: 25.646 min Scan# 3178
Delta R.T. -0.006 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 276 Resp: 33943
Ion Ratio Lower Upper
276 100
138 24.0 19.8 29.6
277 24.9 19.9 29.9

Instrument :
BNA_N
ClientSampleId :
PB135330BS




Manual Integrations
APPROVED


mohammad
3/25/2021 11:50:47 AM



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.441 min Scan# 2534
Delta R.T. -0.006 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion: 264 Resp: 20605
Ion Ratio Lower Upper
264 100
260 24.4 19.3 28.9
265 36.4 66.2 99.4#





#37

Benzo(b)fluoranthene

Concen: 0.43 ng

RT: 22.781 min Scan# 2341

Delta R.T. -0.003 min

Lab File: BN014074.D

Acq: 24 Mar 2021 17:23

Tgt Ion:	252	Resp:	32002
Ion Ratio	Lower	Upper	
252	100		
253	24.6	20.4	30.6
125	12.9	11.6	17.4

Instrument :





BNA_N

ClientSampleId :

PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM



#38

Benzo(k)fluoranthene

Concen: 0.43 ng m




RT: 22.822 min Scan# 2353


Delta R.T. -0.006 min

Lab File: BN014074.D


Acq: 24 Mar 2021 17:23

Tgt Ion:	252	Resp:	31344
Ion Ratio	Lower	Upper	
252	100		
253	24.9	20.5	30.7
125	11.1	12.3	18.5#









#39
Benzo(a)pyrene
Concen: 0.44 ng
RT: 23.342 min Scan# 2505
Delta R.T. -0.010 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23





Tgt Ion:252 Resp: 27935
Ion Ratio Lower Upper
252 100
253 27.0 21.8 32.6
125 16.0 14.6 21.8



#40
Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 25.659 min Scan# 3182
Delta R.T. -0.010 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23




Tgt Ion:278 Resp: 28746
Ion Ratio Lower Upper
278 100
139 16.6 13.1 19.7
279 25.6 21.4 32.0



Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM



#41
Benzo(g,h,i)perylene
Concen: 0.42 ng
RT: 26.327 min Scan# 3377
Delta R.T. -0.003 min
Lab File: BN014074.D
Acq: 24 Mar 2021 17:23

Tgt Ion:	276	Resp:	30340
Ion Ratio	Lower	Upper	
276	100		
277	24.2	20.0	30.0
138	20.8	17.8	26.8

Instrument :
BNA_N
ClientSampleId :
PB135330BS

Manual Integrations
APPROVED

mohammad
3/25/2021 11:50:47 AM

